

2092374

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/26/2017 12:28

Test: *Reactive*
~~Total~~ CN *Mg/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.689	0.0	0.090	
ICB1	-0.848	0.0	0.004	
CCV1	249.964	0.0	0.227	
CCB1	-0.900	0.0	0.004	
PB105329BL	-1.180	0.0	0.003	
I6684-06	-1.142	0.0	0.003	
I6684-06DUP	-0.890	0.0	0.004	
I6686-06	-1.038	0.0	0.003	
I7049-03	-0.730	0.0	0.004	
I7055-01	-0.890	0.0	0.004	
I7055-02	-0.815	0.0	0.004	
I7056-03	-1.004	0.0	0.004	
I7061-01	-0.697	0.0	0.004	
I7069-11	-0.715	0.0	0.004	
CCV2	249.081	0.0	0.226	
CCB2	-0.710	0.0	0.004	
I7069-12	-0.804	0.0	0.004	
I7069-13	-0.979	0.0	0.004	
I7069-14	-1.052	0.0	0.003	
I7069-15	-1.059	0.0	0.003	
CCV3	267.533	0.0	0.242	
CCB3	-0.659	0.0	0.004	

N 22
Mean 38.507
SD 90.7218
CV% 235.60

AS

Aquakem v. 7.2AQ1

Results from time period:

Tue Dec 26 10:44:25 2017

Tue Dec 26 12:12:55 2017

Sample Id	Sam/Ctr/c/	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-1.141	µg/l	12/26/2017 10:44:25	
5.OPPBCN	A	Reactive CN	P	4.1128	µg/l	12/26/2017 10:44:26	
10PPBCN	A	Reactive CN	P	9.2438	µg/l	12/26/2017 10:44:27	
50PPBCN	A	Reactive CN	P	49.6282	µg/l	12/26/2017 10:44:28	
100PPBCN	A	Reactive CN	P	102.5852	µg/l	12/26/2017 10:44:29	
250PPBCN	A	Reactive CN	P	252.0538	µg/l	12/26/2017 10:44:30	
500PPBCN	A	Reactive CN	P	498.5173	µg/l	12/26/2017 10:44:31	
ICV1	S	Reactive CN	P	96.6888	µg/l	12/26/2017 12:01:01	
ICB1	S	Reactive CN	P	-0.8481	µg/l	12/26/2017 12:01:02	
CCV1	S	Reactive CN	P	249.9643	µg/l	12/26/2017 12:01:05	
CCB1	S	Reactive CN	P	-0.8995	µg/l	12/26/2017 12:01:06	
PB105329BL	S	Reactive CN	P	-1.1795	µg/l	12/26/2017 12:01:08	
I6684-06	S	Reactive CN	P	-1.1422	µg/l	12/26/2017 12:01:09	
I6684-06DUP	S	Reactive CN	P	-0.8905	µg/l	12/26/2017 12:01:10	
I6686-06	S	Reactive CN	P	-1.0384	µg/l	12/26/2017 12:08:33	
I7049-03	S	Reactive CN	P	-0.7302	µg/l	12/26/2017 12:08:34	
I7055-01	S	Reactive CN	P	-0.8897	µg/l	12/26/2017 12:08:35	
I7055-02	S	Reactive CN	P	-0.8146	µg/l	12/26/2017 12:08:36	
I7056-03	S	Reactive CN	P	-1.0044	µg/l	12/26/2017 12:08:37	
I7061-01	S	Reactive CN	P	-0.697	µg/l	12/26/2017 12:08:38	
I7069-11	S	Reactive CN	P	-0.7149	µg/l	12/26/2017 12:08:39	
CCV2	S	Reactive CN	P	249.0809	µg/l	12/26/2017 12:08:40	
CCB2	S	Reactive CN	P	-0.71	µg/l	12/26/2017 12:08:41	
I7069-12	S	Reactive CN	P	-0.8041	µg/l	12/26/2017 12:08:42	
I7069-13	S	Reactive CN	P	-0.9792	µg/l	12/26/2017 12:08:43	
I7069-14	S	Reactive CN	P	-1.0519	µg/l	12/26/2017 12:12:52	
I7069-15	S	Reactive CN	P	-1.0585	µg/l	12/26/2017 12:12:53	
CCV3	S	Reactive CN	P	267.5329	µg/l	12/26/2017 12:12:54	
CCB3	S	Reactive CN	P	-0.6589	µg/l	12/26/2017 12:12:55	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS.

12/26/2017 10:48


Test Total CN

Accepted 12/26/2017 10:48

Factor 1125
Bias 0.004

Coeff. of det. 0.999923

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-1.1410	0.0000	
2	5.0PPBCN	0.008	4.1128	5.0000	
3	10PPBCN	0.013	9.2438	10.0000	
4	50PPBCN	0.049	49.6282	50.0000	
5	100PPBCN	0.096	102.5852	100.0000	
6	250PPBCN	0.229	252.0538	250.0000	
7	500PPBCN	0.448	498.5173	500.0000	