

CB92492

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

1/5/2018 13:43

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.121	0.0	0.092	
ICB1	0.166	0.0	0.004	
CCV1	243.593	0.0	0.228	
CCB1	0.247	0.0	0.004	
PB105494BL	0.064	0.0	0.004	
J1011-02	0.022	0.0	0.004	
J1011-02DUP	0.026	0.0	0.004	
J1011-04	-0.132	0.0	0.004	
J1011-06	0.056	0.0	0.004	
J1011-08	0.176	0.0	0.004	
J1012-02	-0.076	0.0	0.004	
J1012-04	-0.001	0.0	0.004	
J1012-08	-0.005	0.0	0.004	
J1012-10	0.293	0.0	0.004	
CCV2	243.883	0.0	0.228	
CCB2	0.364	0.0	0.004	
J1012-12	0.105	0.0	0.004	
J1012-14	-0.051	0.0	0.004	
J1012-16	-0.008	0.0	0.004	
J1016-02	0.088	0.0	0.004	
J1033-07	-0.039	0.0	0.004	
J1033-08	-0.045	0.0	0.004	
J1033-09	0.027	0.0	0.004	
J1033-10	0.205	0.0	0.004	
J1033-11	0.053	0.0	0.004	
J1033-12	0.141	0.0	0.004	
CCV3	245.100	0.0	0.230	
CCB3	0.484	0.0	0.004	

N 28
Mean 29.673
SD 77.8117
CV% 262.23

CB92492

Aquakem v. 7.2AQ1

Results from time period:

AK

Fri Jan 05 13:28:09 2018

Fri Jan 05 13:42:57 2018

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time
0.OPPBCN	A	Reactive CN	P	-0.4067	µg/l	1/5/2018 12:03:23
5.OPPBCN	A	Reactive CN	P	5.7922	µg/l	1/5/2018 12:03:24
10PPBCN	A	Reactive CN	P	9.1858	µg/l	1/5/2018 12:03:25
50PPBCN	A	Reactive CN	P	50.6381	µg/l	1/5/2018 12:03:26
100PPBCN	A	Reactive CN	P	99.5216	µg/l	1/5/2018 12:03:27
250PPBCN	A	Reactive CN	P	250.4576	µg/l	1/5/2018 12:03:28
500PPBCN	A	Reactive CN	P	499.8114	µg/l	1/5/2018 12:03:29
ICV1	S	Reactive CN	P	96.1208	µg/l	1/5/2018 13:28:10
ICB1	S	Reactive CN	P	0.166	µg/l	1/5/2018 13:28:11
CCV1	S	Reactive CN	P	243.593	µg/l	1/5/2018 13:28:14
CCB1	S	Reactive CN	P	0.2467	µg/l	1/5/2018 13:28:15
PB105494BL	S	Reactive CN	P	0.064	µg/l	1/5/2018 13:28:17
J1011-02	S	Reactive CN	P	0.0219	µg/l	1/5/2018 13:28:18
J1011-02DUP	S	Reactive CN	P	0.0264	µg/l	1/5/2018 13:28:19
J1011-04	S	Reactive CN	P	-0.1322	µg/l	1/5/2018 13:35:41
J1011-06	S	Reactive CN	P	0.0563	µg/l	1/5/2018 13:35:42
J1011-08	S	Reactive CN	P	0.1759	µg/l	1/5/2018 13:35:43
J1012-02	S	Reactive CN	P	-0.0761	µg/l	1/5/2018 13:35:44
J1012-04	S	Reactive CN	P	-0.0008	µg/l	1/5/2018 13:35:45
J1012-08	S	Reactive CN	P	-0.0051	µg/l	1/5/2018 13:35:46
J1012-10	S	Reactive CN	P	0.2928	µg/l	1/5/2018 13:35:47
CCV2	S	Reactive CN	P	243.8826	µg/l	1/5/2018 13:35:48
CCB2	S	Reactive CN	P	0.3636	µg/l	1/5/2018 13:35:49
J1012-12	S	Reactive CN	P	0.1051	µg/l	1/5/2018 13:35:50
J1012-14	S	Reactive CN	P	-0.051	µg/l	1/5/2018 13:35:51
J1012-16	S	Reactive CN	P	-0.0077	µg/l	1/5/2018 13:42:48
J1016-02	S	Reactive CN	P	0.088	µg/l	1/5/2018 13:42:49
J1033-07	S	Reactive CN	P	-0.0385	µg/l	1/5/2018 13:42:50
J1033-08	S	Reactive CN	P	-0.0453	µg/l	1/5/2018 13:42:51
J1033-09	S	Reactive CN	P	0.0273	µg/l	1/5/2018 13:42:52
J1033-10	S	Reactive CN	P	0.2051	µg/l	1/5/2018 13:42:53
J1033-11	S	Reactive CN	P	0.0531	µg/l	1/5/2018 13:42:54
J1033-12	S	Reactive CN	P	0.1413	µg/l	1/5/2018 13:42:55
CCV3	S	Reactive CN	P	245.1004	µg/l	1/5/2018 13:42:56
CCB3	S	Reactive CN	P	0.4841	µg/l	1/5/2018 13:42:57

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AIC

1/5/2018 13:53


Test Total CN

Accepted 1/5/2018 12:03

Factor 1085
 Bias 0.004

Coeff. of det. 0.999989

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.4067	0.0000	
2	5.0PPBCN	0.009	5.7922	5.0000	
3	10PPBCN	0.012	9.1858	10.0000	
4	50PPBCN	0.050	50.6381	50.0000	
5	100PPBCN	0.095	99.5216	100.0000	
6	250PPBCN	0.235	250.4576	250.0000	
7	500PPBCN	0.464	499.8114	500.0000	