

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/10/2018 13:32

Test: *Reactive*
Total CN *Hg Id*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.701	0.0	0.093	
ICB1	-0.034	0.0	0.004	
CCV1	240.035	0.0	0.231	
CCB1	-0.056	0.0	0.004	
PB105550BL	-0.136	0.0	0.004	
J1013-02	-0.351	0.0	0.004	
J1013-02DUP	-0.213	0.0	0.004	
J1064-15	-0.346	0.0	0.004	
J1067-03	0.052	0.0	0.004	
CCV2	249.783	0.0	0.240	
CCB2	0.080	0.0	0.004	

N	11
Mean	53.047
SD	99.0067
CV%	186.64

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/10/2018 11:52


Test Total CN

Accepted 1/10/2018 11:52

Factor 1057
Bias 0.004

Coeff. of det. 0.999989

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.3139	0.0000	
2	5.0PPBCN	0.008	4.7201	5.0000	
3	10PPBCN	0.013	9.5690	10.0000	
4	50PPBCN	0.051	50.2891	50.0000	
5	100PPBCN	0.099	100.1973	100.0000	
6	250PPBCN	0.242	251.1906	250.0000	
7	500PPBCN	0.476	499.3477	500.0000	

AS

Aquakem v. 7.2AQ1

Results from time period:

Wed Jan 10 13:27:01 2018

Wed Jan 10 13:31:23 2018

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.3139	µg/l	1/10/2018 11:45:37	
5.OPPBCN	A	Reactive CN	P	4.7201	µg/l	1/10/2018 11:45:38	
10PPBCN	A	Reactive CN	P	9.569	µg/l	1/10/2018 11:45:39	
50PPBCN	A	Reactive CN	P	50.2891	µg/l	1/10/2018 11:45:40	
100PPBCN	A	Reactive CN	P	100.1973	µg/l	1/10/2018 11:45:41	
250PPBCN	A	Reactive CN	P	251.1906	µg/l	1/10/2018 11:45:42	
500PPBCN	A	Reactive CN	P	499.3477	µg/l	1/10/2018 11:45:43	
ICV1	S	Reactive CN	P	94.7008	µg/l	1/10/2018 13:27:01	
ICB1	S	Reactive CN	P	-0.0337	µg/l	1/10/2018 13:27:03	
CCV1	S	Reactive CN	P	240.0346	µg/l	1/10/2018 13:27:05	
CCB1	S	Reactive CN	P	-0.0559	µg/l	1/10/2018 13:27:08	
PB105550BL	S	Reactive CN	P	-0.1364	µg/l	1/10/2018 13:27:09	
J1013-02	S	Reactive CN	P	-0.3506	µg/l	1/10/2018 13:27:10	
J1013-02DUP	S	Reactive CN	P	-0.2134	µg/l	1/10/2018 13:27:11	
J1064-15	S	Reactive CN	P	-0.346	µg/l	1/10/2018 13:31:20	
J1067-03	S	Reactive CN	P	0.0523	µg/l	1/10/2018 13:31:21	
CCV2	S	Reactive CN	P	249.7832	µg/l	1/10/2018 13:31:22	
CCB2	S	Reactive CN	P	0.0796	µg/l	1/10/2018 13:31:23	