

21392568

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/11/2018 15:02

Test: *Reactive*
~~Total~~ CN *ug/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.697	0.0	0.095	
ICB1	0.854	0.0	0.004	
CCV1	248.142	0.0	0.235	
CCB1	0.724	0.0	0.004	
PB105583BL	0.800	0.0	0.004	
J1078-05	0.386	0.0	0.004	
J1078-05DUP	0.564	0.0	0.004	
J1088-04	0.388	0.0	0.004	
CCV2	264.032	0.0	0.250	
CCB2	0.951	0.0	0.004	

N	10
Mean	61.454
SD	107.0149
CV%	174.14

AS

Aquakem v. 7.2AQ1

Results from time period:

Thu Jan 11 14:54:58 2018

Thu Jan 11 14:59:49 2018

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	0.1524	µg/l	1/11/2018 11:35:55	
5.0PPBCN	A	Reactive CN	P	4.747	µg/l	1/11/2018 11:35:56	
10PPBCN	A	Reactive CN	P	10.2514	µg/l	1/11/2018 11:35:57	
50PPBCN	A	Reactive CN	P	49.3987	µg/l	1/11/2018 11:35:58	
100PPBCN	A	Reactive CN	P	100.2679	µg/l	1/11/2018 11:35:59	
250PPBCN	A	Reactive CN	P	250.3571	µg/l	1/11/2018 11:36:00	
500PPBCN	A	Reactive CN	P	499.8255	µg/l	1/11/2018 11:36:01	
ICV1	S	Reactive CN	P	97.6966	µg/l	1/11/2018 14:54:59	
ICB1	S	Reactive CN	P	0.8541	µg/l	1/11/2018 14:55:00	
CCV1	S	Reactive CN	P	248.1421	µg/l	1/11/2018 14:55:02	
CCB1	S	Reactive CN	P	0.7241	µg/l	1/11/2018 14:55:05	
PB105583BL	S	Reactive CN	P	0.8004	µg/l	1/11/2018 14:55:06	
J1078-05	S	Reactive CN	P	0.3862	µg/l	1/11/2018 14:55:07	
J1078-05DUP	S	Reactive CN	P	0.5642	µg/l	1/11/2018 14:55:08	
J1088-04	S	Reactive CN	P	0.3881	µg/l	1/11/2018 14:59:45	
CCV2	S	Reactive CN	P	264.0318	µg/l	1/11/2018 14:59:46	
CCB2	S	Reactive CN	P	0.9506	µg/l	1/11/2018 14:59:47	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/11/2018 11:45


Test Total CN

Accepted 1/11/2018 11:45

Factor 1071
Bias 0.003

Coeff. of det. 0.999996

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.1524	0.0000	
2	5.0PPBCN	0.008	4.7470	5.0000	
3	10PPBCN	0.013	10.2514	10.0000	
4	50PPBCN	0.049	49.3987	50.0000	
5	100PPBCN	0.097	100.2679	100.0000	
6	250PPBCN	0.237	250.3571	250.0000	
7	500PPBCN	0.470	499.8255	500.0000	