

=====

Test results

Aquakem 7.2AQ1

2B92656
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/17/2018 12:59

Test: Phenolics- mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.018	0.0	0.124	
ICB1	0.004	0.0	0.017	
CCV1	0.977	0.0	0.120	
CCB1	0.005	0.0	0.017	
PB105701BL	0.003	0.0	0.017	
PB105701BS	1.026	0.0	0.125	
J1112-03	0.004	0.0	0.017	
J1112-03DUP	0.001	0.0	0.016	
J1112-03MS	0.700	0.0	0.090	
J1112-03MSD	0.706	0.0	0.091	
J1112-04	0.007	0.0	0.017	
J1112-05	0.007	0.0	0.017	
J1112-08	0.004	0.0	0.017	
CCV2	0.982	0.0	0.120	
CCB2	0.004	0.0	0.017	

N	15
Mean	0.363
SD	0.4643
CV%	127.83

AS

Aquakem v. 7.2AQ1

Results from time period:

Wed Jan 17 12:11:50 2018

Wed Jan 17 12:55:36 2018

Sample Id	Sam/Ctr/c/	Test short n:	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0528	mg/l	1/17/2018 12:11:50	
0.1PPM	A	Phenolics-	P	0.1029	mg/l	1/17/2018 12:11:51	
0.25PPM	A	Phenolics-	P	0.2506	mg/l	1/17/2018 12:11:52	
0.50PPM	A	Phenolics-	P	0.4964	mg/l	1/17/2018 12:11:53	
1.0PPM	A	Phenolics-	P	0.9933	mg/l	1/17/2018 12:11:54	
2.0PPM	A	Phenolics-	P	2.004	mg/l	1/17/2018 12:11:55	
ICV1	S	Phenolics-	P	1.0183	mg/l	1/17/2018 12:48:44	
ICB1	S	Phenolics-	P	0.0038	mg/l	1/17/2018 12:48:45	
CCV1	S	Phenolics-	P	0.9767	mg/l	1/17/2018 12:48:48	
CCB1	S	Phenolics-	P	0.0047	mg/l	1/17/2018 12:48:49	
PB105701BL	S	Phenolics-	P	0.0034	mg/l	1/17/2018 12:48:51	
PB105701BS	S	Phenolics-	P	1.0263	mg/l	1/17/2018 12:48:52	
J1112-03	S	Phenolics-	P	0.0038	mg/l	1/17/2018 12:48:53	
J1112-03DUP	S	Phenolics-	P	0.001	mg/l	1/17/2018 12:48:54	
J1112-03MS	S	Phenolics-	P	0.6996	mg/l	1/17/2018 12:55:29	
J1112-03MSD	S	Phenolics-	P	0.7061	mg/l	1/17/2018 12:55:30	
J1112-04	S	Phenolics-	P	0.0073	mg/l	1/17/2018 12:55:31	
J1112-05	S	Phenolics-	P	0.0069	mg/l	1/17/2018 12:55:32	
J1112-08	S	Phenolics-	P	0.0039	mg/l	1/17/2018 12:55:33	
CCV2	S	Phenolics-	P	0.9824	mg/l	1/17/2018 12:55:34	
CCB2	S	Phenolics-	P	0.0044	mg/l	1/17/2018 12:55:35	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/17/2018 12:20


Test Phenolics-

Accepted 1/17/2018 12:20

Factor 9.429
Bias 0.016

Coeff. of det. 0.999968

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.022	0.0528	0.0500	
2	Phenol-2	0.027	0.1029	0.1000	
3	Phenol-2	0.043	0.2506	0.2500	
4	Phenol-2	0.069	0.4964	0.5000	
5	Phenol-2	0.122	0.9933	1.0000	
6	Phenol-2	0.229	2.0040	2.0000	