

=====

Test results

Aquakem 7.2AQ1

LB92765
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/22/2018 14:41

Test: Total CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.432	0.0	0.093	
ICB1	-0.308	0.0	0.004	
CCV1	250.631	0.0	0.236	
CCB1	-0.183	0.0	0.004	
PB105786BL	-0.620	0.0	0.004	
PB105786BS	96.702	0.0	0.094	
MIDPB105786	244.857	0.0	0.230	
J1192-30	708.823	0.0	0.658	Test limit high
J1192-30DLX2	351.189	0.0	0.328	
CCV2	254.590	0.0	0.239	
CCB2	-0.252	0.0	0.004	

N	11
Mean	181.896
SD	216.3007
CV%	118.91

AS

Aquakem v. 7.2AQ1

Results from time period:

Mon Jan 22 13:59:39 2018

Mon Jan 22 14:32:12 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-1.1274	µg/l	1/22/2018 9:10:11	
5.0PPBCN	A	Total CN	P	5.166	µg/l	1/22/2018 9:10:12	
10PPBCN	A	Total CN	P	10.1166	µg/l	1/22/2018 9:10:13	
50PPBCN	A	Total CN	P	49.9282	µg/l	1/22/2018 9:10:14	
100PPBCN	A	Total CN	P	100.4692	µg/l	1/22/2018 9:10:15	
250PPBCN	A	Total CN	P	251.0761	µg/l	1/22/2018 9:10:16	
500PPBCN	A	Total CN	P	499.3713	µg/l	1/22/2018 9:10:17	
ICV1	S	Total CN	P	95.432	µg/l	1/22/2018 13:59:39	
ICB1	S	Total CN	P	-0.3081	µg/l	1/22/2018 13:59:42	
CCV1	S	Total CN	P	250.6313	µg/l	1/22/2018 13:59:44	
CCB1	S	Total CN	P	-0.1833	µg/l	1/22/2018 13:59:45	
PB105786BL	S	Total CN	P	-0.6196	µg/l	1/22/2018 13:59:47	
PB105786BS	S	Total CN	P	96.7015	µg/l	1/22/2018 13:59:48	
MIDPB105786	S	Total CN	P	244.8568	µg/l	1/22/2018 13:59:49	
J1192-30	S	Total CN	P	708.8226	µg/l	1/22/2018 14:05:22	
J1192-30DLX2	S	Total CN	P	351.1887	µg/l	1/22/2018 14:05:23	
CCV2	S	Total CN	P	254.5904	µg/l	1/22/2018 14:05:27	
CCB2	S	Total CN	P	-0.2519	µg/l	1/22/2018 14:05:28	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

1/22/2018 9:14


Test Total CN

Accepted 1/22/2018 9:14

Factor 1085
 Bias 0.005

Coeff. of det. 0.999985

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-1.1274	0.0000	
2	5.0PPBCN	0.009	5.1660	5.0000	
3	10PPBCN	0.014	10.1166	10.0000	
4	50PPBCN	0.051	49.9282	50.0000	
5	100PPBCN	0.097	100.4692	100.0000	
6	250PPBCN	0.236	251.0761	250.0000	
7	500PPBCN	0.465	499.3713	500.0000	