



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB92794	Date Received:	
Client Sample ID:	HIGHPB105882	SDG No.:	LB92794
Lab Sample ID:	HIGHPB105882	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.49		1	0.003		0.005	mg/L		01/23/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/22/2018 12:00:00 AM
Project:	LB92794	Date Received:	1/22/2018 12:00:00 AM
Client Sample ID:	CL-01-012218-A	SDG No.:	LB92794
Lab Sample ID:	J1015-12	Matrix:	Solid
Level (low/med):	low	% Solid:	82.2

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.08	J	1	0.039		0.3	mg/Kg	01/23/2018	01/23/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/22/2018 12:00:00 AM
Project:	LB92794	Date Received:	1/22/2018 12:00:00 AM
Client Sample ID:	CL-01-012218-B	SDG No.:	LB92794
Lab Sample ID:	J1015-14	Matrix:	Solid
Level (low/med):	low	% Solid:	90.5

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.034	U	1	0.036		0.27	mg/Kg	01/23/2018	01/23/2018	9012B

Report of Analysis

Client:		Date Collected:	1/22/2018 12:00:00 AM
Project:	LB92794	Date Received:	1/22/2018 12:00:00 AM
Client Sample ID:	CL-01-012218-C	SDG No.:	LB92794
Lab Sample ID:	J1015-16	Matrix:	Solid
Level (low/med):	low	% Solid:	78.6

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.093	J	1	0.041		0.31	mg/Kg	01/23/2018	01/23/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/22/2018 12:00:00 AM
Project:	LB92794	Date Received:	1/22/2018 12:00:00 AM
Client Sample ID:	CL-01-012218-D	SDG No.:	LB92794
Lab Sample ID:	J1015-18	Matrix:	Solid
Level (low/med):	low	% Solid:	85.3

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.22	J	1	0.037		0.28	mg/Kg	01/23/2018	01/23/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/22/2018 12:00:00 AM
Project:	LB92794	Date Received:	1/22/2018 12:00:00 AM
Client Sample ID:	OK-01-012218	SDG No.:	LB92794
Lab Sample ID:	J1016-05	Matrix:	Solid
Level (low/med):	low	% Solid:	90.9

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.33		1	0.035		0.26	mg/Kg	01/23/2018	01/23/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/22/2018 12:00:00 AM
Project:	LB92794	Date Received:	1/22/2018 12:00:00 AM
Client Sample ID:	SU-04-012218	SDG No.:	LB92794
Lab Sample ID:	J1022-08	Matrix:	Solid
Level (low/med):	low	% Solid:	86.1

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.078	J	1	0.038		0.29	mg/Kg	01/23/2018	01/23/2018	9012B

Report of Analysis

Client:		Date Collected:	1/22/2018 12:00:00 AM
Project:	LB92794	Date Received:	1/22/2018 12:00:00 AM
Client Sample ID:	W12-1(0-2)	SDG No.:	LB92794
Lab Sample ID:	J1251-01	Matrix:	Solid
Level (low/med):	low	% Solid:	73.5

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.028	U	1	0.044		0.33	mg/Kg	01/23/2018	01/23/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/22/2018 12:00:00 AM
Project:	LB92794	Date Received:	1/22/2018 12:00:00 AM
Client Sample ID:	W12-2(0-2)	SDG No.:	LB92794
Lab Sample ID:	J1251-03	Matrix:	Solid
Level (low/med):	low	% Solid:	71.6

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.014	U	1	0.046		0.35	mg/Kg	01/23/2018	01/23/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/22/2018 12:00:00 AM
Project:	LB92794	Date Received:	1/22/2018 12:00:00 AM
Client Sample ID:	W12-3(0-2)	SDG No.:	LB92794
Lab Sample ID:	J1251-05	Matrix:	Solid
Level (low/med):	low	% Solid:	72.1

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.0074	U	1	0.045		0.34	mg/Kg	01/23/2018	01/23/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/22/2018 12:00:00 AM
Project:	LB92794	Date Received:	1/22/2018 12:00:00 AM
Client Sample ID:	W12-4(0-2)	SDG No.:	LB92794
Lab Sample ID:	J1251-07	Matrix:	Solid
Level (low/med):	low	% Solid:	75.8

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.021	U	1	0.042		0.32	mg/Kg	01/23/2018	01/23/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/22/2018 12:00:00 AM
Project:	LB92794	Date Received:	1/22/2018 12:00:00 AM
Client Sample ID:	CONCRETE-PIERS	SDG No.:	LB92794
Lab Sample ID:	J1258-01	Matrix:	Solid
Level (low/med):	low	% Solid:	100

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.07	J	1	0.032		0.25	mg/Kg	01/23/2018	01/23/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/22/2018 12:00:00 AM
Project:	LB92794	Date Received:	1/22/2018 12:00:00 AM
Client Sample ID:	MH-198	SDG No.:	LB92794
Lab Sample ID:	J1266-01	Matrix:	Solid
Level (low/med):	low	% Solid:	87.5

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.21	J	1	0.037		0.28	mg/Kg	01/23/2018	01/23/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/22/2018 12:00:00 AM
Project:	LB92794	Date Received:	1/22/2018 12:00:00 AM
Client Sample ID:	MH-199	SDG No.:	LB92794
Lab Sample ID:	J1266-05	Matrix:	Solid
Level (low/med):	low	% Solid:	88.6

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.045	J	1	0.037		0.28	mg/Kg	01/23/2018	01/23/2018	9012B

Report of Analysis

Client:		Date Collected:	1/22/2018 12:00:00 AM
Project:	LB92794	Date Received:	1/22/2018 12:00:00 AM
Client Sample ID:	MH-200	SDG No.:	LB92794
Lab Sample ID:	J1266-09	Matrix:	Solid
Level (low/med):	low	% Solid:	87.1

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.014	U	1	0.036		0.28	mg/Kg	01/23/2018	01/23/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB92794	Date Received:	
Client Sample ID:	LOWPB105882	SDG No.:	LB92794
Lab Sample ID:	LOWPB105882	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.0094		1	0.003		0.005	mg/L		01/23/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB92794	Date Received:	
Client Sample ID:	MDL-WATER-01	SDG No.:	LB92794
Lab Sample ID:	MDL-WATER-01	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.00072	U	1	0.003		0.005	mg/L		01/23/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB92794	Date Received:	
Client Sample ID:	MDL-WATER-02	SDG No.:	LB92794
Lab Sample ID:	MDL-WATER-02	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.0012	U	1	0.003		0.005	mg/L		01/23/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB92794	Date Received:	
Client Sample ID:	MDL-WATER-03	SDG No.:	LB92794
Lab Sample ID:	MDL-WATER-03	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.00085	U	1	0.003		0.005	mg/L		01/23/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB92794	Date Received:	
Client Sample ID:	MDL-WATER-03	SDG No.:	LB92794
Lab Sample ID:	MDL-WATER-03	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
-----	-----------	-------	------	----	-----	-----	------------	-------	-----------	-----------	----------

Color Before:	Clarity Before:	Texture:
Color After:	Clarity After:	Artifacts:
Comments:		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB92794

SDG No.: LB92794

Contract: _____

Lab Code: CHEM

Case No.: LB92794

SAS No.: LB92794

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV1	Cyanide	0.096	0.099	97		85 - 115		01/23/2018	13:23	LB92794



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB92794

SDG No.: LB92794

Contract: _____ Lab Code: CHEM

Case No.: LB92794

SAS No.: LB92794

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	Cyanide	0.25	0.25	100		90 - 110		01/23/2018	13:23	LB92794
CCV2	Cyanide	0.25	0.25	100		90 - 110		01/23/2018	13:31	LB92794
CCV3	Cyanide	0.25	0.25	100		90 - 110		01/23/2018	13:39	LB92794
CCV4	Cyanide	0.26	0.25	104		90 - 110		01/23/2018	14:05	LB92794



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB92794 **SDG No.:** LB92794
Contract: **Lab Code:** CHEM **Case No.:** LB92794 **SAS No.:** LB92794

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	Cyanide	-0	+/-0.005	U		0.005		01/23/2018	13:23	LB92794



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB92794

SDG No.: LB92794

Contract:

Lab Code: CHEM

Case No.: LB92794

SAS No.: LB92794

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	Cyanide	0.0001	+/-0.005	U		0.005		01/23/2018	13:23	LB92794
CCB2	Cyanide	0.00015	+/-0.005	U		0.005		01/23/2018	13:31	LB92794
CCB3	Cyanide	0.00027	+/-0.005	U		0.005		01/23/2018	13:39	LB92794
CCB4	Cyanide	-0.00015	+/-0.005	U		0.005		01/23/2018	14:05	LB92794



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB92794

SDG No.: LB92794

Contract: Lab Code: CHEM

Case No.: LB92794

SAS No.: LB92794

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



GENCHEM
- 3b -
PREPARATION BLANK SUMMARY

Client: LB92794

SDG No.: LB92794

Instrument: Konelab 20

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	LOD mg/Kg	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB105882BL	Cyanide	SOLID		Batch Number:		PB105882		Prep Date:	01/23/2018	
		-0.0077	<0.25	U		0.25		01/23/2018	13:24	LB92794
Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	LOD mg/Kg	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB105883BL	Cyanide	SOLID		Batch Number:		PB105883		Prep Date:	01/23/2018	
		-0.0075	<0.25	U		0.25		01/23/2018	13:39	LB92794



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB92794 level: low sdg no.: LB92794
contract: lab code: CHEM case no.: LB92794 sas no.: LB92794
matrix: Solid sample id: J1015-12 client id: CL-01-012218-AMS
Percent Solids for Sample: 82.2 Spiked ID: J1015-12MS Percent Solids for Spike Sample: 82.2

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	2.1		0.08	J	2.4	84		



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB92794 level: low sdg no.: LB92794
contract: lab code: CHEM case no.: LB92794 sas no.: LB92794
matrix: Solid sample id: J1015-12 client id: CL-01-012218-AMSD
Percent Solids for Sample: 82.2 Spiked ID: J1015-12MSD Percent Solids for Spike Sample: 82.2

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	2.2		0.08	J	2.4	88		



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB92794 level: low sdg no.: LB92794
contract: lab code: CHEM case no.: LB92794 sas no.: LB92794
matrix: Solid sample id: J1251-01 client id: W12-1(0-2)MS
Percent Solids for Sample: 73.5 Spiked ID: J1251-01MS Percent Solids for Spike Sample: 73.5

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	1.8		0.044	U	2.6	69.2	*	



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB92794 level: low sdg no.: LB92794
contract: lab code: CHEM case no.: LB92794 sas no.: LB92794
matrix: Solid sample id: J1251-01 client id: W12-1(0-2)MSD
Percent Solids for Sample: 73.5 Spiked ID: J1251-01MSD Percent Solids for Spike Sample: 73.5

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	1.9		0.044	U	2.7	70.4	*	

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB92794 **Level:** LOW **SDG No.:** LB92794
Contract: **Lab Code:** CHEM **Case No.:** LB92794 **SAS No.:** LB92794
Matrix: Solid **Sample ID:** J1015-12 **Client ID:** CL-01-012218-ADUP
Percent Solids for Sample: 82.2 **Duplicate ID** J1015-12DUP **Percent Solids for Spike Sample:** 82.2

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	0.08	J	0.088	J	10		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB92794 **Level:** LOW **SDG No.:** LB92794
Contract: **Lab Code:** CHEM **Case No.:** LB92794 **SAS No.:** LB92794
Matrix: Solid **Sample ID:** J1015-12MS **Client ID:** CL-01-012218-AMSD
Percent Solids for Sample: 82.2 **Duplicate ID** J1015-12MSD **Percent Solids for Spike Sample:** 82.2

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	2.1		2.2		5		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB92794 **Level:** LOW **SDG No.:** LB92794
Contract: **Lab Code:** CHEM **Case No.:** LB92794 **SAS No.:** LB92794
Matrix: Solid **Sample ID:** J1251-01 **Client ID:** W12-1(0-2)DUP
Percent Solids for Sample: 73.5 **Duplicate ID** J1251-01DUP **Percent Solids for Spike Sample:** 73.5

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	0.044	U	0.02	U			

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB92794 **Level:** LOW **SDG No.:** LB92794
Contract: **Lab Code:** CHEM **Case No.:** LB92794 **SAS No.:** LB92794
Matrix: Solid **Sample ID:** J1251-01MS **Client ID:** W12-1(0-2)MSD
Percent Solids for Sample: 73.5 **Duplicate ID** J1251-01MSD **Percent Solids for Spike Sample:** 73.5

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	1.8		1.9		5		



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB92794

SDG No.: LB92794

Contract:

Lab Code: CHEM

Case No.: LB92794

SAS No.: LB92794

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB105882BS Cyanide	mg/Kg	4.8	4.6		95.8	85 - 115	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB92794

SDG No.: LB92794

Contract:

Lab Code: CHEM

Case No.: LB92794

SAS No.: LB92794

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB105883BS Cyanide	mg/Kg	4.8	4.7		97.9	85 - 115	

Daily Analysis Runlog For Sequence/QC Batch ID # LB92794

Review By

SubDirectory

Review On

Test : Cyanide

LB92794

STD. NAME		STD REF.#				
ICAL Standard		WP61179,WP61180,WP61181,WP61182,WP61183,WP61184,WP61185				
ICV Standard		W1541				
CCV Standard		WP61186				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		WP58910				
Chk Standard		WP61188,WP59958,WP60878				

Sr#	SampleID	ClientID	QcType	Date	Comment	Status
1	0.0PPBCN	0.0PPBCN	CAL1	01/23/18 10:06		OK
2	5.0PPBCN	5.0PPBCN	CAL2	01/23/18 10:06		OK
3	10PPBCN	10PPBCN	CAL3	01/23/18 10:06		OK
4	50PPBCN	50PPBCN	CAL4	01/23/18 10:06		OK
5	100PPBCN	100PPBCN	CAL5	01/23/18 10:06		OK
6	250PPBCN	250PPBCN	CAL6	01/23/18 10:06		OK
7	500PPBCN	500PPBCN	CAL7	01/23/18 10:06		OK
8	ICV1	ICV1	ICV	01/23/18 13:23		OK
9	ICB1	ICB1	ICB	01/23/18 13:23		OK
10	CCV1	CCV1	CCV	01/23/18 13:23		OK
11	CCB1	CCB1	CCB	01/23/18 13:23		OK
12	PB105882BL	PB105882BL	MB	01/23/18 13:24		OK
13	PB105882BS	PB105882BS	LCS	01/23/18 13:24		OK
14	LOWPB105882	LOWPB105882	SAM	01/23/18 13:24		OK
15	HIGHPB105882	HIGHPB105882	SAM	01/23/18 13:31		OK
16	J1015-12	CL-01-012218-A	SAM	01/23/18 13:31		OK
17	J1015-12DUP	CL-01-012218-ADUP	DUP	01/23/18 13:31		OK
18	J1015-12MS	CL-01-012218-AMS	MS	01/23/18 13:31		OK
19	J1015-12MSD	CL-01-012218-AMSD	MSD	01/23/18 13:31		OK
20	J1015-14	CL-01-012218-B	SAM	01/23/18 13:31		OK
21	J1015-16	CL-01-012218-C	SAM	01/23/18 13:31		OK

Daily Analysis Runlog For Sequence/QC Batch ID # LB92794

Review By		Review On				
SubDirectory		LB92794		Test : Cyanide		
STD. NAME		STD REF.#				
ICAL Standard		WP61179,WP61180,WP61181,WP61182,WP61183,WP61184,WP61185				
ICV Standard		W1541				
CCV Standard		WP61186				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		WP58910				
Chk Standard		WP61188,WP59958,WP60878				
22	CCV2	CCV2	CCV	01/23/18 13:31		OK
23	CCB2	CCB2	CCB	01/23/18 13:31		OK
24	J1015-18	CL-01-012218-D	SAM	01/23/18 13:31		OK
25	J1016-05	OK-01-012218	SAM	01/23/18 13:31		OK
26	J1022-08	SU-04-012218	SAM	01/23/18 13:38		OK
27	J1258-01	CONCRETE-PIERS	SAM	01/23/18 13:39		OK
28	J1266-01	MH-198	SAM	01/23/18 13:39		OK
29	J1266-05	MH-199	SAM	01/23/18 13:39		OK
30	J1266-09	MH-200	SAM	01/23/18 13:39		OK
31	PB105883BL	PB105883BL	MB	01/23/18 13:39		OK
32	PB105883BS	PB105883BS	LCS	01/23/18 13:39		OK
33	J1251-01	W12-1(0-2)	SAM	01/23/18 13:39		OK
34	CCV3	CCV3	CCV	01/23/18 13:39		OK
35	CCB3	CCB3	CCB	01/23/18 13:39		OK
36	J1251-01DUP	W12-1(0-2)DUP	DUP	01/23/18 13:39		OK
37	J1251-01MS	W12-1(0-2)MS	MS	01/23/18 13:46		OK
38	J1251-01MSD	W12-1(0-2)MSD	MSD	01/23/18 13:46		OK
39	J1251-03	W12-2(0-2)	SAM	01/23/18 13:46		OK
40	J1251-05	W12-3(0-2)	SAM	01/23/18 13:46		OK
41	J1251-07	W12-4(0-2)	SAM	01/23/18 13:46		OK
42	MDL-WATER-01	MDL-WATER-01	SAM	01/23/18 14:05		OK
43	MDL-WATER-02	MDL-WATER-02	SAM	01/23/18 14:05		OK



Instrument ID : KONELAB

284 Sheffield Street, Mountainside, New Jersey 07092 Phone : 908 789 8900 Fax : 908 789 8922

Daily Analysis Runlog For Sequence/QC Batch ID # LB92794

Review By		Review On				
SubDirectory		Test : Cyanide				
LB92794						
STD. NAME		STD REF.#				
ICAL Standard		WP61179,WP61180,WP61181,WP61182,WP61183,WP61184,WP61185				
ICV Standard		W1541				
CCV Standard		WP61186				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		WP58910				
Chk Standard		WP61188,WP59958,WP60878				
44	MDL-WATER-03	MDL-WATER-03	SAM	01/23/18 14:05		OK
45	CCV4	CCV4	CCV	01/23/18 14:05		OK
46	CCB4	CCB4	CCB	01/23/18 14:05		OK