


7471B

LB 92811
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44045



A= 0.0000e+000

B= 2.2575e-004

C= 5.5930e-003

Rho= 0.9999490

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.020	-0.020	-113	0.000	-113				
0.2	0.200	0.176	-0.024	756	0.0 %	756	0			
2.5	2.500	2.508	0.008	11086	0.0 %	11086				
5.0	5.000	5.065	0.065	22410	0.0 %	22410				
7.5	7.500	7.522	0.022	33297	0.0 %	33297				
10.0	10.000	9.949	-0.051	44045	0.0 %	44045				

LB92811 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Jan 2018 09:06:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	-113	-	0.0000	7471B	PPB	24 Jan 2018 09:19:48	S	Std
0.2 - 1	S0.2	756	-	0.2000	7471B	PPB	24 Jan 2018 09:22:05	S	Std
2.5 - 1	S2.5	11086	-	2.5000	7471B	PPB	24 Jan 2018 09:24:24	S	Std
5.0 - 1	S5	22410	-	5.0000	7471B	PPB	24 Jan 2018 09:26:41	S	Std
7.5 - 1	S7.5	33297	-	7.5000	7471B	PPB	24 Jan 2018 09:28:58	S	Std
10.0 - 1	S10	44045	-	10.0000	7471B	PPB	24 Jan 2018 09:36:27	S	Std
ICV61 - 1	ICV61	16957	3.8336	-7471B	PPB	24 Jan 2018 09:40:19	U	SMPL	
ICB61 - 1	ICB61	-234	-0.0472	-7471B	PPB	24 Jan 2018 09:42:35	U	SMPL	
CCV37 - 1	CCV37	22335	5.0476	-7471B	PPB	24 Jan 2018 09:44:50	U	SMPL	
CCB37 - 1	CCB37	-306	-0.0635	-7471B	PPB	24 Jan 2018 09:47:06	U	SMPL	
CRA - 1	CRA	738	0.1722	-7471B	PPB	24 Jan 2018 09:49:23	U	SMPL	
HighStd - 1	HighStd	45535	10.2850	-7471B	PPB	24 Jan 2018 09:51:38	U	SMPL	
ChkStd - 1	ChkStd	31407	7.0956	-7471B	PPB	24 Jan 2018 09:53:55	U	SMPL	
PB105880BL - 1	PBS	49	0.0167	-7471B	PPB	24 Jan 2018 09:56:12	U	SMPL	
PB105880BS - 1	LCSS	15020	3.3963	-7471B	PPB	24 Jan 2018 09:58:30	U	SMPL	
J1230-09 - 1	OILY-DEBRIS	287	0.0704	-7471B	PPB	24 Jan 2018 10:00:47	U	SMPL	
J1230-10 - 1	CO-1-6	748	0.1745	-7471B	PPB	24 Jan 2018 10:03:05	U	SMPL	
J1248-01 - 1	P003-S004-0001-01	5357	1.2149	-7471B	PPB	24 Jan 2018 10:05:22	U	SMPL	
J1248-01DUP - 1	P003-S004-0001-01DUP	5149	1.1680	-7471B	PPB	24 Jan 2018 10:07:39	U	SMPL	
J1248-02 - 1	J1248-01MS	22577	5.1023	-7471B	PPB	24 Jan 2018 10:09:55	U	SMPL	
CCV38 - 1	CCV38	22431	5.0693	-7471B	PPB	24 Jan 2018 10:12:10	U	SMPL	
CCB38 - 1	CCB38	-269	-0.0551	-7471B	PPB	24 Jan 2018 10:14:27	U	SMPL	
J1248-03 - 1	J1248-01MSD	22437	5.0707	-7471B	PPB	24 Jan 2018 10:16:43	U	SMPL	
J1248-04 - 1	P003-S004-0001-02	4670	1.0598	-7471B	PPB	24 Jan 2018 10:18:59	U	SMPL	
J1251-01 - 1	W12-1(0-2)	361	0.0871	-7471B	PPB	24 Jan 2018 10:21:16	U	SMPL	
J1251-03 - 1	W12-2(0-2)	518	0.1225	-7471B	PPB	24 Jan 2018 10:23:32	U	SMPL	
J1251-05 - 1	W12-3(0-2)	614	0.1442	-7471B	PPB	24 Jan 2018 10:25:49	U	SMPL	
J1251-07 - 1	W12-4(0-2)	366	0.0882	-7471B	PPB	24 Jan 2018 10:28:06	U	SMPL	
J1266-01 - 1	MH-198	822	0.1912	-7471B	PPB	24 Jan 2018 10:30:23	U	SMPL	
J1266-05 - 1	MH-199	247	0.0614	-7471B	PPB	24 Jan 2018 10:32:41	U	SMPL	
J1266-09 - 1	MH-200	441	0.1051	-7471B	PPB	24 Jan 2018 10:34:58	U	SMPL	
J1269-01 - 1	ORI-TS	1856	0.4246	-7471B	PPB	24 Jan 2018 10:37:14	U	SMPL	
CCV39 - 1	CCV39	22599	5.1072	-7471B	PPB	24 Jan 2018 10:39:31	U	SMPL	
CCB39 - 1	CCB39	-234	-0.0472	-7471B	PPB	24 Jan 2018 10:41:47	U	SMPL	
J1271-01 - 1	WC-012218	281	0.0690	-7471B	PPB	24 Jan 2018 10:44:04	U	SMPL	
J1272-01 - 1	SB1R-0.5-1.0	6515	1.4763	-7471B	PPB	24 Jan 2018 10:46:21	U	SMPL	
J1272-02 - 1	SB2R-0.5-1.0	5063	1.1485	-7471B	PPB	24 Jan 2018 10:48:36	U	SMPL	
J1272-03 - 1	SB4R-0.5-1.0	5067	1.1495	-7471B	PPB	24 Jan 2018 10:50:53	U	SMPL	
J1272-04 - 1	SB7-0.0-0.5	3412	0.7758	-7471B	PPB	24 Jan 2018 10:53:10	U	SMPL	
J1272-06 - 1	SB8-0.0-0.5	1017	0.2352	-7471B	PPB	24 Jan 2018 10:55:26	U	SMPL	
J1272-08 - 1	SB9-0.0-0.5	9734	2.2030	-7471B	PPB	24 Jan 2018 10:57:44	U	SMPL	
J1272-11 - 1	DUP-2	4103	0.9318	-7471B	PPB	24 Jan 2018 11:00:02	U	SMPL	
PB105927BS - 1	LCSS	14809	3.3487	-7471B	PPB	24 Jan 2018 11:04:36	U	SMPL	
CCV40 - 1	CCV40	22712	5.1327	-7471B	PPB	24 Jan 2018 11:06:53	U	SMPL	
CCB40 - 1	CCB40	-254	-0.0517	-7471B	PPB	24 Jan 2018 11:09:09	U	SMPL	
I1281-01 - 1	SP-A	27409	6.1931	-7471B	PPB	24 Jan 2018 11:11:26	U	SMPL	
I1283-01 - 1	TM-66507-01	59452	13.4267	-7471B	PPB	24 Jan 2018 11:13:43	U	SMPL	
I1283-02 - 1	TM-69-01	85280	19.2573	-7471B	PPB	24 Jan 2018 11:16:00	U	SMPL	
I1283-03 - 1	TM-41446854-01	68605	15.4929	-7471B	PPB	24 Jan 2018 11:18:19	U	SMPL	
I1283-04 - 1	TM-41446854-02	71442	16.1334	-7471B	PPB	24 Jan 2018 11:20:38	U	SMPL	
I1286-01 - 1	GL-WF-01-COMPOSITE	2256	0.5149	-7471B	PPB	24 Jan 2018 11:22:57	U	SMPL	
I1288-01 - 1	EPE-108-4(69-71)	652	0.1528	-7471B	PPB	24 Jan 2018 11:25:17	U	SMPL	
I1288-02 - 1	EPE-108-4(75-77)	745	0.1738	-7471B	PPB	24 Jan 2018 11:27:34	U	SMPL	
I1288-03 - 1	EPE-108-4(80-82)	525	0.1241	-7471B	PPB	24 Jan 2018 11:29:52	U	SMPL	
I1288-04 - 1	EPE-108-4(84-86)	729	0.1702	-7471B	PPB	24 Jan 2018 11:32:08	U	SMPL	
CCV41 - 1	CCV41	21983	4.9682	-7471B	PPB	24 Jan 2018 11:34:25	U	SMPL	
CCB41 - 1	CCB41	-109	-0.0190	-7471B	PPB	24 Jan 2018 11:36:41	U	SMPL	
1288-05 - 1	EPE-108-4(90-92)	583	0.1372	-7471B	PPB	24 Jan 2018 11:38:58	U	SMPL	
1288-06 - 1	EPE-108-4(95-97)	1623	0.3720	-7471B	PPB	24 Jan 2018 11:41:14	U	SMPL	
1288-07 - 1	EPE-121-2(20-22)	660	0.1546	-7471B	PPB	24 Jan 2018 11:43:31	U	SMPL	
1288-08 - 1	EPE-121-2(25-27)	1065	0.2460	-7471B	PPB	24 Jan 2018 11:45:47	U	SMPL	
1288-08DUP - 1	EPE-121-2(25-27)DUP	1038	0.2399	-7471B	PPB	24 Jan 2018 11:48:04	U	SMPL	
1288-08MS - 1	EPE-121-2(25-27)MS	17453	3.9455	-7471B	PPB	24 Jan 2018 11:50:21	U	SMPL	
1288-08MSD - 1	EPE-121-2(25-27)MSD	17287	3.9081	-7471B	PPB	24 Jan 2018 11:52:38	U	SMPL	
PB105927BL - 1	PBS	-164	-0.0314	-7471B	PPB	24 Jan 2018 11:54:56	U	SMPL	
CV42 - 1	CCV42	21814	4.9300	-7471B	PPB	24 Jan 2018 11:57:14	U	SMPL	
CB42 - 1	CCB42	-128	-0.0233	-7471B	PPB	24 Jan 2018 11:59:30	U	SMPL	
1248-01LX5 - 1		927	0.2149	-7471B	PPB	24 Jan 2018 12:01:47	U	SMPL	
1248-01A - 1		22468	5.0777	-7471B	PPB	24 Jan 2018 12:04:04	U	SMPL	
1288-08LX5 - 1		69	0.0212	-7471B	PPB	24 Jan 2018 12:06:20	U	SMPL	
1288-08A - 1		17828	4.0302	-7471B	PPB	24 Jan 2018 12:08:37	U	SMPL	
1283-01DLX5 - 1		10520	2.3804	-7471B	PPB	24 Jan 2018 12:10:54	U	SMPL	
1283-02DLX5 - 1		15680	3.5453	-7471B	PPB	24 Jan 2018 12:13:11	U	SMPL	
1283-03DLX5 - 1		12489	2.8249	-7471B	PPB	24 Jan 2018 12:15:28	U	SMPL	
1283-04DLX5 - 1		12874	2.9119	-7471B	PPB	24 Jan 2018 12:17:45	U	SMPL	
CV43 - 1	CCV43	21749	4.9154	-7471B	PPB	24 Jan 2018 12:20:02	U	SMPL	
CB43 - 1	CCB43	-120	-0.0215	-7471B	PPB	24 Jan 2018 12:22:20	U	SMPL	