

=====

Test results

Aquakem 7.2AQ1

21392924
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/26/2018 17:29

Test: ^{Reactive} ~~Total~~ CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.676	0.0	0.092	
ICB1	-0.336	0.0	0.004	
CCV1	251.832	0.0	0.236	
CCB1	-0.483	0.0	0.004	
PB106076BL	-0.577	0.0	0.004	
J1023-06	1.079	0.0	0.005	
J1023-06DUP	-0.793	0.0	0.004	
J1303-02	-0.916	0.0	0.003	
J1303-03	-0.843	0.0	0.004	
J1305-01	-1.854	0.0	0.003	
J1305-02	-0.722	0.0	0.004	
J1305-03	-0.690	0.0	0.004	
J1305-04	-1.343	0.0	0.003	
J1303-01	-0.457	0.0	0.004	
CCV2	253.151	0.0	0.237	
CCB2	-0.345	0.0	0.004	
J1305-05	-0.022	0.0	0.004	
CCV3	259.434	0.0	0.243	
CCV3	259.833	0.0	0.243	

N	19
Mean	58.506
SD	107.0957
CV%	183.05

AS

Aquakem v. 7.2AQ1

Results from time period:

Fri Jan 26 16:53:34 2018

Fri Jan 26 17:27:47 2018

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.652	µg/l	1/26/2018 10:00:34	
5.OPPBCN	A	Reactive CN	P	4.151	µg/l	1/26/2018 10:00:35	
10PPBCN	A	Reactive CN	P	9.0246	µg/l	1/26/2018 10:00:36	
50PPBCN	A	Reactive CN	P	48.9515	µg/l	1/26/2018 10:00:37	
100PPBCN	A	Reactive CN	P	100.4822	µg/l	1/26/2018 10:00:38	
250PPBCN	A	Reactive CN	P	256.0126	µg/l	1/26/2018 10:00:39	
500PPBCN	A	Reactive CN	P	497.0302	µg/l	1/26/2018 10:00:40	
ICV1	S	Reactive CN	P	95.6758	µg/l	1/26/2018 16:53:34	
ICB1	S	Reactive CN	P	-0.3357	µg/l	1/26/2018 16:53:36	
CCV1	S	Reactive CN	P	251.8315	µg/l	1/26/2018 16:53:38	
CCB1	S	Reactive CN	P	-0.483	µg/l	1/26/2018 16:53:39	
PB106076BL	S	Reactive CN	P	-0.5775	µg/l	1/26/2018 16:53:41	
J1023-06	S	Reactive CN	P	1.0795	µg/l	1/26/2018 16:53:42	
J1023-06DUP	S	Reactive CN	P	-0.793	µg/l	1/26/2018 16:53:43	
J1303-02	S	Reactive CN	P	-0.9165	µg/l	1/26/2018 17:00:55	
J1303-03	S	Reactive CN	P	-0.8431	µg/l	1/26/2018 17:00:56	
J1305-01	S	Reactive CN	P	-1.8538	µg/l	1/26/2018 17:00:57	
J1305-02	S	Reactive CN	P	-0.7216	µg/l	1/26/2018 17:00:58	
J1305-03	S	Reactive CN	P	-0.6903	µg/l	1/26/2018 17:00:59	
J1305-04	S	Reactive CN	P	-1.3427	µg/l	1/26/2018 17:01:00	
J1303-01	S	Reactive CN	P	-0.4571	µg/l	1/26/2018 17:27:40	
CCV2	S	Reactive CN	P	253.1508	µg/l	1/26/2018 17:27:41	
CCB2	S	Reactive CN	P	-0.3449	µg/l	1/26/2018 17:27:42	
J1305-05	S	Reactive CN	P	-0.0223	µg/l	1/26/2018 17:27:43	
CCV3	S	Reactive CN	P	259.4341	µg/l	1/26/2018 17:27:44	
CCV3	S	Reactive CN	P	259.8328	µg/l	1/26/2018 17:27:45	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/26/2018 10:01


Test Total CN

Accepted 1/26/2018 10:01

Factor 1089
Bias 0.004

Coeff. of det. 0.999765

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.6520	0.0000	
2	5.0PPBCN	0.008	4.1510	5.0000	
3	10PPBCN	0.013	9.0246	10.0000	
4	50PPBCN	0.049	48.9515	50.0000	
5	100PPBCN	0.097	100.4822	100.0000	
6	250PPBCN	0.239	256.0126	250.0000	
7	500PPBCN	0.461	497.0302	500.0000	