

=====

Test results

Aquakem 7.2AQ1

2093004
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/30/2018 15:32

Test: Total CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	98.894	0.0	0.094	
ICB1	0.534	0.0	0.004	
CCV1	248.920	0.0	0.232	
CCB1	0.475	0.0	0.004	
PB106149BL	0.752	0.0	0.004	
PB106149BS	101.338	0.0	0.097	
MIDPB106149	253.492	0.0	0.237	
J1161-10	1.214	0.0	0.004	
J1161-12	1.394	0.0	0.005	
J1343-01	5.665	0.0	0.009	
J1343-01DUP	5.448	0.0	0.008	
J1343-01MS	44.243	0.0	0.044	
J1343-01MSD	44.941	0.0	0.045	
J1161-11	5.143	0.0	0.008	
CCV2	267.327	0.0	0.249	
CCB2	1.023	0.0	0.004	

N	16
Mean	67.550
SD	99.5975
CV%	147.44

AS

Aquakem v. 7.2AQ1

Results from time period:

Tue Jan 30 15:05:13 2018

Tue Jan 30 15:31:45 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.5066	µg/l	1/30/2018 12:44:51	
5.0PPBCN	A	Total CN	P	4.9974	µg/l	1/30/2018 12:44:52	
10PPBCN	A	Total CN	P	10.2652	µg/l	1/30/2018 12:44:53	
50PPBCN	A	Total CN	P	50.3216	µg/l	1/30/2018 12:44:54	
100PPBCN	A	Total CN	P	99.0676	µg/l	1/30/2018 12:44:55	
250PPBCN	A	Total CN	P	249.3852	µg/l	1/30/2018 12:44:56	
500PPBCN	A	Total CN	P	500.4564	µg/l	1/30/2018 12:44:57	
ICV1	S	Total CN	P	98.8944	µg/l	1/30/2018 15:05:13	
ICB1	S	Total CN	P	0.5342	µg/l	1/30/2018 15:05:14	
CCV1	S	Total CN	P	248.9199	µg/l	1/30/2018 15:05:16	
CCB1	S	Total CN	P	0.4746	µg/l	1/30/2018 15:05:19	
PB106149BL	S	Total CN	P	0.7521	µg/l	1/30/2018 15:05:20	
PB106149BS	S	Total CN	P	101.3378	µg/l	1/30/2018 15:05:21	
MIDPB106149	S	Total CN	P	253.492	µg/l	1/30/2018 15:05:22	
J1161-10	S	Total CN	P	1.2138	µg/l	1/30/2018 15:05:23	
J1161-12	S	Total CN	P	1.3944	µg/l	1/30/2018 15:11:25	
J1343-01	S	Total CN	P	5.6654	µg/l	1/30/2018 15:11:26	
J1343-01DUP	S	Total CN	P	5.4475	µg/l	1/30/2018 15:11:27	
J1343-01MS	S	Total CN	P	44.243	µg/l	1/30/2018 15:11:28	
J1343-01MSD	S	Total CN	P	44.9406	µg/l	1/30/2018 15:11:29	
J1161-11	S	Total CN	P	5.1427	µg/l	1/30/2018 15:31:43	
CCV2	S	Total CN	P	267.3273	µg/l	1/30/2018 15:31:44	
CCB2	S	Total CN	P	1.0235	µg/l	1/30/2018 15:31:45	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/30/2018 12:46


Test Total CN

Accepted 1/30/2018 12:46

Factor 1087
Bias 0.003

Coeff. of det. 0.999991

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.5066	0.0000	
2	5.0PPBCN	0.008	4.9974	5.0000	
3	10PPBCN	0.013	10.2652	10.0000	
4	50PPBCN	0.050	50.3216	50.0000	
5	100PPBCN	0.095	99.0676	100.0000	
6	250PPBCN	0.233	249.3852	250.0000	
7	500PPBCN	0.464	500.4564	500.0000	