

7471B


LB 93086

INSTRUMENT ID: 242

Linear

 μ Abs.:

42557



A= 0.0000e+000

B= 2.3745e-004

C= -1.4613e-002

Rho= 0.9997830

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.035	-0.035	-87	0.000	-87				
0.2	0.200	0.227	0.027	1017	0.0 %	1017	0			
2.5	2.500	2.585	0.085	10948	0.0 %	10948				
5.0	5.000	4.962	-0.038	20959	0.0 %	20959				
7.5	7.500	7.371	-0.129	31104	0.0 %	31104				
10.0	10.000	10.090	0.090	42557	0.0 %	42557				

LB93086 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Feb 2018 18:47:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-87	-	0.0000	7471B	PPB	02 Feb 2018 18:49:50	S	Std
0.2 - 1	50.2	1017	-	0.2000	7471B	PPB	02 Feb 2018 18:52:07	S	Std
2.5 - 1	52.5	10948	-	2.5000	7471B	PPB	02 Feb 2018 18:54:25	S	Std
5.0 - 1	55	20959	-	5.0000	7471B	PPB	02 Feb 2018 18:56:42	S	Std
7.5 - 1	57.5	31104	-	7.5000	7471B	PPB	02 Feb 2018 18:59:00	S	Std
10.0 - 1	58	42557	-	10.0000	7471B	PPB	02 Feb 2018 19:01:17	S	Std
ICV82 - 1	ICV82	16726	3.9569	-	7471B	PPB	02 Feb 2018 19:04:08	U	SMPL
ICB82 - 1	ICB82	-247	-0.0733	-	7471B	PPB	02 Feb 2018 19:06:23	U	SMPL
CCV63 - 1	CCV63	21043	4.9820	-	7471B	PPB	02 Feb 2018 19:08:39	U	SMPL
CCB63 - 1	CCB63	-250	-0.0740	-	7471B	PPB	02 Feb 2018 19:10:54	U	SMPL
CRA - 1	CRA	971	0.2159	-	7471B	PPB	02 Feb 2018 19:13:10	U	SMPL
HighStd - 1	HighStd	42875	10.1659	-	7471B	PPB	02 Feb 2018 19:15:26	U	SMPL
ChkStd - 1	ChkStd	30028	7.1155	-	7471B	PPB	02 Feb 2018 19:17:42	U	SMPL
PB106266BL - 1	PBS	-187	-0.0590	-	7471B	PPB	02 Feb 2018 19:20:01	U	SMPL
PB106266BS - 1	LCSS	14660	3.4664	-	7471B	PPB	02 Feb 2018 19:22:18	U	SMPL
J1395-01 - 1	NB-08-020118	2560	0.5933	-	7471B	PPB	02 Feb 2018 19:24:35	U	SMPL
J1430-01 - 1	EPE-125-7A(0-5)	2365	0.5470	-	7471B	PPB	02 Feb 2018 19:26:52	U	SMPL
J1430-02 - 1	EPE-125-7A(5-10)	7711	1.8163	-	7471B	PPB	02 Feb 2018 19:30:55	U	SMPL
J1431-01 - 1	132-1-PH-3	-26	-0.0208	-	7471B	PPB	02 Feb 2018 19:33:12	U	SMPL
J1431-04 - 1	132-1-PH-2	163	0.0241	-	7471B	PPB	02 Feb 2018 19:35:28	U	SMPL
CCV64 - 1	CCV64	20824	4.9300	-	7471B	PPB	02 Feb 2018 19:37:44	U	SMPL
CCB64 - 1	CCB64	-229	-0.0690	-	7471B	PPB	02 Feb 2018 19:40:00	U	SMPL
J1431-07 - 1	132-1-PH-1	62	0.0001	-	7471B	PPB	02 Feb 2018 19:42:17	U	SMPL
J1436-01 - 1	B15	1282	0.2898	-	7471B	PPB	02 Feb 2018 19:44:33	U	SMPL
J1436-05 - 1	B13	8062	1.8997	-	7471B	PPB	02 Feb 2018 19:46:50	U	SMPL
J1436-09 - 1	B12	2392	0.5534	-	7471B	PPB	02 Feb 2018 19:49:06	U	SMPL
J1436-13 - 1	B18	1640	0.3748	-	7471B	PPB	02 Feb 2018 19:51:23	U	SMPL
J1436-17 - 1	B25	1109	0.2487	-	7471B	PPB	02 Feb 2018 19:53:40	U	SMPL
J1436-17DUP - 1	B25DUP	965	0.2145	-	7471B	PPB	02 Feb 2018 19:55:57	U	SMPL
J1436-17MS - 1	B25MS	16441	3.8893	-	7471B	PPB	02 Feb 2018 19:58:15	U	SMPL
J1436-17MSD - 1	B25MSD	17669	4.1808	-	7471B	PPB	02 Feb 2018 20:00:32	U	SMPL
CCV65 - 1	CCV65	20918	4.9523	-	7471B	PPB	02 Feb 2018 20:02:48	U	SMPL
CCB65 - 1	CCB65	-239	-0.0714	-	7471B	PPB	02 Feb 2018 20:05:05	U	SMPL
J1436-17LX5 - 1		110	0.0115	-	7471B	PPB	02 Feb 2018 20:07:22	U	SMPL
J1436-17A - 1		17871	4.2288	-	7471B	PPB	02 Feb 2018 20:09:38	U	SMPL
CCV66 - 1	CCV66	21138	5.0046	-	7471B	PPB	02 Feb 2018 20:15:15	U	SMPL
CCB66 - 1	CCB66	-231	-0.0695	-	7471B	PPB	02 Feb 2018 20:17:31	U	SMPL