

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/13/18
 Analyst : AK
 Data File : LB93186.csv

Approved By : AK
 Approved Date : 2/13/18
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis		Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method		Line 2
Sulfate										
STD1	STD1				W	1/15/18				
Sulfate		PASS	0.000		0.000					mg/L
STD2	STD2				W	1/15/18				
Sulfate		PASS	1.304		1.30					mg/L
STD3	STD3				W	1/15/18				
Sulfate		PASS	3.216		3.22					mg/L
STD4	STD4				W	1/15/18				
Sulfate		PASS	5.638		5.64					mg/L
STD5	STD5				W	1/15/18				
Sulfate		PASS	13.997		14.0					mg/L
STD6	STD6				W	1/15/18				
Sulfate		PASS	28.428		28.4					mg/L
STD7	STD7				W	1/15/18				
Sulfate		PASS	38.159		38.2					mg/L
ICV1	ICV1				W	1/15/18				
Sulfate		PASS	13.963		14.00	93.0	90	110		mg/L
ICB1	ICB1				W	1/15/18				
Sulfate		PASS	0.000		0.000			+/-0.1320		mg/L
CCV1	CCV1				W	2/9/18				
Sulfate		PASS	13.766		13.80	92.0	90	110		mg/L
CCB1	CCB1				W	2/9/18				
Sulfate		PASS	0.000		0.000			+/-0.1320		mg/L
LB93186BLS	LB93186BLS				S	2/9/18				
Sulfate		PASS	0.000		0.000			+/-0.0960		mg/Kg
LB93186BSS	LB93186BSS				S	2/9/18				
Sulfate		PASS	14.405		288.00	96.0	90.00	110.00		mg/Kg
J1511-01	MM-1			1	S	2/9/18				
Sulfate		OR	57.009		1310					mg/Kg
J1511-02	MM-2			1	S	2/9/18				
Sulfate		OR	37.968		925					mg/Kg
J1511-03	MM-3			1	S	2/9/18				
Sulfate		PASS	30.363		707					mg/Kg
J1511-03D	MM-3D			1	S	2/9/18				
Sulfate		PASS	29.396		690.00			2.4	20	mg/Kg
J1511-03S	MM-3S			1	S	2/9/18				
Sulfate		FAIL	35.110		824.0	33.0	80	120		mg/Kg
J1511-03SD	MM-3SD			1	S	2/9/18				
Sulfate		FAIL	39.378		921.0	61.0	80	120	11.1	15 mg/Kg
J1511-01DL	MM-1DL			2	S	2/9/18				
Sulfate		PASS	26.552		1220					mg/Kg

Chemtech Consulting Group


Analytical Review Report

Date Printed : 2/13/18

Analyst : AK

Data File : LB93186.csv

Approved By : 

Approved Date : 

Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units
											Line 2
Sulfate											
J1511-02DL	MM-2DL			2	S	2/9/18					
Sulfate		PASS	17.893		872						mg/Kg
CCV2	CCV2				W	2/9/18					
Sulfate		PASS	14.081		14.10	94.0	90	110			mg/L
CCB2	CCB2				W	2/9/18					
Sulfate		PASS	0.000		0.000			+/-0.1320			mg/L
J1525-01	SAMPLE-1			1	S	2/9/18					
Sulfate		PASS	2.169		51.9						mg/Kg
J1525-03	SAMPLE-2			1	S	2/9/18					
Sulfate		PASS	0.951		22.6						mg/Kg
J1525-05	SAMPLE-3			1	S	2/9/18					
Sulfate		PASS	0.840		19.1						mg/Kg
CCV3	CCV3				W	2/9/18					
Sulfate		PASS	14.114		14.10	94.0	90	110			mg/L
CCB3	CCB3				W	2/9/18					
Sulfate		PASS	0.000		0.000			+/-0.1320			mg/L