

=====

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

2/19/2018 14:55

Test: Total CN ug/l

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	98.610	0.0	0.091	
ICB1	0.785	0.0	0.004	
CCV1	252.594	0.0	0.228	
CCB1	0.696	0.0	0.004	
PB106642BL	0.613	0.0	0.004	
PB106642BS	100.111	0.0	0.092	
J1601-02	1.095	0.0	0.004	
J1601-02DUP	1.259	0.0	0.004	
J1601-02MS	1.511	0.0	0.005	
J1601-02MSD	1.063	0.0	0.004	
CCV2	263.342	0.0	0.238	
CCB2	0.338	0.0	0.004	

N	12
Mean	60.168
SD	99.7561
CV%	165.80

AS

Aquakem v. 7.2AQ1

Results from time period:

Mon Feb 19 14:43:54 2018

Mon Feb 19 14:54:53 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	0.1759	µg/l	2/19/2018 12:31:21	
5.OPPBCN	A	Total CN	P	4.9028	µg/l	2/19/2018 12:31:22	
10PPBCN	A	Total CN	P	9.9386	µg/l	2/19/2018 12:31:23	
50PPBCN	A	Total CN	P	50.094	µg/l	2/19/2018 12:31:24	
100PPBCN	A	Total CN	P	98.8742	µg/l	2/19/2018 12:31:25	
250PPBCN	A	Total CN	P	251.5932	µg/l	2/19/2018 12:31:26	
500PPBCN	A	Total CN	P	499.4214	µg/l	2/19/2018 12:31:27	
ICV1	S	Total CN	P	98.6105	µg/l	2/19/2018 14:43:54	
ICB1	S	Total CN	P	0.7854	µg/l	2/19/2018 14:43:55	
CCV1	S	Total CN	P	252.5939	µg/l	2/19/2018 14:43:56	
CCB1	S	Total CN	P	0.6959	µg/l	2/19/2018 14:43:57	
PB106642BL	S	Total CN	P	0.6129	µg/l	2/19/2018 14:43:58	
PB106642BS	S	Total CN	P	100.1107	µg/l	2/19/2018 14:43:59	
J1601-02	S	Total CN	P	1.0947	µg/l	2/19/2018 14:44:00	
J1601-02DUP	S	Total CN	P	1.2592	µg/l	2/19/2018 14:44:01	
J1601-02MS	S	Total CN	P	1.5106	µg/l	2/19/2018 14:44:02	
J1601-02MSD	S	Total CN	P	1.0628	µg/l	2/19/2018 14:44:03	
CCV2	S	Total CN	P	263.3416	µg/l	2/19/2018 14:54:52	
CCB2	S	Total CN	P	0.338	µg/l	2/19/2018 14:54:53	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

2/19/2018 12:33


Test Total CN

Accepted 2/19/2018 12:33

Factor 1122
Bias 0.003

Coeff. of det. 0.999980

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.1759	0.0000	
2	5.0PPBCN	0.008	4.9028	5.0000	
3	10PPBCN	0.012	9.9386	10.0000	
4	50PPBCN	0.048	50.0940	50.0000	
5	100PPBCN	0.091	98.8742	100.0000	
6	250PPBCN	0.227	251.5932	250.0000	
7	500PPBCN	0.448	499.4214	500.0000	