

Daily Analysis Runlog For Sequence/QC Batch ID # LB93475

Review By

SubDirectory

Review On

Test : TOC

LB93475

STD. NAME		STD REF.#				
ICAL Standard		WP61660,WP61661,WP61662,WP61663,WP61664,WP61665,WP61666,WP61667,WP61668				
ICV Standard		WP61669				
CCV Standard		WP61667				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		WP61669				
Chk Standard		WP57902				

Sr#	SampleId	ClientID	QcType	Date	Comment	Status
1	0.0PPM	0.0PPM	CAL1	02/09/18 09:25		OK
2	0.5PPM	0.5PPM	CAL2	02/09/18 09:40		OK
3	1.0PPM	1.0PPM	CAL3	02/09/18 09:53		OK
4	2.0PPM	2.0PPM	CAL4	02/09/18 10:07		OK
5	5.0PPM	5.0PPM	CAL5	02/09/18 10:21		OK
6	10.0PPM	10.0PPM	CAL6	02/09/18 10:36		OK
7	20.0PPM	20.0PPM	CAL7	02/09/18 10:51		OK
8	ICV1	ICV1	ICV	02/09/18 14:51		OK
9	ICB1	ICB1	ICB	02/09/18 15:05		OK
10	CCV1	CCV1	CCV	02/23/18 10:29		OK
11	CCB1	CCB1	CCB	02/23/18 10:58		OK
12	LB93475BLW	LB93475BLW	MB	02/23/18 11:22		OK
13	LB93475BSW	LB93475BSW	LCS	02/23/18 11:50		OK
14	J1632-01	EFF-180220-01	SAM	02/23/18 12:44		OK
15	J1632-01MS	EFF-180220-01MS	MS	02/23/18 13:16		OK
16	J1632-01MSD	EFF-180220-01MSD	MSD	02/23/18 13:47		OK
17	J1161-08	LOQ-WATER-02-QT	LOQ	02/23/18 14:43		OK
18	J1161-07	LOD-MDL-WATER-0	SAM	02/23/18 16:04		OK
19	J1161-09	MDL-WATER-03-QT	SAM	02/23/18 16:55		OK
20	CCV2	CCV2	CCV	02/23/18 18:07		OK
21	CCB2	CCB2	CCB	02/23/18 18:20		OK



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/16/2018 12:00:00 AM
Project:	LB93475	Date Received:	1/16/2018 12:00:00 AM
Client Sample ID:	LOD-MDL-WATER-01-QT1-2018	SDG No.:	LB93475
Lab Sample ID:	J1161-07	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
10-35-5	TOC	0.34	J	1	0.097		0.5	mg/L		02/23/2018	9060



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/16/2018 12:00:00 AM
Project:	LB93475	Date Received:	1/16/2018 12:00:00 AM
Client Sample ID:	MDL-WATER-03-QT1-2018	SDG No.:	LB93475
Lab Sample ID:	J1161-09	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
10-35-5	TOC	0.25	J	1	0.097		0.5	mg/L		02/23/2018	9060



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	2/20/2018 12:00:00 AM
Project:	LB93475	Date Received:	2/21/2018 12:00:00 AM
Client Sample ID:	EFF-180220-01	SDG No.:	LB93475
Lab Sample ID:	J1632-01	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
10-35-5	TOC	0.41	J	1	0.097		0.5	mg/L		02/23/2018	9060



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	2/20/2018 12:00:00 AM
Project:	LB93475	Date Received:	2/21/2018 12:00:00 AM
Client Sample ID:	EFF-180220-01	SDG No.:	LB93475
Lab Sample ID:	J1632-01	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
-----	-----------	-------	------	----	-----	-----	------------	-------	-----------	-----------	----------

Color Before:

Clarity Before:

Texture:

Color After:

Clarity After:

Artifacts:

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB93475

SDG No.: LB93475

Contract: _____ Lab Code: CHEM

Case No.: LB93475

SAS No.: LB93475

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	TOC	9.6	10	96		90 - 110		02/23/2018	10:29	LB93475
CCV2	TOC	10	10	100		90 - 110		02/23/2018	18:07	LB93475



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB93475 **SDG No.:** LB93475
Contract: **Lab Code:** CHEM **Case No.:** LB93475 **SAS No.:** LB93475

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	TOC	0.29	+/-0.5	J		0.5		02/09/2018	15:05	LB93475



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB93475

SDG No.: LB93475

Contract:

Lab Code: CHEM

Case No.: LB93475

SAS No.: LB93475

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	TOC	0.15	+/-0.5	J		0.5		02/23/2018	10:58	LB93475
CCB2	TOC	0.23	+/-0.5	J		0.5		02/23/2018	18:20	LB93475



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB93475

SDG No.: LB93475

Contract: Lab Code: CHEM

Case No.: LB93475

SAS No.: LB93475

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
TOC	mg/L	75 - 125	10.7		0.41	J	10	103		



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB93475 level: low sdg no.: LB93475
contract: lab code: CHEM case no.: LB93475 sas no.: LB93475
matrix: WATER sample id: J1632-01 client id: EFF-180220-01MSD
Percent Solids for Sample: 0 Spiked ID: J1632-01MSD Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
TOC	mg/L	75 - 125	10.4		0.41	J	10	100		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB93475 **Level:** LOW **SDG No.:** LB93475
Contract: **Lab Code:** CHEM **Case No.:** LB93475 **SAS No.:** LB93475
Matrix: WATER **Sample ID:** J1632-01MS **Client ID:** EFF-180220-01MSD
Percent Solids for Sample: 0 **Duplicate ID** J1632-01MSD **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
TOC	mg/L	20	10.7		10.4		3		



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB93475

SDG No.: LB93475

Contract:

Lab Code: CHEM

Case No.: LB93475

SAS No.: LB93475

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
J1161-08 TOC	mg/L	1	1.2		120	50 - 150	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB93475

SDG No.: LB93475

Contract:

Lab Code: CHEM

Case No.: LB93475

SAS No.: LB93475

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
LB93475BSW TOC	mg/L	10	10		100	90 - 110	

TOC Raw Data

Sequence ID : LB93475

Method ID : 9060

LabSampleID	Response	Result(ppm)	%RSD	Anal DateTime	Sample Vol/Wt
0.0PPM	17741	0.0PPM		2/9/2018 9:25:00 AM	0.500
		0.0PPM			
0.5PPM	73879	0.5PPM		2/9/2018 9:40:00 AM	0.5
		0.5PPM			
1.0PPM	132822	1.0PPM		2/9/2018 9:53:00 AM	0.5
		1.0PPM			
2.0PPM	252602	2.0PPM		2/9/2018 10:07:00 AM	0.5
		2.0PPM			
5.0PPM	581752	5.0PPM		2/9/2018 10:21:00 AM	0.5
		5.0PPM			
10.0PPM	1201476	10.0PPM		2/9/2018 10:36:00 AM	0.5
		10.0PPM			
20.0PPM	2356976	20.0PPM		2/9/2018 10:51:00 AM	0.5
		20.0PPM			
ICV1	1145690	9.7716		2/9/2018 2:51:00 PM	0.500
ICB1	33422	0.2851		2/9/2018 3:05:00 PM	0.500
CCV1	1129047	9.6297		2/23/2018 10:29:00 AM	0.5
CCB1	17298	0.1475		2/23/2018 10:58:00 AM	0.5
LB93475BLW	1257	0.0107		2/23/2018 11:11:00 AM	0.5
LB93475BLW	9120	0.0778		2/23/2018 11:15:00 AM	0.5
LB93475BLW	8602	0.0734		2/23/2018 11:18:00 AM	0.5
LB93475BLW	18938	0.1615		2/23/2018 11:22:00 AM	0.5
		0.08085	76.5463		
LB93475BSW	1207140	10.2957		2/23/2018 11:37:00 AM	0.5
LB93475BSW	1152087	9.8262		2/23/2018 11:42:00 AM	0.5
LB93475BSW	1165029	9.9365		2/23/2018 11:46:00 AM	0.5
LB93475BSW	1151627	9.8222		2/23/2018 11:50:00 AM	0.5
		9.97015	2.2407		
J1632-01	53303	0.4546		2/23/2018 12:34:00 PM	0.5
J1632-01	39849	0.3399		2/23/2018 12:37:00 PM	0.5
J1632-01	47927	0.4088		2/23/2018 12:41:00 PM	0.5
J1632-01	50236	0.4285		2/23/2018 12:44:00 PM	0.5
		0.40795	12.0338		
J1632-01MS	1206238	10.288		2/23/2018 1:00:00 PM	0.5
J1632-01MS	1231198	10.5009		2/23/2018 1:05:00 PM	0.5
J1632-01MS	1334760	11.3842		2/23/2018 1:11:00 PM	0.5
J1632-01MS	1234351	10.5278		2/23/2018 1:16:00 PM	0.5
		10.6752	4.5401		
J1632-01MSD	1301924	11.1041		2/23/2018 1:33:00 PM	0.5
J1632-01MSD	1193410	10.1786		2/23/2018 1:37:00 PM	0.5
J1632-01MSD	1157205	9.8698		2/23/2018 1:42:00 PM	0.5
J1632-01MSD	1239966	10.5757		2/23/2018 1:47:00 PM	0.5
		10.432	5.1104		
J1161-08	147850	1.261		2/23/2018 2:30:00 PM	0.5



TOC Raw Data

Sequence ID : LB93475

Method ID : 9060

LabSampleID	Response	Result(ppm)	%RSD	Anal DateTime	Sample Vol/Wt
J1161-08	137068	1.1691		2/23/2018 2:34:00 PM	0.5
J1161-08	174476	1.4881		2/23/2018 2:39:00 PM	0.5
J1161-08	115294	0.9833		2/23/2018 2:43:00 PM	0.5
		1.22537	17.1221		
J1161-07	67671	0.5772		2/23/2018 3:53:00 PM	0.5
J1161-07	34115	0.291		2/23/2018 3:57:00 PM	0.5
J1161-07	34673	0.2957		2/23/2018 4:01:00 PM	0.5
J1161-07	22666	0.1933		2/23/2018 4:04:00 PM	0.5
		0.3393	48.7696		
J1161-09	33303	0.284		2/23/2018 4:45:00 PM	0.5
J1161-09	21304	0.1817		2/23/2018 4:48:00 PM	0.5
J1161-09	30396	0.2592		2/23/2018 4:52:00 PM	0.5
J1161-09	32724	0.2791		2/23/2018 4:55:00 PM	0.5
		0.251	18.8957		
CCV2	1167220	9.9552		2/23/2018 6:07:00 PM	0.5
CCB2	27276	0.2326		2/23/2018 6:20:00 PM	0.5

Daily Analysis Runlog For Sequence/QC Batch ID # LB93475

Review By

SubDirectory

Review On

Test : TOC

LB93475

STD. NAME		STD REF.#				
ICAL Standard		WP61660,WP61661,WP61662,WP61663,WP61664,WP61665,WP61666,WP61667,WP61668				
ICV Standard		WP61669				
CCV Standard		WP61667				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		WP61669				
Chk Standard		WP57902				

Sr#	SampleID	ClientID	QcType	Date	Comment	Status
1	0.0PPM	0.0PPM	CAL1	02/09/18 09:25		OK
2	0.5PPM	0.5PPM	CAL2	02/09/18 09:40		OK
3	1.0PPM	1.0PPM	CAL3	02/09/18 09:53		OK
4	2.0PPM	2.0PPM	CAL4	02/09/18 10:07		OK
5	5.0PPM	5.0PPM	CAL5	02/09/18 10:21		OK
6	10.0PPM	10.0PPM	CAL6	02/09/18 10:36		OK
7	20.0PPM	20.0PPM	CAL7	02/09/18 10:51		OK
8	ICV1	ICV1	ICV	02/09/18 14:51		OK
9	ICB1	ICB1	ICB	02/09/18 15:05		OK
10	CCV1	CCV1	CCV	02/23/18 10:29		OK
11	CCB1	CCB1	CCB	02/23/18 10:58		OK
12	LB93475BLW	LB93475BLW	MB	02/23/18 11:22		OK
13	LB93475BSW	LB93475BSW	LCS	02/23/18 11:50		OK
14	J1632-01	EFF-180220-01	SAM	02/23/18 12:44		OK
15	J1632-01MS	EFF-180220-01MS	MS	02/23/18 13:16		OK
16	J1632-01MSD	EFF-180220-01MSD	MSD	02/23/18 13:47		OK
17	J1161-08	LOQ-WATER-02-QT	LOQ	02/23/18 14:43		OK
18	J1161-07	LOD-MDL-WATER-0	SAM	02/23/18 16:04		OK
19	J1161-09	MDL-WATER-03-QT	SAM	02/23/18 16:55		OK
20	CCV2	CCV2	CCV	02/23/18 18:07		OK
21	CCB2	CCB2	CCB	02/23/18 18:20		OK



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/16/2018 12:00:00 AM
Project:	LB93475	Date Received:	1/16/2018 12:00:00 AM
Client Sample ID:	LOD-MDL-WATER-01-QT1-2018	SDG No.:	LB93475
Lab Sample ID:	J1161-07	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
10-35-5	TOC	0.34	J	1	0.097		0.5	mg/L		02/23/2018	9060



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/16/2018 12:00:00 AM
Project:	LB93475	Date Received:	1/16/2018 12:00:00 AM
Client Sample ID:	MDL-WATER-03-QT1-2018	SDG No.:	LB93475
Lab Sample ID:	J1161-09	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
10-35-5	TOC	0.25	J	1	0.097		0.5	mg/L		02/23/2018	9060



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	2/20/2018 12:00:00 AM
Project:	LB93475	Date Received:	2/21/2018 12:00:00 AM
Client Sample ID:	EFF-180220-01	SDG No.:	LB93475
Lab Sample ID:	J1632-01	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
10-35-5	TOC	0.41	J	1	0.097		0.5	mg/L		02/23/2018	9060



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	2/20/2018 12:00:00 AM
Project:	LB93475	Date Received:	2/21/2018 12:00:00 AM
Client Sample ID:	EFF-180220-01	SDG No.:	LB93475
Lab Sample ID:	J1632-01	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
-----	-----------	-------	------	----	-----	-----	------------	-------	-----------	-----------	----------

Color Before:

Clarity Before:

Texture:

Color After:

Clarity After:

Artifacts:

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB93475

SDG No.: LB93475

Contract: _____

Lab Code: CHEM

Case No.: LB93475

SAS No.: LB93475

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV1	TOC	9.8	10	98		90 - 110		02/09/2018	14:51	LB93475



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB93475

SDG No.: LB93475

Contract: _____ Lab Code: CHEM

Case No.: LB93475

SAS No.: LB93475

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	TOC	9.6	10	96		90 - 110		02/23/2018	10:29	LB93475
CCV2	TOC	10	10	100		90 - 110		02/23/2018	18:07	LB93475



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB93475

SDG No.: LB93475

Contract:

Lab Code: CHEM

Case No.: LB93475

SAS No.: LB93475

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	TOC	0.29	+/-0.5	J			0.5	02/09/2018	15:05	LB93475



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB93475

SDG No.: LB93475

Contract:

Lab Code: CHEM

Case No.: LB93475

SAS No.: LB93475

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	TOC	0.15	+/-0.5	J		0.5		02/23/2018	10:58	LB93475
CCB2	TOC	0.23	+/-0.5	J		0.5		02/23/2018	18:20	LB93475



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB93475

SDG No.: LB93475

Contract: Lab Code: CHEM

Case No.: LB93475

SAS No.: LB93475

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB93475 level: low sdg no.: LB93475
contract: lab code: CHEM case no.: LB93475 sas no.: LB93475
matrix: WATER sample id: J1632-01 client id: EFF-180220-01MS
Percent Solids for Sample: 0 Spiked ID: J1632-01MS Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
TOC	mg/L	75 - 125	10.7		0.41	J	10	103		



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB93475 level: low sdg no.: LB93475
contract: lab code: CHEM case no.: LB93475 sas no.: LB93475
matrix: WATER sample id: J1632-01 client id: EFF-180220-01MSD
Percent Solids for Sample: 0 Spiked ID: J1632-01MSD Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
TOC	mg/L	75 - 125	10.4		0.41	J	10	100		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB93475 **Level:** LOW **SDG No.:** LB93475
Contract: **Lab Code:** CHEM **Case No.:** LB93475 **SAS No.:** LB93475
Matrix: WATER **Sample ID:** J1632-01MS **Client ID:** EFF-180220-01MSD
Percent Solids for Sample: 0 **Duplicate ID** J1632-01MSD **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
TOC	mg/L	20	10.7		10.4		3		

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
J1161-08							
TOC	mg/L	1	1.2		120	50 - 150	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB93475

SDG No.: LB93475

Contract:

Lab Code: CHEM

Case No.: LB93475

SAS No.: LB93475

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
LB93475BSW TOC	mg/L	10	10		100	90 - 110	