

LB93797

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

3/9/2018 13:58

Reactive
Test: Total CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.718	0.0	0.095	
ICB1	-0.372	0.0	0.004	
CCV1	242.638	0.0	0.232	
CCB1	-0.317	0.0	0.004	
PB107205BL	-0.696	0.0	0.003	
J1755-02	-0.566	0.0	0.004	
J1755-02DUP	-0.667	0.0	0.003	
J1853-02	-0.495	0.0	0.004	
J1855-04	-0.660	0.0	0.003	
J1855-10	-0.771	0.0	0.003	
J1855-14	-0.642	0.0	0.003	
J1855-07	-0.537	0.0	0.004	
CCV2	256.942	0.0	0.245	
CCB2	-0.511	0.0	0.004	

N	14
Mean	42.147
SD	91.7273
CV%	217.63

AS

Aquakem v. 7.2AQ1

Results from time period:

Fri Mar 09 13:25:21 2018

Fri Mar 09 13:54:47 2018

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.9215	µg/l	3/9/2018 9:50:05	
5.OPPBCN	A	Reactive CN	P	4.0491	µg/l	3/9/2018 9:50:06	
10PPBCN	A	Reactive CN	P	9.5887	µg/l	3/9/2018 9:50:07	
50PPBCN	A	Reactive CN	P	49.8315	µg/l	3/9/2018 9:50:08	
100PPBCN	A	Reactive CN	P	100.8181	µg/l	3/9/2018 9:50:09	
250PPBCN	A	Reactive CN	P	253.5263	µg/l	3/9/2018 9:50:10	
500PPBCN	A	Reactive CN	P	498.1079	µg/l	3/9/2018 9:50:11	
ICV1	S	Reactive CN	P	96.7183	µg/l	3/9/2018 13:25:21	
ICB1	S	Reactive CN	P	-0.3725	µg/l	3/9/2018 13:25:22	
CCV1	S	Reactive CN	P	242.6379	µg/l	3/9/2018 13:25:23	
CCB1	S	Reactive CN	P	-0.317	µg/l	3/9/2018 13:25:24	
PB107205BL	S	Reactive CN	P	-0.6959	µg/l	3/9/2018 13:25:25	
J1755-02	S	Reactive CN	P	-0.566	µg/l	3/9/2018 13:25:26	
J1755-02DUP	S	Reactive CN	P	-0.6669	µg/l	3/9/2018 13:25:27	
J1853-02	S	Reactive CN	P	-0.4949	µg/l	3/9/2018 13:25:28	
J1855-04	S	Reactive CN	P	-0.6601	µg/l	3/9/2018 13:25:29	
J1855-10	S	Reactive CN	P	-0.7707	µg/l	3/9/2018 13:25:31	
J1855-14	S	Reactive CN	P	-0.6417	µg/l	3/9/2018 13:29:12	
J1855-07	S	Reactive CN	P	-0.5371	µg/l	3/9/2018 13:54:45	
CCV2	S	Reactive CN	P	256.9417	µg/l	3/9/2018 13:54:46	
CCB2	S	Reactive CN	P	-0.5108	µg/l	3/9/2018 13:54:47	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

3/9/2018 9:56


Test Total CN

Accepted 3/9/2018 9:56

Factor 1067
Bias 0.004

Coeff. of det. 0.999909

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.9215	0.0000	
2	5.0PPBCN	0.008	4.0491	5.0000	
3	10PPBCN	0.013	9.5887	10.0000	
4	50PPBCN	0.051	49.8315	50.0000	
5	100PPBCN	0.099	100.8181	100.0000	
6	250PPBCN	0.242	253.5263	250.0000	
7	500PPBCN	0.471	498.1079	500.0000	