

=====

Test results

Aquakem 7.2AQ1

2094009
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

3/19/2018 14:51

Test: Reactive
~~Total~~ CN Mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.571	0.0	0.097	
ICB1	-0.229	0.0	0.004	
CCV1	247.007	0.0	0.240	
CCB1	-0.208	0.0	0.004	
PB107453BL	-0.318	0.0	0.004	
J1960-02	-0.308	0.0	0.004	
J1960-02DUP	-0.289	0.0	0.004	
J1966-08	-0.296	0.0	0.004	
J1966-10	-0.272	0.0	0.004	
PB107454BL	-0.318	0.0	0.004	
J1755-04	-0.215	0.0	0.004	
J1755-04DUP	-0.157	0.0	0.004	
J1963-04	-0.351	0.0	0.004	
CCV2	260.636	0.0	0.253	
CCB2	-0.299	0.0	0.004	

N	15
Mean	40.130
SD	90.3595
CV%	225.17

AS

Aquakem v. 7.2AQ1

Results from time period:

Mon Mar 19 14:41:57 2018

Mon Mar 19 14:46:19 2018

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.1103	µg/l	3/19/2018 12:21:42	
5.OPPBCN	A	Reactive CN	P	4.5038	µg/l	3/19/2018 12:21:43	
10PPBCN	A	Reactive CN	P	9.7701	µg/l	3/19/2018 12:21:44	
50PPBCN	A	Reactive CN	P	49.8971	µg/l	3/19/2018 12:21:45	
100PPBCN	A	Reactive CN	P	100.2507	µg/l	3/19/2018 12:21:46	
250PPBCN	A	Reactive CN	P	251.4379	µg/l	3/19/2018 12:21:47	
500PPBCN	A	Reactive CN	P	499.2507	µg/l	3/19/2018 12:21:48	
ICV1	S	Reactive CN	P	97.5711	µg/l	3/19/2018 14:41:57	
ICB1	S	Reactive CN	P	-0.2293	µg/l	3/19/2018 14:41:58	
CCV1	S	Reactive CN	P	247.0065	µg/l	3/19/2018 14:41:59	
CCB1	S	Reactive CN	P	-0.208	µg/l	3/19/2018 14:42:00	
PB107453BL	S	Reactive CN	P	-0.3183	µg/l	3/19/2018 14:42:01	
J1960-02	S	Reactive CN	P	-0.3084	µg/l	3/19/2018 14:42:02	
J1960-02DUP	S	Reactive CN	P	-0.2892	µg/l	3/19/2018 14:42:03	
J1966-08	S	Reactive CN	P	-0.2956	µg/l	3/19/2018 14:42:04	
J1966-10	S	Reactive CN	P	-0.272	µg/l	3/19/2018 14:42:05	
PB107454BL	S	Reactive CN	P	-0.3182	µg/l	3/19/2018 14:42:06	
J1755-04	S	Reactive CN	P	-0.2148	µg/l	3/19/2018 14:42:07	
J1755-04DUP	S	Reactive CN	P	-0.1567	µg/l	3/19/2018 14:46:16	
J1963-04	S	Reactive CN	P	-0.351	µg/l	3/19/2018 14:46:17	
CCV2	S	Reactive CN	P	260.6356	µg/l	3/19/2018 14:46:18	
CCB2	S	Reactive CN	P	-0.2989	µg/l	3/19/2018 14:46:19	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

3/19/2018 12:22


Test Total CN

Accepted 3/19/2018 12:22

Factor 1045
Bias 0.004

Coeff. of det. 0.999985

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.1103	0.0000	
2	5.0PPBCN	0.008	4.5038	5.0000	
3	10PPBCN	0.013	9.7701	10.0000	
4	50PPBCN	0.052	49.8971	50.0000	
5	100PPBCN	0.100	100.2507	100.0000	
6	250PPBCN	0.245	251.4379	250.0000	
7	500PPBCN	0.482	499.2507	500.0000	