

LB94695

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

3/23/2018 11:08

Test: ^{Reactive} ~~Total~~ CN Mg/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	93.955	0.0	0.093	
ICB1	0.504	0.0	0.004	
CCV1	239.856	0.0	0.232	
CCB1	0.716	0.0	0.004	
PB107594BL	0.302	0.0	0.004	
J1748-08	0.375	0.0	0.004	
J1748-08DUP	0.292	0.0	0.004	
J1748-10	0.339	0.0	0.004	
J1748-12	0.217	0.0	0.004	
J1748-14	0.338	0.0	0.004	
CCV2	239.528	0.0	0.232	
CCB2	0.048	0.0	0.003	

N	12
Mean	48.039
SD	93.4395
CV%	194.51

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

3/23/2018 7:38


Test Total CN

Accepted 3/23/2018 7:38

Factor 1050
 Bias 0.003

Coeff. of det. 0.999982

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.2596	0.0000	
2	5.0PPBCN	0.008	5.1040	5.0000	
3	10PPBCN	0.013	10.2687	10.0000	
4	50PPBCN	0.051	50.4093	50.0000	
5	100PPBCN	0.098	99.7890	100.0000	
6	250PPBCN	0.240	248.3490	250.0000	
7	500PPBCN	0.480	500.8203	500.0000	

AS

Aquakem v. 7.2AQ1

Results from time period:

Fri Mar 23 11:05:26 2018

Fri Mar 23 11:08:14 2018

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	0.2596	µg/l	3/23/2018 7:36:49	
5.0PPBCN	A	Reactive CN	P	5.104	µg/l	3/23/2018 7:36:50	
10PPBCN	A	Reactive CN	P	10.2687	µg/l	3/23/2018 7:36:51	
50PPBCN	A	Reactive CN	P	50.4093	µg/l	3/23/2018 7:36:52	
100PPBCN	A	Reactive CN	P	99.789	µg/l	3/23/2018 7:36:53	
250PPBCN	A	Reactive CN	P	248.349	µg/l	3/23/2018 7:36:54	
500PPBCN	A	Reactive CN	P	500.8203	µg/l	3/23/2018 7:36:55	
ICV1	S	Reactive CN	P	93.9552	µg/l	3/23/2018 11:05:26	
ICB1	S	Reactive CN	P	0.5037	µg/l	3/23/2018 11:05:27	
CCV1	S	Reactive CN	P	239.8563	µg/l	3/23/2018 11:05:28	
CCB1	S	Reactive CN	P	0.7159	µg/l	3/23/2018 11:05:29	
PB107594BL	S	Reactive CN	P	0.3016	µg/l	3/23/2018 11:05:30	
J1748-08	S	Reactive CN	P	0.3746	µg/l	3/23/2018 11:05:31	
J1748-08DUP	S	Reactive CN	P	0.2922	µg/l	3/23/2018 11:05:32	
J1748-10	S	Reactive CN	P	0.3385	µg/l	3/23/2018 11:05:33	
J1748-12	S	Reactive CN	P	0.217	µg/l	3/23/2018 11:05:34	
J1748-14	S	Reactive CN	P	0.3382	µg/l	3/23/2018 11:05:35	
CCV2	S	Reactive CN	P	239.5276	µg/l	3/23/2018 11:05:36	
CCB2	S	Reactive CN	P	0.0481	µg/l	3/23/2018 11:08:14	