

## Daily Analysis Runlog For Sequence/QC Batch ID # LB94301

Review By

Review On

SubDirectory

LB94301

Test :

Cyanide

| STD. NAME     |  | STD REF.#                                               |  |  |  |  |
|---------------|--|---------------------------------------------------------|--|--|--|--|
| ICAL Standard |  | WP63245,WP63246,WP63247,WP63248,WP63249,WP63250,WP63251 |  |  |  |  |
| ICV Standard  |  | W1542                                                   |  |  |  |  |
| CCV Standard  |  | WP63252                                                 |  |  |  |  |
| ICSA Standard |  | N/A                                                     |  |  |  |  |
| CRI Standard  |  | N/A                                                     |  |  |  |  |
| LCS Standard  |  | WP62802                                                 |  |  |  |  |
| Chk Standard  |  | WP63254,WP59958,WP60878                                 |  |  |  |  |

| Sr# | SampleID    | ClientID      | QcType | Date           | Comment | Status |
|-----|-------------|---------------|--------|----------------|---------|--------|
| 1   | 0.0PPBCN    | 0.0PPBCN      | CAL1   | 04/02/18 09:04 |         | OK     |
| 2   | 5.0PPBCN    | 5.0PPBCN      | CAL2   | 04/02/18 09:04 |         | OK     |
| 3   | 10PPBCN     | 10PPBCN       | CAL3   | 04/02/18 09:04 |         | OK     |
| 4   | 50PPBCN     | 50PPBCN       | CAL4   | 04/02/18 09:04 |         | OK     |
| 5   | 100PPBCN    | 100PPBCN      | CAL5   | 04/02/18 09:04 |         | OK     |
| 6   | 250PPBCN    | 250PPBCN      | CAL6   | 04/02/18 09:04 |         | OK     |
| 7   | 500PPBCN    | 500PPBCN      | CAL7   | 04/02/18 09:04 |         | OK     |
| 8   | ICV1        | ICV1          | ICV    | 04/02/18 10:27 |         | OK     |
| 9   | ICB1        | ICB1          | ICB    | 04/02/18 10:27 |         | OK     |
| 10  | CCV1        | CCV1          | CCV    | 04/02/18 10:27 |         | OK     |
| 11  | CCB1        | CCB1          | CCB    | 04/02/18 10:27 |         | OK     |
| 12  | PB107847BL  | PB107847BL    | MB     | 04/02/18 10:27 |         | OK     |
| 13  | PB107847BS  | PB107847BS    | LCS    | 04/02/18 10:27 |         | OK     |
| 14  | J2142-01    | GB-RT-2334    | SAM    | 04/02/18 10:27 |         | OK     |
| 15  | J2142-01DUP | GB-RT-2334DUP | DUP    | 04/02/18 10:27 |         | OK     |
| 16  | J2142-01MS  | GB-RT-2334MS  | MS     | 04/02/18 10:27 |         | OK     |
| 17  | J2142-01MSD | GB-RT-2334MSD | MSD    | 04/02/18 10:35 |         | OK     |
| 18  | J2142-05    | GB-RT-3025    | SAM    | 04/02/18 10:35 |         | OK     |
| 19  | CCV2        | CCV2          | CCV    | 04/02/18 10:35 |         | OK     |
| 20  | CCB2        | CCB2          | CCB    | 04/02/18 10:35 |         | OK     |
| 21  | J2142-07    | GB-RT-4252    | SAM    | 04/02/18 10:35 |         | OK     |

284 Sheffield Street, Mountainside, New Jersey 07092 Phone : 908 789 8900 Fax : 908 789 8922

**Daily Analysis Runlog For Sequence/QC Batch ID # LB94301**

|               |              |                                                         |     |                |  |    |
|---------------|--------------|---------------------------------------------------------|-----|----------------|--|----|
| Review By     |              | Review On                                               |     |                |  |    |
| SubDirectory  |              | LB94301                                                 |     | Test : Cyanide |  |    |
| STD. NAME     |              | STD REF.#                                               |     |                |  |    |
| ICAL Standard |              | WP63245,WP63246,WP63247,WP63248,WP63249,WP63250,WP63251 |     |                |  |    |
| ICV Standard  |              | W1542                                                   |     |                |  |    |
| CCV Standard  |              | WP63252                                                 |     |                |  |    |
| ICSA Standard |              | N/A                                                     |     |                |  |    |
| CRI Standard  |              | N/A                                                     |     |                |  |    |
| LCS Standard  |              | WP62802                                                 |     |                |  |    |
| Chk Standard  |              | WP63254,WP59958,WP60878                                 |     |                |  |    |
| 22            | J2142-09     | GB-RT-2342                                              | SAM | 04/02/18 10:35 |  | OK |
| 23            | J2146-03     | TP-CD                                                   | SAM | 04/02/18 10:35 |  | OK |
| 24            | J2146-01     | TP-CS                                                   | SAM | 04/02/18 10:35 |  | OK |
| 25            | J2142-03     | GB-RT-4632                                              | SAM | 04/02/18 11:14 |  | OK |
| 26            | LOWPB107847  | LOWPB107847                                             | SAM | 04/02/18 11:30 |  | OK |
| 27            | HIGHPB107847 | HIGHPB107847                                            | SAM | 04/02/18 11:31 |  | OK |
| 28            | CCV3         | CCV3                                                    | CCV | 04/02/18 11:31 |  | OK |
| 29            | CCB3         | CCB3                                                    | CCB | 04/02/18 11:31 |  | OK |



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

## Report of Analysis

|                   |              |                 |         |
|-------------------|--------------|-----------------|---------|
| Client:           |              | Date Collected: |         |
| Project:          | LB94301      | Date Received:  |         |
| Client Sample ID: | HIGHPB107847 | SDG No.:        | LB94301 |
| Lab Sample ID:    | HIGHPB107847 | Matrix:         | WATER   |
| Level (low/med):  | low          | % Solid:        | 0       |

| Cas     | Parameter | Conc. | Qua. | DF | MDL   | LOD | LOQ / CRQL | Units | Prep Date | Date Ana.  | Ana Met. |
|---------|-----------|-------|------|----|-------|-----|------------|-------|-----------|------------|----------|
| 57-12-5 | Cyanide   | 0.51  | OR   | 1  | 0.003 |     | 0.005      | mg/L  |           | 04/02/2018 | 9012B    |



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

## Report of Analysis

|                   |            |                 |                       |
|-------------------|------------|-----------------|-----------------------|
| Client:           |            | Date Collected: | 3/29/2018 12:00:00 AM |
| Project:          | LB94301    | Date Received:  | 3/29/2018 12:00:00 AM |
| Client Sample ID: | GB-RT-2334 | SDG No.:        | LB94301               |
| Lab Sample ID:    | J2142-01   | Matrix:         | Solid                 |
| Level (low/med):  | low        | % Solid:        | 81.8                  |

| Cas     | Parameter | Conc. | Qua. | DF | MDL  | LOD | LOQ / CRQL | Units | Prep Date  | Date Ana.  | Ana Met. |
|---------|-----------|-------|------|----|------|-----|------------|-------|------------|------------|----------|
| 57-12-5 | Cyanide   | 0.17  | J    | 1  | 0.04 |     | 0.3        | mg/Kg | 03/30/2018 | 04/02/2018 | 9012B    |



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

## Report of Analysis

|                   |            |                 |                       |
|-------------------|------------|-----------------|-----------------------|
| Client:           |            | Date Collected: | 3/29/2018 12:00:00 AM |
| Project:          | LB94301    | Date Received:  | 3/29/2018 12:00:00 AM |
| Client Sample ID: | GB-RT-4632 | SDG No.:        | LB94301               |
| Lab Sample ID:    | J2142-03   | Matrix:         | Solid                 |
| Level (low/med):  | low        | % Solid:        | 83.7                  |

| Cas     | Parameter | Conc. | Qua. | DF | MDL   | LOD | LOQ / CRQL | Units | Prep Date  | Date Ana.  | Ana Met. |
|---------|-----------|-------|------|----|-------|-----|------------|-------|------------|------------|----------|
| 57-12-5 | Cyanide   | 0.082 | J    | 1  | 0.039 |     | 0.3        | mg/Kg | 03/30/2018 | 04/02/2018 | 9012B    |



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

## Report of Analysis

|                   |            |                 |                       |
|-------------------|------------|-----------------|-----------------------|
| Client:           |            | Date Collected: | 3/29/2018 12:00:00 AM |
| Project:          | LB94301    | Date Received:  | 3/29/2018 12:00:00 AM |
| Client Sample ID: | GB-RT-3025 | SDG No.:        | LB94301               |
| Lab Sample ID:    | J2142-05   | Matrix:         | Solid                 |
| Level (low/med):  | low        | % Solid:        | 76.9                  |

| Cas     | Parameter | Conc. | Qua. | DF | MDL   | LOD | LOQ / CRQL | Units | Prep Date  | Date Ana.  | Ana Met. |
|---------|-----------|-------|------|----|-------|-----|------------|-------|------------|------------|----------|
| 57-12-5 | Cyanide   | 0.15  | J    | 1  | 0.042 |     | 0.32       | mg/Kg | 03/30/2018 | 04/02/2018 | 9012B    |



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

## Report of Analysis

|                   |            |                 |                       |
|-------------------|------------|-----------------|-----------------------|
| Client:           |            | Date Collected: | 3/29/2018 12:00:00 AM |
| Project:          | LB94301    | Date Received:  | 3/29/2018 12:00:00 AM |
| Client Sample ID: | GB-RT-4252 | SDG No.:        | LB94301               |
| Lab Sample ID:    | J2142-07   | Matrix:         | Solid                 |
| Level (low/med):  | low        | % Solid:        | 84.9                  |

| Cas     | Parameter | Conc. | Qua. | DF | MDL   | LOD | LOQ / CRQL | Units | Prep Date  | Date Ana.  | Ana Met. |
|---------|-----------|-------|------|----|-------|-----|------------|-------|------------|------------|----------|
| 57-12-5 | Cyanide   | 0.083 | J    | 1  | 0.037 |     | 0.28       | mg/Kg | 03/30/2018 | 04/02/2018 | 9012B    |



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

## Report of Analysis

|                   |            |                 |                       |
|-------------------|------------|-----------------|-----------------------|
| Client:           |            | Date Collected: | 3/29/2018 12:00:00 AM |
| Project:          | LB94301    | Date Received:  | 3/29/2018 12:00:00 AM |
| Client Sample ID: | GB-RT-2342 | SDG No.:        | LB94301               |
| Lab Sample ID:    | J2142-09   | Matrix:         | Solid                 |
| Level (low/med):  | low        | % Solid:        | 84.2                  |

| Cas     | Parameter | Conc. | Qua. | DF | MDL   | LOD | LOQ / CRQL | Units | Prep Date  | Date Ana.  | Ana Met. |
|---------|-----------|-------|------|----|-------|-----|------------|-------|------------|------------|----------|
| 57-12-5 | Cyanide   | 0.073 | J    | 1  | 0.037 |     | 0.28       | mg/Kg | 03/30/2018 | 04/02/2018 | 9012B    |



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

## Report of Analysis

|                   |          |                 |                       |
|-------------------|----------|-----------------|-----------------------|
| Client:           |          | Date Collected: | 3/29/2018 12:00:00 AM |
| Project:          | LB94301  | Date Received:  | 3/29/2018 12:00:00 AM |
| Client Sample ID: | TP-CS    | SDG No.:        | LB94301               |
| Lab Sample ID:    | J2146-01 | Matrix:         | Solid                 |
| Level (low/med):  | low      | % Solid:        | 93.2                  |

| Cas     | Parameter | Conc. | Qua. | DF | MDL   | LOD | LOQ / CRQL | Units | Prep Date  | Date Ana.  | Ana Met. |
|---------|-----------|-------|------|----|-------|-----|------------|-------|------------|------------|----------|
| 57-12-5 | Cyanide   | 0.032 | U    | 1  | 0.034 |     | 0.26       | mg/Kg | 03/30/2018 | 04/02/2018 | 9012B    |



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

## Report of Analysis

|                   |          |                 |                       |
|-------------------|----------|-----------------|-----------------------|
| Client:           |          | Date Collected: | 3/29/2018 12:00:00 AM |
| Project:          | LB94301  | Date Received:  | 3/29/2018 12:00:00 AM |
| Client Sample ID: | TP-CD    | SDG No.:        | LB94301               |
| Lab Sample ID:    | J2146-03 | Matrix:         | Solid                 |
| Level (low/med):  | low      | % Solid:        | 94.1                  |

| Cas     | Parameter | Conc. | Qua. | DF | MDL   | LOD | LOQ / CRQL | Units | Prep Date  | Date Ana.  | Ana Met. |
|---------|-----------|-------|------|----|-------|-----|------------|-------|------------|------------|----------|
| 57-12-5 | Cyanide   | 0.029 | U    | 1  | 0.035 |     | 0.27       | mg/Kg | 03/30/2018 | 04/02/2018 | 9012B    |



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

## Report of Analysis

|                   |             |                 |         |
|-------------------|-------------|-----------------|---------|
| Client:           |             | Date Collected: |         |
| Project:          | LB94301     | Date Received:  |         |
| Client Sample ID: | LOWPB107847 | SDG No.:        | LB94301 |
| Lab Sample ID:    | LOWPB107847 | Matrix:         | WATER   |
| Level (low/med):  | low         | % Solid:        | 0       |

| Cas     | Parameter | Conc. | Qua. | DF | MDL   | LOD | LOQ / CRQL | Units | Prep Date | Date Ana.  | Ana Met. |
|---------|-----------|-------|------|----|-------|-----|------------|-------|-----------|------------|----------|
| 57-12-5 | Cyanide   | 0.011 |      | 1  | 0.003 |     | 0.005      | mg/L  |           | 04/02/2018 | 9012B    |



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

## Report of Analysis

|                   |             |                 |         |
|-------------------|-------------|-----------------|---------|
| Client:           |             | Date Collected: |         |
| Project:          | LB94301     | Date Received:  |         |
| Client Sample ID: | LOWPB107847 | SDG No.:        | LB94301 |
| Lab Sample ID:    | LOWPB107847 | Matrix:         | WATER   |
| Level (low/med):  | low         | % Solid:        | 0       |

| Cas | Parameter | Conc. | Qua. | DF | MDL | LOD | LOQ / CRQL | Units | Prep Date | Date Ana. | Ana Met. |
|-----|-----------|-------|------|----|-----|-----|------------|-------|-----------|-----------|----------|
|-----|-----------|-------|------|----|-----|-----|------------|-------|-----------|-----------|----------|

|               |                 |            |
|---------------|-----------------|------------|
| Color Before: | Clarity Before: | Texture:   |
| Color After:  | Clarity After:  | Artifacts: |
| Comments:     |                 |            |

U = Not Detected  
LOQ = Limit of Quantitation  
MDL = Method Detection Limit  
LOD = Limit of Detection  
D = Dilution  
Q = indicates LCS control criteria did not meet requirements

J = Estimated Value  
B = Analyte Found in Associated Method Blank  
\* = indicates the duplicate analysis is not within control limits.  
E = Indicates the reported value is estimated because of the presence of interference.  
OR = Over Range  
N = Spiked sample recovery not within control limits



## GENCHEM

- 2a -

### INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB94301

SDG No.: LB94301

Contract: \_\_\_\_\_

Lab Code: CHEM

Case No.: LB94301

SAS No.: LB94301

Initial Calibration Source: \_\_\_\_\_

Continuing Calibration Source: \_\_\_\_\_

---

| Sample ID | Analyte | Result<br>mg/L | True Value | %<br>Recovery | Qual | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|------|---------------------------|---|------------------|------------------|---------------|
| ICV1      | Cyanide | 0.095          | 0.099      | 96            |      | 85 - 115                  |   | 04/02/2018       | 10:27            | LB94301       |

---



## GENCHEM

- 2a -

### INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB94301

SDG No.: LB94301

Contract: \_\_\_\_\_ Lab Code: CHEM

Case No.: LB94301

SAS No.: LB94301

Initial Calibration Source: \_\_\_\_\_

Continuing Calibration Source: \_\_\_\_\_

| Sample ID | Analyte | Result<br>mg/L | True Value | %<br>Recovery | Qual | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|------|---------------------------|---|------------------|------------------|---------------|
| CCV1      | Cyanide | 0.24           | 0.25       | 96            |      | 90 - 110                  |   | 04/02/2018       | 10:27            | LB94301       |
| CCV2      | Cyanide | 0.24           | 0.25       | 96            |      | 90 - 110                  |   | 04/02/2018       | 10:35            | LB94301       |
| CCV3      | Cyanide | 0.26           | 0.25       | 104           |      | 90 - 110                  |   | 04/02/2018       | 11:31            | LB94301       |



**GENCHEM**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** LB94301

**SDG No.:** LB94301

**Contract:**                     

**Lab Code:** CHEM

**Case No.:** LB94301

**SAS No.:** LB94301

| Sample ID | Analyte | Result<br>mg/L | Acceptance<br>Limit | Conc<br>Qual | LOD | CRQL  | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|-----|-------|---|------------------|------------------|---------------|
| ICB1      | Cyanide | 0.0016         | +/-0.005            | U            |     | 0.005 |   | 04/02/2018       | 10:27            | LB94301       |



**GENCHEM**

- 3a -

### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** LB94301

SDG No.: LB94301

**Contract:**

Lab Code: CHEM

**Case No.:** LB94301

**SAS No.:** LB94301

| Sample ID | Analyte | Result<br>mg/L | Acceptance<br>Limit | Conc<br>Qual | LOD | CRQL  | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|-----|-------|---|------------------|------------------|---------------|
| CCB1      | Cyanide | 0.002          | +/-0.005            | U            |     | 0.005 |   | 04/02/2018       | 10:27            | LB94301       |
| CCB2      | Cyanide | 0.00057        | +/-0.005            | U            |     | 0.005 |   | 04/02/2018       | 10:35            | LB94301       |
| CCB3      | Cyanide | 0.00062        | +/-0.005            | U            |     | 0.005 |   | 04/02/2018       | 11:31            | LB94301       |



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB94301

SDG No.: LB94301

Contract: Lab Code: CHEM

Case No.: LB94301

SAS No.: LB94301

| Sample ID | Analyte | Result<br>mg/L | Acceptance<br>Limit | Conc<br>Qual | LOD | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|-----|------|---|------------------|------------------|---------------|
|-----------|---------|----------------|---------------------|--------------|-----|------|---|------------------|------------------|---------------|



**GENCHEM**  
**- 3b -**  
**PREPARATION BLANK SUMMARY**

**Client:** LB94301

**SDG No.:** LB94301

**Instrument:** Konelab 20

| Sample ID         | Analyte | Result<br>(mg/Kg)     | Acceptance<br>Limit | Conc<br>Qual       | LOD<br>mg/Kg | CRQL<br>mg/Kg    | M | Analysis<br>Date                | Analysis<br>Time           | Run     |
|-------------------|---------|-----------------------|---------------------|--------------------|--------------|------------------|---|---------------------------------|----------------------------|---------|
| <b>PB107847BL</b> | Cyanide | <b>SOLID</b><br>0.045 | <0.25               | Batch Number:<br>J |              | PB107847<br>0.25 |   | <b>Prep Date:</b><br>04/02/2018 | <b>03/30/2018</b><br>10:27 | LB94301 |



GENCHEM  
- 5a -  
MATRIX SPIKE SUMMARY

client: LB94301 level: low sdg no.: LB94301  
contract: lab code: CHEM case no.: LB94301 sas no.: LB94301  
matrix: Solid sample id: J2142-01 client id: GB-RT-2334MS  
Percent Solids for Sample: 81.8 Spiked ID: J2142-01MS Percent Solids for Spike Sample: 81.8

| Analyte | Units | Acceptance<br>Limit %R | Spiked<br>Result | C | Sample<br>Result | C | Spike<br>Added | %<br>Recovery | Qual | M |
|---------|-------|------------------------|------------------|---|------------------|---|----------------|---------------|------|---|
| Cyanide | mg/Kg | 75 - 125               | 1.4              |   | 0.17             | J | 2.4            | 51            | *    |   |



GENCHEM  
- 5a -  
MATRIX SPIKE DUPLICATE SUMMARY

client: LB94301 level: low sdg no.: LB94301  
contract: lab code: CHEM case no.: LB94301 sas no.: LB94301  
matrix: Solid sample id: J2142-01 client id: GB-RT-2334MSD  
Percent Solids for Sample: 81.8 Spiked ID: J2142-01MSD Percent Solids for Spike Sample: 81.8

| Analyte | Units | Acceptance<br>Limit %R | MSD<br>Result | C | Sample<br>Result | C | Spike<br>Added | %<br>Recovery | Qual | M |
|---------|-------|------------------------|---------------|---|------------------|---|----------------|---------------|------|---|
| Cyanide | mg/Kg | 75 - 125               | 1.3           |   | 0.17             | J | 2.4            | 47            | *    |   |

# GENCHEM

- 6 -

## DUPLICATE SAMPLE SUMMARY

**Client:** LB94301      **Level:** LOW      **SDG No.:** LB94301  
**Contract:**                           **Lab Code:** CHEM      **Case No.:** LB94301      **SAS No.:** LB94301  
**Matrix:** Solid      **Sample ID:** J2142-01      **Client ID:** GB-RT-2334DUP  
**Percent Solids for Sample:** 81.8      **Duplicate ID** J2142-01DUP      **Percent Solids for Spike Sample:** 81.8

| Analyte | Units | Acceptance<br>Limit | Sample<br>Result | C | Duplicate<br>Result | C | RPD | Qual | M |
|---------|-------|---------------------|------------------|---|---------------------|---|-----|------|---|
| Cyanide | mg/Kg | 20                  | 0.17             | J | 0.14                | J | 19  |      |   |

# GENCHEM

- 6 -

## DUPLICATE SAMPLE SUMMARY

**Client:** LB94301      **Level:** LOW      **SDG No.:** LB94301  
**Contract:**                           **Lab Code:** CHEM      **Case No.:** LB94301      **SAS No.:** LB94301  
**Matrix:** Solid      **Sample ID:** J2142-01MS      **Client ID:** GB-RT-2334MSD  
**Percent Solids for Sample:** 81.8      **Duplicate ID** J2142-01MSD      **Percent Solids for Spike Sample:** 81.8

| Analyte | Units | Acceptance<br>Limit | Sample<br>Result | C | Duplicate<br>Result | C | RPD | Qual | M |
|---------|-------|---------------------|------------------|---|---------------------|---|-----|------|---|
| Cyanide | mg/Kg | 20                  | 1.4              |   | 1.3                 |   | 7   |      |   |



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB94301

SDG No.: LB94301

Contract:

Lab Code: CHEM

Case No.: LB94301

SAS No.: LB94301

| Analyte               | Units | True Value | Result | C | %<br>Recovery | Acceptance<br>Limits | M |
|-----------------------|-------|------------|--------|---|---------------|----------------------|---|
| PB107847BS<br>Cyanide | mg/Kg | 4.8        | 4.6    |   | 95.8          | 85 - 115             |   |