

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/2/2018 13:25

Test: Phenolics-mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.023	0.0	0.109	
ICB1	0.003	0.0	0.003	
CCV1	0.971	0.0	0.103	
CCB1	0.004	0.0	0.003	
PB107868BL	0.004	0.0	0.003	
PB107868BS	1.041	0.0	0.111	
J2083-01	0.005	0.0	0.003	
J2083-02	0.003	0.0	0.003	
J2116-01	0.002	0.0	0.003	
J2116-02	0.583	0.0	0.063	
J2116-03	0.591	0.0	0.064	
J2116-04	0.003	0.0	0.003	
J2116-05	0.002	0.0	0.003	
J2116-06	0.003	0.0	0.003	
CCV2	0.964	0.0	0.103	
CCB2	0.004	0.0	0.003	
J2116-07	0.003	0.0	0.003	
J2136-01	0.003	0.0	0.003	
J2136-01DUP	0.003	0.0	0.003	
J2136-02	0.002	0.0	0.003	
J2136-03	0.003	0.0	0.003	
CCV3	0.940	0.0	0.100	
CCB3	0.003	0.0	0.003	

N	23
Mean	0.268
SD	0.4224
CV%	157.63

AS

Aquakem v. 7.2AQ1

Results from time period:

Mon Apr 02 12:42:29 2018

Mon Apr 02 13:23:59 2018

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0528	mg/l	4/2/2018 12:42:29	
0.1PPM	A	Phenolics-	P	0.1021	mg/l	4/2/2018 12:42:30	
0.25PPM	A	Phenolics-	P	0.2502	mg/l	4/2/2018 12:42:31	
0.50PPM	A	Phenolics-	P	0.4974	mg/l	4/2/2018 12:42:32	
1.0PPM	A	Phenolics-	P	0.9943	mg/l	4/2/2018 12:42:33	
2.0PPM	A	Phenolics-	P	2.0033	mg/l	4/2/2018 12:42:34	
ICV1	S	Phenolics-	P	1.0228	mg/l	4/2/2018 13:13:46	
ICB1	S	Phenolics-	P	0.0028	mg/l	4/2/2018 13:13:47	
CCV1	S	Phenolics-	P	0.9706	mg/l	4/2/2018 13:13:48	
CCB1	S	Phenolics-	P	0.0041	mg/l	4/2/2018 13:13:49	
PB107868BL	S	Phenolics-	P	0.0035	mg/l	4/2/2018 13:13:50	
PB107868BS	S	Phenolics-	P	1.0415	mg/l	4/2/2018 13:13:51	
J2083-01	S	Phenolics-	P	0.0048	mg/l	4/2/2018 13:13:52	
J2083-02	S	Phenolics-	P	0.0031	mg/l	4/2/2018 13:13:53	
J2116-01	S	Phenolics-	P	0.0016	mg/l	4/2/2018 13:13:54	
J2116-02	S	Phenolics-	P	0.5831	mg/l	4/2/2018 13:13:55	
J2116-03	S	Phenolics-	P	0.5915	mg/l	4/2/2018 13:13:56	
J2116-04	S	Phenolics-	P	0.0032	mg/l	4/2/2018 13:13:57	
J2116-05	S	Phenolics-	P	0.0022	mg/l	4/2/2018 13:23:48	
J2116-06	S	Phenolics-	P	0.0033	mg/l	4/2/2018 13:23:49	
CCV2	S	Phenolics-	P	0.9644	mg/l	4/2/2018 13:23:50	
CCB2	S	Phenolics-	P	0.0041	mg/l	4/2/2018 13:23:51	
J2116-07	S	Phenolics-	P	0.0033	mg/l	4/2/2018 13:23:52	
J2136-01	S	Phenolics-	P	0.0031	mg/l	4/2/2018 13:23:53	
J2136-01DUP	S	Phenolics-	P	0.0027	mg/l	4/2/2018 13:23:54	
J2136-02	S	Phenolics-	P	0.0017	mg/l	4/2/2018 13:23:55	
J2136-03	S	Phenolics-	P	0.0028	mg/l	4/2/2018 13:23:56	
CCV3	S	Phenolics-	P	0.94	mg/l	4/2/2018 13:23:57	
CCB3	S	Phenolics-	P	0.0027	mg/l	4/2/2018 13:23:58	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/2/2018 12:46


Test Phenolics-

Accepted 4/2/2018 12:46

Factor 9.663
Bias 0.003

Coeff. of det. 0.999978

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0528	0.0500	
2	Phenol-2	0.013	0.1021	0.1000	
3	Phenol-2	0.029	0.2502	0.2500	
4	Phenol-2	0.054	0.4974	0.5000	
5	Phenol-2	0.106	0.9943	1.0000	
6	Phenol-2	0.210	2.0033	2.0000	