

=====

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/2/2018 14:35

Test: Total CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.774	0.0	0.097	
ICB1	1.509	0.0	0.004	
CCV1	244.034	0.0	0.237	
CCB1	1.312	0.0	0.004	
PB107892BL	1.103	0.0	0.004	
PB107892BS	96.635	0.0	0.096	
J1757-07	1.951	0.0	0.005	
J1757-07DUP	1.738	0.0	0.005	
J1757-07MS	37.757	0.0	0.039	
J1757-07MSD	37.700	0.0	0.039	
J1757-09	1.195	0.0	0.004	
J1757-11	1.511	0.0	0.004	
J2148-01	1.299	0.0	0.004	
J2148-04	1.370	0.0	0.004	
CCV2	246.857	0.0	0.240	
CCB2	1.393	0.0	0.004	
J2169-01	1.147	0.0	0.004	
J2175-01	6.743	0.0	0.009	
J2178-01	6.097	0.0	0.009	
CCV3	249.754	0.0	0.242	
CCB3	1.285	0.0	0.004	

N 21
Mean 49.532
SD 87.5212
CV% 176.70

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/2/2018 10:01


Test Total CN

Accepted 4/2/2018 10:01

Factor 1042
Bias 0.003

Coeff. of det. 0.999922

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.6135	0.0000	
2	5.0PPBCN	0.008	5.4770	5.0000	
3	10PPBCN	0.013	10.9758	10.0000	
4	50PPBCN	0.052	50.7719	50.0000	
5	100PPBCN	0.097	98.4056	100.0000	
6	250PPBCN	0.240	247.0776	250.0000	
7	500PPBCN	0.484	501.6786	500.0000	

AS

Aquakem v. 7.2AQ1

Results from time period:

Mon Apr 02 14:22:24 2018

Mon Apr 02 14:53:48 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.6135	µg/l	4/2/2018 9:04:12	
0.0PPBCN	A	Total CN	P	0.6135	µg/l	4/2/2018 9:04:12	
0.0PPBCN	A	Total CN	P	0.6135	µg/l	4/2/2018 9:04:12	
0.0PPBCN	A	Total CN	P	0.6135	µg/l	4/2/2018 9:04:12	
0.0PPBCN	A	Total CN	P	0.6135	µg/l	4/2/2018 9:04:12	
0.0PPBCN	A	Total CN	P	0.6135	µg/l	4/2/2018 9:04:12	
0.0PPBCN	A	Total CN	P	0.6135	µg/l	4/2/2018 9:04:12	
ICV1	S	Total CN	P	97.7737	µg/l	4/2/2018 14:22:24	
ICB1	S	Total CN	P	1.509	µg/l	4/2/2018 14:22:25	
CCV1	S	Total CN	P	244.0342	µg/l	4/2/2018 14:22:26	
CCB1	S	Total CN	P	1.3122	µg/l	4/2/2018 14:22:27	
PB107892BL	S	Total CN	P	1.1033	µg/l	4/2/2018 14:22:28	
PB107892BS	S	Total CN	P	96.6352	µg/l	4/2/2018 14:22:29	
J1757-07	S	Total CN	P	1.951	µg/l	4/2/2018 14:22:30	
J1757-07DUP	S	Total CN	P	1.7383	µg/l	4/2/2018 14:22:31	
J1757-07MS	S	Total CN	P	37.757	µg/l	4/2/2018 14:22:32	
J1757-07MSD	S	Total CN	P	37.7003	µg/l	4/2/2018 14:22:33	
J1757-09	S	Total CN	P	1.1949	µg/l	4/2/2018 14:22:34	
J1757-11	S	Total CN	P	1.5107	µg/l	4/2/2018 14:29:57	
J2148-01	S	Total CN	P	1.2988	µg/l	4/2/2018 14:29:58	
J2148-04	S	Total CN	P	1.3704	µg/l	4/2/2018 14:29:59	
CCV2	S	Total CN	P	246.8573	µg/l	4/2/2018 14:30:00	
CCB2	S	Total CN	P	1.3927	µg/l	4/2/2018 14:30:01	
J2169-01	S	Total CN	P	1.1474	µg/l	4/2/2018 14:30:02	
J2175-01	S	Total CN	P	6.7425	µg/l	4/2/2018 14:30:03	
J2178-01	S	Total CN	P	6.0975	µg/l	4/2/2018 14:30:04	
CCV3	S	Total CN	P	249.7544	µg/l	4/2/2018 14:30:05	
CCB3	S	Total CN	P	1.2846	µg/l	4/2/2018 14:30:06	