

Chemtech Consulting Group


Analytical Review Report

Date Printed : 4/5/18

Analyst : AK

Data File : LB94369.csv

Approved By : 

Approved Date : 

Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Ortho Phosphate											
STD1	STD1			W	3/19/18						
Ortho Phosphate		PASS		0.000	0.000					mg/L	
STD2	STD2			W	3/19/18						
Ortho Phosphate		PASS		0.156	0.156					mg/L	
STD3	STD3			W	3/19/18						
Ortho Phosphate		PASS		0.916	0.916					mg/L	
STD4	STD4			W	3/19/18						
Ortho Phosphate		PASS		2.096	2.10					mg/L	
STD5	STD5			W	3/19/18						
Ortho Phosphate		PASS		5.113	5.11					mg/L	
STD6	STD6			W	3/19/18						
Ortho Phosphate		PASS		10.139	10.1					mg/L	
STD7	STD7			W	3/19/18						
Ortho Phosphate		PASS		12.330	12.3					mg/L	
ICV1	ICV1			W	3/19/18						
Ortho Phosphate		PASS		5.261	5.26	105.0	90	110		mg/L	
ICB1	ICB1			W	3/19/18						
Ortho Phosphate		PASS		0.000	0.000			+/-0.0100		mg/L	
CCV1	CCV1			W	3/20/18						
Ortho Phosphate		PASS		4.934	4.93	99.0	90	110		mg/L	
CCB1	CCB1			W	3/20/18						
Ortho Phosphate		PASS		0.000	0.000			+/-0.0100		mg/L	
LB94369BLS	LB94369BLS			S	3/20/18						
Ortho Phosphate		PASS		0.000	0.000			+/-0.0020		mg/Kg	
LB94369BSS	LB94369BSS			S	3/20/18						
Ortho Phosphate		PASS		5.193	104.00	104.0	90.00	110.00		mg/Kg	
J1161-01	LOD-MDL-SOIL-01-QT1-2018		1	S	3/20/18						
Ortho Phosphate		PASS		0.116	2.32					mg/Kg	
J1161-03	MDL-SOIL-03-QT1-2018		1	S	3/20/18						
Ortho Phosphate		PASS		0.088	1.76					mg/Kg	
CCV2	CCV2			W	3/20/18						
Ortho Phosphate		PASS		5.299	5.30	106.0	90	110		mg/L	
CCB2	CCB2			W	3/20/18						
Ortho Phosphate		PASS		0.000	0.000			+/-0.0100		mg/L	