

Daily Analysis Runlog For Sequence/QC Batch ID # LB94459

Review By		Review On				
SubDirectory		LB94459		Test : Anions		
STD. NAME		STD REF.#				
ICAL Standard		wp62978,wp62979,wp62980,wp62981,wp62982,wp62983,wp62984				
ICV Standard		wp62985				
CCV Standard		wp62728				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		wp62729				
Chk Standard		WP62458,WP62459,wp62727				
Sr#	SampleID	ClientID	QcType	Date	Comment	Status
1	STD1	STD1	CAL1	03/12/18 14:48		OK
2	STD2	STD2	CAL2	03/12/18 15:19		OK
3	STD3	STD3	CAL3	03/12/18 15:50		OK
4	STD4	STD4	CAL4	03/12/18 16:21		OK
5	STD5	STD5	CAL5	03/12/18 16:52		OK
6	STD6	STD6	CAL6	03/12/18 17:22		OK
7	STD7	STD7	CAL7	03/12/18 17:53		OK
8	ICV1	ICV1	ICV	03/12/18 19:26		OK
9	ICB1	ICB1	ICB	03/12/18 19:57		OK
10	CCV1	CCV1	CCV	03/15/18 19:18		OK
11	CCB1	CCB1	CCB	03/15/18 19:49		OK
12	LB94459BLW	LB94459BLW	MB	03/15/18 20:20		OK
13	LB94459BSW	LB94459BSW	LCS	03/15/18 20:51		OK
14	J1161-07	LOD-MDL-WATER-0	SAM	03/15/18 21:22		OK
15	J1161-09	MDL-WATER-03-QT	SAM	03/15/18 22:23		OK
16	CCV2	CCV2	CCV	03/16/18 03:02		OK
17	CCB2	CCB2	CCB	03/16/18 03:33		OK



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/16/2018 12:00:00 AM
Project:	LB94459	Date Received:	1/16/2018 12:00:00 AM
Client Sample ID:	LOD-MDL-WATER-01-QT1-2018	SDG No.:	LB94459
Lab Sample ID:	J1161-07	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	Bromide	0.54		1	0.066		0.5	mg/L		03/15/2018	9056
	Chloride	0.17		1	0.075		0.15	mg/L		03/15/2018	9056
	Fluoride	0.11		1	0.043		0.1	mg/L		03/15/2018	9056
	Nitrate	0.14		1	0.027		0.11	mg/L		03/15/2018	9056
	Nitrite	0.18		1	0.022		0.15	mg/L		03/15/2018	9056
	Phosphorus, Ortho	0.061	J	1	0.01		0.24	mg/L		03/15/2018	9056
	Sulfate	0.89		1	0.13		0.75	mg/L		03/15/2018	9056



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/16/2018 12:00:00 AM
Project:	LB94459	Date Received:	1/16/2018 12:00:00 AM
Client Sample ID:	MDL-WATER-03-QT1-2018	SDG No.:	LB94459
Lab Sample ID:	J1161-09	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	Bromide	0.54		1	0.066		0.5	mg/L		03/15/2018	9056
	Chloride	0.17		1	0.075		0.15	mg/L		03/15/2018	9056
	Fluoride	0.11		1	0.043		0.1	mg/L		03/15/2018	9056
	Nitrate	0.14		1	0.027		0.11	mg/L		03/15/2018	9056
	Nitrite	0.19		1	0.022		0.15	mg/L		03/15/2018	9056
	Phosphorus, Ortho	0.091	J	1	0.01		0.24	mg/L		03/15/2018	9056
	Sulfate	0.92		1	0.13		0.75	mg/L		03/15/2018	9056



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/16/2018 12:00:00 AM
Project:	LB94459	Date Received:	1/16/2018 12:00:00 AM
Client Sample ID:	MDL-WATER-03-QT1-2018	SDG No.:	LB94459
Lab Sample ID:	J1161-09	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
-----	-----------	-------	------	----	-----	-----	------------	-------	-----------	-----------	----------

Color Before:	Clarity Before:	Texture:
Color After:	Clarity After:	Artifacts:
Comments:		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB94459

SDG No.: LB94459

Contract: _____ Lab Code: CHEM

Case No.: LB94459

SAS No.: LB94459

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV1	Bromide	9.6	10	96		90 - 110		03/12/2018	19:26	LB94459
	Chloride	2.9	3	96.7		90 - 110		03/12/2018	19:26	LB94459
	Fluoride	1.9	2	95		90 - 110		03/12/2018	19:26	LB94459
	Phosphorus, Ortho	4.7	5	94		90 - 110		03/12/2018	19:26	LB94459
	Nitrite	2.8	3	93.3		90 - 110		03/12/2018	19:26	LB94459
	Nitrate	2.4	2.5	96		90 - 110		03/12/2018	19:26	LB94459
	Sulfate	14.2	15	94.7		90 - 110		03/12/2018	19:26	LB94459



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB94459

SDG No.: LB94459

Contract: _____ Lab Code: CHEM

Case No.: LB94459

SAS No.: LB94459

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	Bromide	10.2	10	102		90 - 110		03/15/2018	19:18	LB94459
	Chloride	3.1	3	103.3		90 - 110		03/15/2018	19:18	LB94459
	Fluoride	2	2	100		90 - 110		03/15/2018	19:18	LB94459
	Phosphorus, Ortho	4.8	5	96		90 - 110		03/15/2018	19:18	LB94459
	Nitrite	3	3	100		90 - 110		03/15/2018	19:18	LB94459
	Nitrate	2.5	2.5	100		90 - 110		03/15/2018	19:18	LB94459
	Sulfate	14.6	15	97.3		90 - 110		03/15/2018	19:18	LB94459
CCV2	Bromide	10.3	10	103		90 - 110		03/16/2018	03:02	LB94459
	Chloride	3.1	3	103.3		90 - 110		03/16/2018	03:02	LB94459
	Fluoride	2	2	100		90 - 110		03/16/2018	03:02	LB94459
	Phosphorus, Ortho	4.9	5	98		90 - 110		03/16/2018	03:02	LB94459
	Nitrite	3	3	100		90 - 110		03/16/2018	03:02	LB94459
	Nitrate	2.6	2.5	104		90 - 110		03/16/2018	03:02	LB94459
	Sulfate	14.6	15	97.3		90 - 110		03/16/2018	03:02	LB94459



GENCHEM
- 3b -
PREPARATION BLANK SUMMARY

Client: LB94459

SDG No.: LB94459

Instrument: IC-2

Sample ID	Analyte	Result (mg/L)	Acceptance Limit	Conc Qual	LOD mg/L	CRQL mg/L	M	Analysis Date	Analysis Time	Run
LB94459BLW		WATER		Batch Number:				Prep Date:		
	Bromide	0.066	<0.5	U		0.5		03/15/2018	20:20	LB94459
	Chloride	0.075	<0.15	U		0.15		03/15/2018	20:20	LB94459
	Fluoride	0.043	<0.1	U		0.1		03/15/2018	20:20	LB94459
	Phosphorus, Ortho	0.01	<0.24	U		0.24		03/15/2018	20:20	LB94459
	Nitrite	0.022	<0.15	U		0.15		03/15/2018	20:20	LB94459
	Nitrate	0.027	<0.11	U		0.11		03/15/2018	20:20	LB94459
	Sulfate	0.13	<0.75	U		0.75		03/15/2018	20:20	LB94459



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB94459

SDG No.: LB94459

Contract:

Lab Code: CHEM

Case No.: LB94459

SAS No.: LB94459

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	Bromide	0.066	+/-0.5	U		0.5		03/12/2018	19:57	LB94459
	Chloride	0.075	+/-0.15	U		0.15		03/12/2018	19:57	LB94459
	Fluoride	0.043	+/-0.1	U		0.1		03/12/2018	19:57	LB94459
	Phosphorus, Ortho	0.01	+/-0.24	U		0.24		03/12/2018	19:57	LB94459
	Nitrite	0.022	+/-0.15	U		0.15		03/12/2018	19:57	LB94459
	Nitrate	0.027	+/-0.11	U		0.11		03/12/2018	19:57	LB94459
	Sulfate	0.13	+/-0.75	U		0.75		03/12/2018	19:57	LB94459



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB94459

SDG No.: LB94459

Contract:

Lab Code: CHEM

Case No.: LB94459

SAS No.: LB94459

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	Bromide	0.066	+/-0.5	U		0.5		03/15/2018	19:49	LB94459
	Chloride	0.075	+/-0.15	U		0.15		03/15/2018	19:49	LB94459
	Fluoride	0.043	+/-0.1	U		0.1		03/15/2018	19:49	LB94459
	Phosphorus, Ortho	0.01	+/-0.24	U		0.24		03/15/2018	19:49	LB94459
	Nitrite	0.022	+/-0.15	U		0.15		03/15/2018	19:49	LB94459
	Nitrate	0.027	+/-0.11	U		0.11		03/15/2018	19:49	LB94459
	Sulfate	0.13	+/-0.75	U		0.75		03/15/2018	19:49	LB94459
CCB2	Bromide	0.066	+/-0.5	U		0.5		03/16/2018	03:33	LB94459
	Chloride	0.075	+/-0.15	U		0.15		03/16/2018	03:33	LB94459
	Fluoride	0.043	+/-0.1	U		0.1		03/16/2018	03:33	LB94459
	Phosphorus, Ortho	0.01	+/-0.24	U		0.24		03/16/2018	03:33	LB94459
	Nitrite	0.022	+/-0.15	U		0.15		03/16/2018	03:33	LB94459
	Nitrate	0.027	+/-0.11	U		0.11		03/16/2018	03:33	LB94459
	Sulfate	0.13	+/-0.75	U		0.75		03/16/2018	03:33	LB94459



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB94459

SDG No.: LB94459

Contract: Lab Code: CHEM

Case No.: LB94459

SAS No.: LB94459

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB94459

SDG No.: LB94459

Contract:

Lab Code: CHEM

Case No.: LB94459

SAS No.: LB94459

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
LB94459BSW							
Bromide	mg/L	10	9.8		98	90 - 110	
Chloride	mg/L	3	2.9		96.7	90 - 110	
Fluoride	mg/L	2	2		100	90 - 110	
Phosphorus, Ortho	mg/L	5	4.7		94	90 - 110	
Nitrite	mg/L	3	2.9		96.7	90 - 110	
Nitrate	mg/L	2.5	2.5		100	90 - 110	
Sulfate	mg/L	15	14.8		98.7	90 - 110	
