

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/9/2018 12:09

Test: Phenolics-*mg/L*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.009	0.0	0.107	
ICB1	0.004	0.0	0.004	
CCV1	1.017	0.0	0.108	
CCB1	0.001	0.0	0.004	
PB108073BL	0.002	0.0	0.004	
PB108073BS	1.008	0.0	0.107	
J2252-01	-0.003	0.0	0.004	
J2252-01DUP	0.006	0.0	0.004	
J2252-01MS	0.838	0.0	0.089	
J2252-01MSD	0.842	0.0	0.090	
CCV2	1.012	0.0	0.107	
CCB2	-0.001	0.0	0.004	

N	12
Mean	0.478
SD	0.5013
CV%	104.89

AL.

Aquakem v. 7.2AQ1

Results from time period:

Mon Apr 09 11:36:26 2018

Mon Apr 09 12:04:24 2018

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0526	mg/l	4/9/2018 11:36:26	
0.1PPM	A	Phenolics-	P	0.1059	mg/l	4/9/2018 11:36:27	
0.25PPM	A	Phenolics-	P	0.2529	mg/l	4/9/2018 11:36:28	
0.50PPM	A	Phenolics-	P	0.4916	mg/l	4/9/2018 11:36:29	
1.0PPM	A	Phenolics-	P	0.9913	mg/l	4/9/2018 11:36:30	
2.0PPM	A	Phenolics-	P	2.0057	mg/l	4/9/2018 11:36:31	
ICV1	S	Phenolics-	P	1.0094	mg/l	4/9/2018 12:04:13	
ICB1	S	Phenolics-	P	0.0042	mg/l	4/9/2018 12:04:14	
CCV1	S	Phenolics-	P	1.0166	mg/l	4/9/2018 12:04:15	
CCB1	S	Phenolics-	P	0.0008	mg/l	4/9/2018 12:04:16	
PB108073BL	S	Phenolics-	P	0.0024	mg/l	4/9/2018 12:04:17	
PB108073BS	S	Phenolics-	P	1.0083	mg/l	4/9/2018 12:04:18	
J2252-01	S	Phenolics-	P	-0.0034	mg/l	4/9/2018 12:04:19	
J2252-01DUP	S	Phenolics-	P	0.0055	mg/l	4/9/2018 12:04:20	
J2252-01MS	S	Phenolics-	P	0.8382	mg/l	4/9/2018 12:04:21	
J2252-01MSD	S	Phenolics-	P	0.8421	mg/l	4/9/2018 12:04:22	
CCV2	S	Phenolics-	P	1.0124	mg/l	4/9/2018 12:04:23	
CCB2	S	Phenolics-	P	-0.001	mg/l	4/9/2018 12:04:24	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/9/2018 11:42

Test Phenolics-


Accepted 4/9/2018 11:42

Factor 9.795

Bias 0.004

Coeff. of det. -0.999918

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0526	0.0500	
2	Phenol-2	0.015	0.1059	0.1000	
3	Phenol-2	0.030	0.2529	0.2500	
4	Phenol-2	0.054	0.4916	0.5000	
5	Phenol-2	0.105	0.9913	1.0000	
6	Phenol-2	0.209	2.0057	2.0000	