

=====

Test results

Aquakem 7.2AQ1

LB94537
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/11/2018 13:59

Test: Total CN *ug/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.569	0.0	0.096	
ICB1	0.363	0.0	0.004	
CCV1	242.899	0.0	0.236	
CCB1	0.223	0.0	0.004	
PB108159BL	-0.052	0.0	0.003	
PB108159BS	97.596	0.0	0.097	
MIDPB108159	242.802	0.0	0.236	
J2318-13	7.506	0.0	0.011	
J2318-13DUP	6.946	0.0	0.010	
J2318-13MS	44.218	0.0	0.046	
J2318-13MSD	44.037	0.0	0.046	
CCV2	257.027	0.0	0.250	
CCB2	0.314	0.0	0.004	

N	13
Mean	80.034
SD	101.4865
CV%	126.80

Aquakem v. 7.2AQ1

AS

Results from time period:

Wed Apr 11 13:55:59 2018

Wed Apr 11 13:59:23 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.254	µg/l	4/11/2018 12:05:15	
5.OPPBCN	A	Total CN	P	4.4395	µg/l	4/11/2018 12:05:16	
10PPBCN	A	Total CN	P	9.5353	µg/l	4/11/2018 12:05:17	
50PPBCN	A	Total CN	P	49.7818	µg/l	4/11/2018 12:05:18	
100PPBCN	A	Total CN	P	100.0643	µg/l	4/11/2018 12:05:19	
250PPBCN	A	Total CN	P	252.8187	µg/l	4/11/2018 12:05:20	
500PPBCN	A	Total CN	P	498.6145	µg/l	4/11/2018 12:05:21	
ICV1	S	Total CN	P	96.5694	µg/l	4/11/2018 13:55:59	
ICB1	S	Total CN	P	0.3628	µg/l	4/11/2018 13:56:00	
CCV1	S	Total CN	P	242.8989	µg/l	4/11/2018 13:56:01	
CCB1	S	Total CN	P	0.2225	µg/l	4/11/2018 13:56:02	
PB108159BL	S	Total CN	P	-0.0522	µg/l	4/11/2018 13:56:03	
PB108159BS	S	Total CN	P	97.5964	µg/l	4/11/2018 13:56:04	
MIDPB108159	S	Total CN	P	242.8025	µg/l	4/11/2018 13:56:05	
J2318-13	S	Total CN	P	7.5058	µg/l	4/11/2018 13:56:06	
J2318-13DUP	S	Total CN	P	6.9455	µg/l	4/11/2018 13:56:07	
J2318-13MS	S	Total CN	P	44.2177	µg/l	4/11/2018 13:56:08	
J2318-13MSD	S	Total CN	P	44.0368	µg/l	4/11/2018 13:56:09	
CCV2	S	Total CN	P	257.0269	µg/l	4/11/2018 13:59:22	
CCB2	S	Total CN	P	0.3143	µg/l	4/11/2018 13:59:23	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

4/11/2018 12:14


Test Total CN

Accepted 4/11/2018 12:14

Factor 1043
 Bias 0.003

Coeff. of det. 0.999949

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors

1	0.0PPBCN	0.003	-0.2540	0.0000	
2	5.0PPBCN	0.008	4.4395	5.0000	
3	10PPBCN	0.013	9.5353	10.0000	
4	50PPBCN	0.051	49.7818	50.0000	
5	100PPBCN	0.099	100.0643	100.0000	
6	250PPBCN	0.246	252.8187	250.0000	
7	500PPBCN	0.481	498.6145	500.0000	