


Chemtech Consulting Group

Analytical Review Report

Date Printed : 4/16/18
Analyst : AK
Data File : LB94633.csv


Approved By : 
Approved Date : 
Worksheet # : 

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1	
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Bromide												
STD1	STD1				W	4/4/18						
Bromide		PASS	0.000		0.000						mg/L	
STD2	STD2				W	4/4/18						
Bromide		PASS	0.707		0.707						mg/L	
STD3	STD3				W	4/4/18						
Bromide		PASS	2.080		2.08						mg/L	
STD4	STD4				W	4/4/18						
Bromide		PASS	3.953		3.95						mg/L	
STD5	STD5				W	4/4/18						
Bromide		PASS	9.541		9.54						mg/L	
STD6	STD6				W	4/4/18						
Bromide		PASS	19.573		19.6						mg/L	
STD7	STD7				W	4/4/18						
Bromide		PASS	25.448		25.4						mg/L	
ICV1	ICV1				W	4/4/18						
Bromide		PASS	9.782		9.78	98.0	90	110			mg/L	
ICB1	ICB1				W	4/4/18						
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L	
CCV1	CCV1				W	4/12/18						
Bromide		PASS	9.997		10.00	100.0	90	110			mg/L	
CCB1	CCB1				W	4/12/18						
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L	
LB94633BLS	LB94633BLS				S	4/12/18						
Bromide		PASS	0.000		0.000			+/-0.1900			mg/Kg	
LB94633BSS	LB94633BSS				S	4/12/18						
Bromide		PASS	10.072		201.00	101.0	90.00	110.00			mg/Kg	
J2133-01	LOD-MDL-SOIL-01-QT2-2018		1		S	4/12/18						
Bromide		PASS	0.552		5.51						mg/Kg	
J2133-03	MDL-SOIL-03-QT2-2018		1		S	4/12/18						
Bromide		PASS	0.572		5.72						mg/Kg	
J2133-02	LOQ-SOIL-02-QT2-2018				S	4/12/18						
Bromide		PASS	1.260		25.00	125.0	50.00	150.00			mg/Kg	
CCV2	CCV2				W	4/12/18						
Bromide		PASS	10.048		10.00	100.0	90	110			mg/L	
CCB2	CCB2				W	4/12/18						
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 4/16/18
Analyst : AK
Data File : LB94633.csv


Approved By : 
Approved Date :
Worksheet # :

Lab Sample ID	Client ID		Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method			Line 1	
Parameter						%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Chloride												
STD1	STD1				W	4/4/18						
Chloride		PASS	0.000		0.000						mg/L	
STD2	STD2				W	4/4/18						
Chloride		PASS	0.217		0.217						mg/L	
STD3	STD3				W	4/4/18						
Chloride		PASS	0.635		0.635						mg/L	
STD4	STD4				W	4/4/18						
Chloride		PASS	1.188		1.19						mg/L	
STD5	STD5				W	4/4/18						
Chloride		PASS	2.859		2.86						mg/L	
STD6	STD6				W	4/4/18						
Chloride		PASS	5.871		5.87						mg/L	
STD7	STD7				W	4/4/18						
Chloride		PASS	7.639		7.64						mg/L	
ICV1	ICV1				W	4/4/18						
Chloride		PASS	2.898		2.90	97.0	90	110			mg/L	
ICB1	ICB1				W	4/4/18						
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L	
CCV1	CCV1				W	4/12/18						
Chloride		PASS	2.989		2.99	100.0	90	110			mg/L	
CCB1	CCB1				W	4/12/18						
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L	
LB94633BLS	LB94633BLS				S	4/12/18						
Chloride		PASS	0.000		0.000			+/-0.0320			mg/Kg	
LB94633BSS	LB94633BSS				S	4/12/18						
Chloride		PASS	3.005		60.00	100.0	90.00	110.00			mg/Kg	
J2133-01	LOD-MDL-SOIL-01-QT2-2018		1		S	4/12/18						
Chloride		PASS	0.202		2.02						mg/Kg	
J2133-03	MDL-SOIL-03-QT2-2018		1		S	4/12/18						
Chloride		PASS	0.193		1.93						mg/Kg	
J2133-02	LOQ-SOIL-02-QT2-2018				S	4/12/18						
Chloride		PASS	0.397		7.89	132.0	50.00	150.00			mg/Kg	
CCV2	CCV2				W	4/12/18						
Chloride		PASS	3.008		3.01	100.0	90	110			mg/L	
CCB2	CCB2				W	4/12/18						
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 4/16/18
 Analyst : AK
 Data File : LB94633.csv

Approved By : 
 Approved Date : 4/10/18
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter			PPB			%Rec	Method		Method		Line 2
							LCL	UCL	RPD	Max RPD	Units
Fluoride											
STD1	STD1				W	4/4/18					
Fluoride		PASS	0.000		0.000						mg/L
STD2	STD2				W	4/4/18					
Fluoride		PASS	0.091		0.091						mg/L
STD3	STD3				W	4/4/18					
Fluoride		PASS	0.409		0.409						mg/L
STD4	STD4				W	4/4/18					
Fluoride		PASS	0.763		0.763						mg/L
STD5	STD5				W	4/4/18					
Fluoride		PASS	1.890		1.89						mg/L
STD6	STD6				W	4/4/18					
Fluoride		PASS	3.998		4.00						mg/L
STD7	STD7				W	4/4/18					
Fluoride		PASS	5.033		5.03						mg/L
ICV1	ICV1				W	4/4/18					
Fluoride		PASS	1.945		1.94	97.0	90	110			mg/L
ICB1	ICB1				W	4/4/18					
Fluoride		PASS	0.000		0.000			+/-0.0430			mg/L
CCV1	CCV1				W	4/12/18					
Fluoride		PASS	2.048		2.05	103.0	90	110			mg/L
CCB1	CCB1				W	4/12/18					
Fluoride		PASS	0.000		0.000			+/-0.0430			mg/L
LB94633BLS	LB94633BLS				S	4/12/18					
Fluoride		PASS	0.000		0.000			+/-0.0170			mg/Kg
LB94633BSS	LB94633BSS				S	4/12/18					
Fluoride		PASS	2.040		40.70	102.0	90.00	110.00			mg/Kg
J2133-01	LOD-MDL-SOIL-01-QT2-2018	1			S	4/12/18					
Fluoride		PASS	0.111		1.11						mg/Kg
J2133-03	MDL-SOIL-03-QT2-2018	1			S	4/12/18					
Fluoride		PASS	0.104		1.04						mg/Kg
J2133-02	LOQ-SOIL-02-QT2-2018				S	4/12/18					
Fluoride		PASS	0.276		5.49	137.0	50.00	150.00			mg/Kg
CCV2	CCV2				W	4/12/18					
Fluoride		PASS	2.018		2.02	101.0	90	110			mg/L
CCB2	CCB2				W	4/12/18					
Fluoride		PASS	0.000		0.000			+/-0.0430			mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 4/16/18
 Analyst : AK
 Data File : LB94633.csv

Approved By : 
 Approved Date : 
 Worksheet # : 

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method			Line 2
Nitrate											
STD1	STD1				W	4/4/18					
Nitrate		PASS	0.000		0.000					mg/L	
STD2	STD2				W	4/4/18					
Nitrate		PASS	0.182		0.182					mg/L	
STD3	STD3				W	4/4/18					
Nitrate		PASS	0.522		0.522					mg/L	
STD4	STD4				W	4/4/18					
Nitrate		PASS	0.988		0.988					mg/L	
STD5	STD5				W	4/4/18					
Nitrate		PASS	2.383		2.38					mg/L	
STD6	STD6				W	4/4/18					
Nitrate		PASS	4.889		4.89					mg/L	
STD7	STD7				W	4/4/18					
Nitrate		PASS	6.364		6.36					mg/L	
ICV1	ICV1				W	4/4/18					
Nitrate		PASS	2.446		2.45	98.0	90	110		mg/L	
ICB1	ICB1				W	4/4/18					
Nitrate		PASS	0.000		0.000			+/-0.0270		mg/L	
CCV1	CCV1				W	4/12/18					
Nitrate		PASS	2.494		2.49	100.0	90	110		mg/L	
CCB1	CCB1				W	4/12/18					
Nitrate		PASS	0.000		0.000			+/-0.0270		mg/L	
LB94633BLS	LB94633BLS				S	4/12/18					
Nitrate		PASS	0.000		0.000			+/-0.0110		mg/Kg	
LB94633BSS	LB94633BSS				S	4/12/18					
Nitrate		PASS	2.520		50.30	101.0	90.00	110.00		mg/Kg	
J2133-01	LOD-MDL-SOIL-01-QT2-2018	1			S	4/12/18					
Nitrate		PASS	0.143		1.43					mg/Kg	
J2133-03	MDL-SOIL-03-QT2-2018	1			S	4/12/18					
Nitrate		PASS	0.143		1.43					mg/Kg	
J2133-02	LOQ-SOIL-02-QT2-2018				S	4/12/18					
Nitrate		PASS	0.320		6.36	127.0	50.00	150.00		mg/Kg	
CCV2	CCV2				W	4/12/18					
Nitrate		PASS	2.510		2.51	100.0	90	110		mg/L	
CCB2	CCB2				W	4/12/18					
Nitrate		PASS	0.000		0.000			+/-0.0270		mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 4/16/18
 Analyst : AK
 Data File : LB94633.csv


Approved By : SL
 Approved Date : 4/16/18
 Worksheet # : 2/10/18

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis		Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method		Line 2
Nitrite										
STD1	STD1				W	4/4/18				
Nitrite		PASS	0.000		0.000					mg/L
STD2	STD2				W	4/4/18				
Nitrite		PASS	0.245		0.245					mg/L
STD3	STD3				W	4/4/18				
Nitrite		PASS	0.626		0.626					mg/L
STD4	STD4				W	4/4/18				
Nitrite		PASS	1.182		1.18					mg/L
STD5	STD5				W	4/4/18				
Nitrite		PASS	2.854		2.85					mg/L
STD6	STD6				W	4/4/18				
Nitrite		PASS	5.841		5.84					mg/L
STD7	STD7				W	4/4/18				
Nitrite		PASS	7.677		7.68					mg/L
ICV1	ICV1				W	4/4/18				
Nitrite		PASS	2.918		2.92	97.0	90	110		mg/L
ICB1	ICB1				W	4/4/18				
Nitrite		PASS	0.000		0.000			+/-0.0220		mg/L
CCV1	CCV1				W	4/12/18				
Nitrite		PASS	2.973		2.97	99.0	90	110		mg/L
CCB1	CCB1				W	4/12/18				
Nitrite		PASS	0.000		0.000			+/-0.0220		mg/L
LB94633BLS	LB94633BLS				S	4/12/18				
Nitrite		PASS	0.000		0.000			+/-0.0100		mg/Kg
LB94633BSS	LB94633BSS				S	4/12/18				
Nitrite		PASS	2.978		59.40	99.0	90.00	110.00		mg/Kg
J2133-01	LOD-MDL-SOIL-01-QT2-2018	1			S	4/12/18				
Nitrite		PASS	0.207		2.07					mg/Kg
J2133-03	MDL-SOIL-03-QT2-2018	1			S	4/12/18				
Nitrite		PASS	0.210		2.10					mg/Kg
J2133-02	LOQ-SOIL-02-QT2-2018				S	4/12/18				
Nitrite		PASS	0.401		7.97	133.0	50.00	150.00		mg/Kg
CCV2	CCV2				W	4/12/18				
Nitrite		PASS	2.923		2.92	97.0	90	110		mg/L
CCB2	CCB2				W	4/12/18				
Nitrite		PASS	0.000		0.000			+/-0.0220		mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 4/16/18
 Analyst : AK
 Data File : LB94633.csv

Approved By : 
 Approved Date : 
 Worksheet # : 

Lab Sample ID	Client ID	Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method	Line 1
Parameter			Final Conc	%Rec		LCL	UCL	Line 2
Sulfate								
STD1	STD1			W	4/4/18			
Sulfate		PASS	0.000	0.000				mg/L
STD2	STD2			W	4/4/18			
Sulfate		PASS	1.326	1.33				mg/L
STD3	STD3			W	4/4/18			
Sulfate		PASS	3.154	3.15				mg/L
STD4	STD4			W	4/4/18			
Sulfate		PASS	5.810	5.81				mg/L
STD5	STD5			W	4/4/18			
Sulfate		PASS	14.046	14.0				mg/L
STD6	STD6			W	4/4/18			
Sulfate		PASS	29.205	29.2				mg/L
STD7	STD7			W	4/4/18			
Sulfate		PASS	38.335	38.3				mg/L
ICV1	ICV1			W	4/4/18			
Sulfate		PASS	15.115	15.10	101.0	90	110	mg/L
ICB1	ICB1			W	4/4/18			
Sulfate		PASS	0.000	0.000			+/-0.1320	mg/L
CCV1	CCV1			W	4/12/18			
Sulfate		PASS	14.522	14.50	97.0	90	110	mg/L
CCB1	CCB1			W	4/12/18			
Sulfate		PASS	0.000	0.000			+/-0.1320	mg/L
LB94633BLS	LB94633BLS			S	4/12/18			
Sulfate		PASS	0.000	0.000			+/-0.0960	mg/Kg
LB94633BSS	LB94633BSS			S	4/12/18			
Sulfate		PASS	15.322	306.00	102.0	90.00	110.00	mg/Kg
J2133-01	LOD-MDL-SOIL-01-QT2-2018		1	S	4/12/18			
Sulfate		PASS	1.099	11.0				mg/Kg
J2133-03	MDL-SOIL-03-QT2-2018		1	S	4/12/18			
Sulfate		PASS	1.095	11.0				mg/Kg
J2133-02	LOQ-SOIL-02-QT2-2018			S	4/12/18			
Sulfate		PASS	2.043	40.60	135.0	50.00	150.00	mg/Kg
CCV2	CCV2			W	4/12/18			
Sulfate		PASS	14.355	14.40	96.0	90	110	mg/L
CCB2	CCB2			W	4/12/18			
Sulfate		PASS	0.000	0.000			+/-0.1320	mg/L