

Daily Analysis Runlog For Sequence/QC Batch ID # LB94640

Review By		Review On				
SubDirectory		LB94640		Test : TOC		

STD. NAME	STD REF.#
ICAL Standard	WP63327,WP63328,WP63329,WP63330,WP63331,WP63332,WP63333,WP63334,WP63335
ICV Standard	WP63336
CCV Standard	WP63334
ICSA Standard	N/A
CRI Standard	N/A
LCS Standard	WP63336
Chk Standard	wp62726

Sr#	SampleID	ClientID	QcType	Date	Comment	Status
1	0.0PPM	0.0PPM	CAL1	04/11/18 19:05		
2	0.5PPM	0.5PPM	CAL2	04/11/18 19:18		
3	1.0PPM	1.0PPM	CAL3	04/11/18 19:43		
4	2.0PPM	2.0PPM	CAL4	04/11/18 19:56		
5	5.0PPM	5.0PPM	CAL5	04/11/18 20:09		
6	10.0PPM	10.0PPM	CAL6	04/11/18 20:23		
7	20.0PPM	20.0PPM	CAL7	04/11/18 20:37		
8	ICV1	ICV1	ICV	04/11/18 20:51		
9	ICB1	ICB1	ICB	04/11/18 21:04		
10	CCV1	CCV1	CCV	04/16/18 14:05		
11	CCB1	CCB1	CCB	04/16/18 14:18		
12	LB94640BLW	LB94640BLW	MB	04/16/18 14:54		
13	LB94640BSW	LB94640BSW	LCS	04/16/18 15:22		
14	J2133-08	LOQ-WATER-02-QT2	LOQ	04/16/18 15:47		
15	J2133-07	LOD-MDL-WATER-0	SAM	04/16/18 16:11		
16	J2133-09	MDL-WATER-03-QT2	SAM	04/16/18 16:35		
17	CCV2	CCV2	CCV	04/16/18 16:50		
18	CCB2	CCB2	CCB	04/16/18 17:03		



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/29/2018 12:00:00 AM
Project:	LB94640	Date Received:	3/29/2018 12:00:00 AM
Client Sample ID:	LOD-MDL-WATER-01-QT2-2018	SDG No.:	LB94640
Lab Sample ID:	J2133-07	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
10-35-5	TOC	0.44	J	1	0.097		0.5	mg/L		04/16/2018	9060



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/29/2018 12:00:00 AM
Project:	LB94640	Date Received:	3/29/2018 12:00:00 AM
Client Sample ID:	MDL-WATER-03-QT2-2018	SDG No.:	LB94640
Lab Sample ID:	J2133-09	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
10-35-5	TOC	0.38	J	1	0.097		0.5	mg/L		04/16/2018	9060



Report of Analysis

Client:		Date Collected:	3/29/2018 12:00:00 AM
Project:	LB94640	Date Received:	3/29/2018 12:00:00 AM
Client Sample ID:	MDL-WATER-03-QT2-2018	SDG No.:	LB94640
Lab Sample ID:	J2133-09	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
-----	-----------	-------	------	----	-----	-----	------------	-------	-----------	-----------	----------

Color Before:

Clarity Before:

Texture:

Color After:

Clarity After:

Artifacts:

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB94640

SDG No.: LB94640

Contract: _____ Lab Code: CHEM

Case No.: LB94640

SAS No.: LB94640

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	TOC	10.4	10	104		90 - 110		04/16/2018	14:05	1b94640
CCV2	TOC	10.6	10	106		90 - 110		04/16/2018	16:50	1b94640



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB94640

SDG No.: LB94640

Contract:

Lab Code: CHEM

Case No.: LB94640

SAS No.: LB94640

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	TOC	0.41	+/-0.5	J			0.5	04/11/2018	21:04	1b94640

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	TOC	0.31	+/-0.5	J			0.5	04/16/2018	14:18	lb94640
CCB2	TOC	0.28	+/-0.5	J			0.5	04/16/2018	17:03	lb94640



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB94640

SDG No.: LB94640

Contract: Lab Code: CHEM

Case No.: LB94640

SAS No.: LB94640

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



GENCHEM
- 3b -
PREPARATION BLANK SUMMARY

Client: LB94640

SDG No.: LB94640

Instrument: Appolo-9000

Sample ID	Analyte	Result (mg/L)	Acceptance Limit	Conc Qual	LOD mg/L	CRQL mg/L	M	Analysis Date	Analysis Time	Run
LB94640BLW	TOC	WATER 0.16	<0.5	Batch Number: J		0.5		Prep Date: 04/16/2018	14:54	lb94640



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB94640

SDG No.: LB94640

Contract:

Lab Code: CHEM

Case No.: LB94640

SAS No.: LB94640

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
J2133-08 TOC	mg/L	1	1.3		130	50 - 150	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB94640

SDG No.: LB94640

Contract:

Lab Code: CHEM

Case No.: LB94640

SAS No.: LB94640

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
LB94640BSW TOC	mg/L	10	10.6		106	90 - 110	