

IC-1 300.0

4-17-18 AK


Clear table

ident	concentrati on F-	concentrati on CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2018-03-19_	3/19/18 11:39
STD2	0.1500	0.2550	0.2370	0.8930	0.2300	0.1560	1.3110	_2018-03-19_	3/19/18 12:11
STD3	0.3910	0.6370	0.6410	2.1570	0.5450	0.9160	3.1970	_2018-03-19_	3/19/18 12:43
STD4	0.7810	1.1970	1.1780	3.8990	0.9790	2.0960	5.8990	_2018-03-19_	3/19/18 13:15
STD5	1.9910	2.8560	2.9190	9.4060	2.3260	5.1130	14.1490	_2018-03-19_	3/19/18 13:47
STD6	3.9180	5.8250	5.7830	19.7170	4.9370	10.1390	29.5500	_2018-03-19_	3/19/18 14:19
STD7	5.0690	7.6800	7.6910	25.4270	6.3580	12.3300	38.1440	_2018-03-19_	3/19/18 14:51
ICV	1.9800	2.9080	2.9290	9.6630	2.3880	5.2610	15.2060	_2018-03-19_	3/19/18 16:27
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2018-03-19_	3/19/18 17:23
CCV	1.9890	2.9140	2.7910	9.5390	2.3580	4.8590	14.4470	_2018-04-16_	4/16/18 8:50
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2018-04-16_	4/16/18 9:22
LB94662BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2018-04-16_	4/16/18 9:53
LB94662BSW	1.8320	2.7760	2.7270	9.2900	2.2890	5.0940	14.6130	_2018-04-16_	4/16/18 10:57
J2383-01	0.0000	460.8000	0.0000	0.0000	0.0000	0.0000	438.1070	_2018-04-16_	4/16/18 11:30
J2383-02	0.2890	893.8350	0.0000	1.4190	2.8830	0.0000	359.3080	_2018-04-16_	4/16/18 12:01
J2383-03	0.0000	393.7910	0.0000	0.0000	0.0000	0.0000	6.4670	_2018-04-16_	4/16/18 12:32
J2383-03DUP	0.0000	386.1130	0.0000	0.0000	0.0000	0.0000	6.4150	_2018-04-16_	4/16/18 13:03
J2383-03MS	2.7290	357.6850	2.5920	9.8880	2.3940	6.4210	21.3440	_2018-04-16_	4/16/18 13:34
J2383-01DLX20	0.0000	16.9520	0.0000	0.0000	0.0000	0.0000	16.3680	_2018-04-16_	4/16/18 14:05
J2383-01DL2X100	0.0000	3.1960	0.0000	0.0000	0.0000	0.0000	3.6190	_2018-04-16_	4/16/18 14:36
J2383-02DLX10	0.0000	77.9670	0.0000	0.0000	0.3640	0.0000	28.5840	_2018-04-16_	4/16/18 15:06
CCV	1.7590	2.9180	2.7610	9.5820	2.3690	5.2570	14.4710	_2018-04-16_	4/16/18 15:37
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2018-04-16_	4/16/18 16:08
J2383-02DL2X200	0.0000	2.9250	0.0000	0.0000	0.0000	0.0000	1.8810	_2018-04-16_	4/16/18 16:39
CCV	1.8890	2.9500	2.7910	9.6710	2.4440	5.2690	14.7820	_2018-04-16_	4/16/18 17:41
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2018-04-16_	4/16/18 18:12
CCV	1.9140	2.9270	2.8000	9.5280	2.3680	4.9100	14.5670	_2018-04-17_	4/17/18 10:34
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2018-04-17_	4/17/18 11:04
LB94662BLW2	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2018-04-17_	4/17/18 11:35
LB94662BSW2	1.8240	2.8230	2.7270	9.3550	2.3080	4.9890	14.7550	_2018-04-17_	4/17/18 12:06

IC-1 3-19-18
cal Ak

CALIBRATION OF COMPONENT F-

Method: ANIONS 03-19-18.mtw
Equation: $Q = 0.827175 \cdot A + 2.04637$
RSD: 2.922 %
Correlation coefficient: 0.999651



K3 = 0 K2 = 0 K1 = 0.827175 K0 = 2.04637


Base: Area
Ref.channel: Cond
ISTD:
Formula: Linear
Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
-------	--------	------	-------	---------	-----------	------	------

1	0	0	0	0	0	No
2	0.04538	1.141	0.1	20	5.921	Yes
3	0.2594	6.989	0.4	20	5.921	Yes
4	0.5635	16.41	0.8	20	5.921	Yes
5	1.518	45.66	2	20	5.921	Yes
6	3.234	92.27	4	20	5.921	Yes
7	4.105	120.1	5	20	5.921	Yes

CALIBRATION OF COMPONENT CL-

Method: ANIONS 03-19-18.mtw
 Equation: $Q = 1.21583 \cdot A + 2.13837$
 RSD: 5.049 %
 Correlation coefficient: 0.998959




K3 = 0 K2 = 0 K1 = 1.21583 K0 = 2.13837

Base: Area
 Ref.channel: Cond
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0	No	
2	0.1382	2.432	0.15	20	8.982	Yes	
3	0.5252	8.718	0.6	20	8.982	Yes	
4	1.075	17.94	1.2	20	8.982	Yes	
5	2.756	45.22	3	20	8.982	Yes	
6	5.764	94.06	6	20	8.982	Yes	
7	7.49	124.6	7.5	20	8.982	Yes	

CALIBRATION OF COMPONENT NO2

Method: ANIONS 03-19-18.mtw
 Equation: $Q = 0.51376 \cdot A + 2.60348$
 RSD: 5.143 %
 Correlation coefficient: 0.998920




K3 = 0 K2 = 0 K1 = 0.51376 K0 = 2.60348

Base: Area
 Ref.channel: Cond
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File	
1	0	0	0	0	0			No
2	0.21	4.156	0.15	20	10.72	Yes		
3	1.035	19.9	0.6	20	10.72	Yes		
4	2.13	40.8	1.2	20	10.72	Yes		
5	5.487	108.6	3	20	10.72	Yes		
6	11.13	220	6	20	10.72	Yes		
7	14.24	294.3	7.5	20	10.72	Yes		

CALIBRATION OF COMPONENT BR-

Method: ANIONS 03-19-18.mtw
 Equation: $Q = 2.68494 \cdot A + 9.86212$
 RSD: 4.375 %
 Correlation coefficient: 0.999218




K3 = 0 K2 = 0 K1 = 2.68494 K0 = 9.86212

Base: Area
 Ref.channel: Cond
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File	
1	0	0	0	0	0			No
2	0.1528	2.981	0.5	20	13.6			Yes
3	0.6454	12.39	2	20	13.6			Yes
4	1.339	25.37	4	20	13.6			Yes
5	3.519	66.39	10	20	13.6			Yes
6	7.35	143.2	20	20	13.6			Yes
7	9.434	185.7	25	20	13.6			Yes

CALIBRATION OF COMPONENT NO3

Method: ANIONS 03-19-18.mtw
 Equation: $Q = 0.473492 \cdot A + 2.64071$
 RSD: 4.760 %
 Correlation coefficient: 0.999075




K3 = 0 K2 = 0 K1 = 0.473492 K0 = 2.64071

Base: Area
 Ref.channel: Cond
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File	
1	0	0	0	0	0			No
2	0.1886	4.149	0.125	20	15.49			Yes
3	0.7944	17.44	0.5	20	15.49			Yes
4	1.643	35.76	1	20	15.49			Yes
5	4.336	92.66	2.5	20	15.49			Yes
6	9.344	203	5	20	15.49			Yes
7	12.2	263	6.25	20	15.49			Yes

CALIBRATION OF COMPONENT HPO4

Method: ANIONS 03-19-18.mtw
 Equation: $Q = 0.841758 \cdot A - 4.5807$
 RSD: 2.868 %
 Correlation coefficient: 0.999664




K3 = 0 K2 = 0 K1 = 0.841758 K0 = -4.5807

Base: Area
 Ref.channel: Cond
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File	
1	0		0	0	0			No
2	0.2149	9.149	0.25	20	22.72			Yes
3	0.6954	27.2	1	20	22.72			Yes
4	1.377	55.25	2	20	22.72			Yes
5	3.348	126.9	5	20	22.72			Yes
6	6.804	246.3	10	20	22.72			Yes
7	8.626	298.4	12.5	20	22.72			Yes

CALIBRATION OF COMPONENT SO4

Method: ANIONS 03-19-18.mtw
 Equation: $Q = 1.55224 \cdot A + 13.2199$
 RSD: 4.247 %
 Correlation coefficient: 0.999263



K3 = 0 K2 = 0 K1 = 1.55224 K0 = 13.2199
 Base: Area
 Ref.channel: Cond
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		No
2	0.2644	8.376	0.75	20	25.7		Yes
3	1.07	32.68	3	20	25.7		Yes
4	2.205	67.48	6	20	25.7		Yes
5	5.865	173.8	15	20	25.7		Yes
6	12.67	372.2	30	20	25.7		Yes
7	16.58	483	37.5	20	25.7		Yes

Report date: 3/19/2018 4:51:12 PM
Printed by: wet


Ident: STD1
Analysis from: 3/19/2018 11:39:46 AM
File: _2018-03-19_

Last save: 3/19/2018 4:49:08 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61448

Last save: 3/19/2018 9:52:10 AM

SAMPLE:
:
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 3/19/2018 4:51:19 PM
Printed by: wet


Ident: STD2
Analysis from: 3/19/2018 12:11:40 PM
File: _2018-03-19_

Last save: 3/19/2018 4:50:02 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61449

Last save: 3/19/2018 9:52:10 AM

SAMPLE:
: AK/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	0.00	0.000	0.00	0.00	0.000	0.00	0.
2	8.74	0.250	0.14	11.78	2.432	7.78	0.
3	10.49	0.286	0.21	17.92	4.156	13.30	0.
4	13.43	0.301	0.15	13.00	2.981	9.54	0.
5	15.38	0.332	0.19	16.07	4.149	13.28	0.
6	21.72	0.523	0.21	18.31	9.149	29.28	0.
7	25.21	0.468	0.26	22.57	8.376	26.81	0.
7	29.00	0.309	1.16	99.65	31.243	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 3/19/2018 4:52:13 PM
Printed by: wet

Ident: STD2
Analysis from: 3/19/2018 12:11:40 PM
File: _2018-03-19_

Last save: 3/19/2018 4:52:12 PM


Manual peaks!

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61449

Last save: 3/19/2018 9:52:10 AM

SAMPLE:

: AK/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.74	0.369	0.05	3.74	1.141	3.52	0.
2	8.74	0.250	0.14	11.34	2.432	7.51	0.
3	10.49	0.286	0.21	17.25	4.156	12.83	0.
4	13.43	0.301	0.15	12.51	2.981	9.20	0.
5	15.38	0.332	0.19	15.47	4.149	12.81	0.
6	21.72	0.523	0.21	17.63	9.149	28.25	0.
7	25.21	0.468	0.26	21.73	8.376	25.86	0.
7	29.00	0.361	1.21	99.67	32.384	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 3/19/2018 5:16:31 PM
Printed by: wet


Ident: STD3
Analysis from: 3/19/2018 12:43:34 PM
File: _2018-03-19_

Last save: 3/19/2018 5:16:16 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61450

Last save: 3/19/2018 9:52:10 AM

SAMPLE:
:
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.79	0.378	0.26	5.15	6.989	5.58	0.
2	8.72	0.248	0.52	10.44	8.718	6.96	0.
3	10.46	0.282	1.03	20.57	19.904	15.88	0.
4	13.36	0.289	0.64	12.83	12.392	9.89	0.
5	15.29	0.326	0.79	15.80	17.438	13.92	0.
6	21.70	0.505	0.70	13.83	27.196	21.70	0.
7	25.21	0.454	1.07	21.29	32.680	26.08	0.
7	29.00	0.355	5.02	99.92	125.316	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 3/19/2018 5:16:37 PM
Printed by: wet


Ident: STD4
Analysis from: 3/19/2018 1:15:28 PM
File: _2018-03-19_

Last save: 3/19/2018 5:16:16 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61451

Last save: 3/19/2018 9:52:10 AM

SAMPLE:
: AK/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.81	0.391	0.56	5.44	16.409	6.33	0.
2	8.73	0.246	1.08	10.41	17.939	6.93	0.
3	10.45	0.280	2.13	20.61	40.803	15.75	0.
4	13.35	0.285	1.34	12.96	25.373	9.80	0.
5	15.27	0.323	1.64	15.89	35.765	13.81	0.
6	21.71	0.507	1.38	13.32	55.249	21.33	0.
7	25.25	0.458	2.21	21.34	67.485	26.05	0.
7	29.00	0.356	10.33	99.98	259.022	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 3/19/2018 5:16:44 PM
Printed by: wet


Ident: STD5
Analysis from: 3/19/2018 1:47:22 PM
File: _2018-03-19_

Last save: 3/19/2018 5:16:16 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61452

Last save: 3/19/2018 9:52:10 AM

SAMPLE:
: AK/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.81	0.382	1.52	5.65	45.665	6.93	0.
2	8.72	0.244	2.76	10.27	45.218	6.86	0.
3	10.44	0.291	5.49	20.45	108.582	16.47	0.
4	13.35	0.283	3.52	13.12	66.393	10.07	0.
5	15.23	0.315	4.34	16.16	92.663	14.06	0.
6	21.68	0.486	3.35	12.48	126.933	19.25	0.
7	25.21	0.443	5.86	21.86	173.784	26.36	0.
7	29.00	0.349	26.83	99.99	659.239	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 3/19/2018 5:16:51 PM
Printed by: wet

Ident: STD6
Analysis from: 3/19/2018 2:19:16 PM
File: _2018-03-19_


Last save: 3/19/2018 5:16:16 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61453

Last save: 3/19/2018 9:52:10 AM

SAMPLE:

: AK/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.83	0.371	3.23	5.74	92.268	6.73	0.
2	8.74	0.241	5.76	10.24	94.056	6.86	0.
3	10.45	0.290	11.13	19.77	220.042	16.05	0.
4	13.42	0.292	7.35	13.06	143.199	10.44	0.
5	15.31	0.316	9.34	16.60	202.969	14.80	0.
6	21.92	0.475	6.80	12.09	246.345	17.97	0.
7	25.51	0.439	12.67	22.50	372.225	27.15	0.
7	29.00	0.346	56.30	100.00	1371.105	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 3/19/2018 5:16:57 PM
Printed by: wet


Ident: STD7
Analysis from: 3/19/2018 2:51:11 PM
File: _2018-03-19_

Last save: 3/19/2018 5:16:16 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61454

Last save: 3/19/2018 9:52:10 AM

SAMPLE:
:
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.83	0.387	4.10	5.65	120.082	6.79	0.
2	8.77	0.245	7.49	10.31	124.582	7.04	0.
3	10.51	0.304	14.24	19.59	294.341	16.64	0.
4	13.53	0.296	9.43	12.98	185.734	10.50	0.
5	15.41	0.314	12.20	16.79	262.984	14.87	0.
6	21.95	0.457	8.63	11.87	298.393	16.87	0.
7	25.49	0.436	16.58	22.81	482.955	27.30	0.
7	29.00	0.349	72.68	99.99	1769.070	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 3/19/2018 6:46:42 PM
 Printed by: wet


Ident: ICV
 Analysis from: 3/19/2018 4:27:28 PM
 File: _2018-03-19_

Last save: 3/19/2018 6:46:38 PM

Method: ANIONS 03-19-18.mtw
 Run operator: wet
 Analysis number: 61457

Last save: 3/19/2018 9:52:10 AM

SAMPLE:
 : AK/AP
 Vial number: 11
 Volume: 20.0 µL
 Dilution: 1.00
 Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.83	0.389	1.56	5.64	45.402	6.66	0.
2	8.75	0.246	2.78	10.02	46.071	6.76	0.
3	10.48	0.285	5.60	20.21	108.955	15.98	0.
4	13.42	0.288	3.56	12.84	68.309	10.02	0.
5	15.32	0.319	4.39	15.85	95.307	13.98	0.
6	21.66	0.482	3.49	12.60	130.436	19.13	0.
7	25.19	0.443	6.33	22.84	187.404	27.48	0.
7	29.00	0.350	27.70	100.00	681.883	100.00	0.

This report has been created by IC Net
 METROHM LTD

Report date: 3/19/2018 6:46:49 PM
Printed by: wet


Ident: ICB
Analysis from: 3/19/2018 5:23:20 PM
File: _2018-03-19_

Last save: 3/19/2018 6:46:38 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61459

Last save: 3/19/2018 9:52:10 AM

SAMPLE:
: AK/AP
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

J2383-03DLX100	0.0000	2.7600	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2018-04-17	4/17/18 12:38
CCV	1.9830	2.9720	2.8510	9.8120	2.4220	5.0920	14.6280	14.6280	14.6280	2018-04-17	4/17/18 13:09
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2018-04-17	4/17/18 13:39

Report date: 4/17/2018 9:13:54 AM
Printed by: wet


Ident: CCV
Analysis from: 4/16/2018 8:50:43 AM
File: _2018-04-16_

Last save: 4/16/2018 9:18:32 AM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61907

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.86	0.501	1.19	4.52	45.613	6.93	0.
2	8.81	0.244	2.81	10.69	46.174	7.01	0.
3	10.60	0.286	5.33	20.30	103.577	15.73	0.
4	13.64	0.295	3.41	12.99	67.382	10.23	0.
5	15.62	0.329	4.19	15.96	94.035	14.28	0.
6	21.44	0.482	3.19	12.14	120.899	18.36	0.
7	25.10	0.440	6.04	23.03	177.627	26.97	0.
7	28.00	0.368	26.14	99.64	655.307	99.51	0.

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 9:14:11 AM
Printed by: wet


Ident: CCB
Analysis from: 4/16/2018 9:22:50 AM
File: _2018-04-16_

Last save: 4/17/2018 9:14:10 AM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61908

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
:
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 9:14:27 AM
Printed by: wet


Ident: LB94662BLW
Analysis from: 4/16/2018 9:53:45 AM
File: _2018-04-16_

Last save: 4/17/2018 9:14:26 AM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61909

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
:
Vial number: 4
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 9:14:45 AM
Printed by: wet


Ident: LB94662BSW
Analysis from: 4/16/2018 10:57:45 AM
File: _2018-04-16_

Last save: 4/17/2018 9:13:42 AM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61911

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.90	0.461	1.19	4.56	41.815	6.40	0.
2	8.78	0.243	2.68	10.26	43.900	6.72	0.
3	10.52	0.288	5.16	19.75	101.105	15.49	0.
4	13.51	0.292	3.36	12.87	65.526	10.04	0.
5	15.41	0.319	4.19	16.02	91.109	13.95	0.
6	21.67	0.483	3.34	12.78	126.466	19.37	0.
7	25.35	0.442	6.12	23.40	179.765	27.53	0.
7	28.00	0.361	26.04	99.65	649.686	99.51	0.

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 9:15:31 AM
Printed by: wet


Ident: J2383-01
Analysis from: 4/16/2018 11:30:36 AM
File: _2018-04-16_

Last save: 4/17/2018 9:15:26 AM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61912

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 31
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	0.00	0.000	0.00	0.00	0.000	0.00	0.
2	8.99	0.421	289.19	62.89	7578.257	57.35	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	0.00	0.000	0.00	0.00	0.000	0.00	0.
6	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	25.14	0.504	170.67	37.11	5636.318	42.65	0.
7	28.00	0.132	459.86	100.00	13214.575	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 9:16:24 AM
Printed by: wet


Ident: J2383-02
Analysis from: 4/16/2018 12:01:31 PM
File: _2018-04-16_

Last save: 4/17/2018 9:16:12 AM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61913

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 32
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.86	0.649	0.10	0.02	4.502	0.02	0.
2	9.23	0.555	448.71	74.94	14701.555	75.59	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	13.55	0.235	0.45	0.07	6.897	0.04	0.
5	15.32	0.269	6.29	1.05	116.180	0.60	0.
6	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	25.17	0.493	143.07	23.89	4621.017	23.76	0.
7	28.00	0.315	598.63	99.97	19450.151	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 9:20:20 AM
Printed by: wet


Ident: J2383-03
Analysis from: 4/16/2018 12:32:25 PM
File: _2018-04-16_

Last save: 4/17/2018 9:20:20 AM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61914

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 33
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	0.00	0.000	0.00	0.00	0.000	0.00	0.
2	8.96	0.319	322.85	99.19	6475.978	98.86	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	0.00	0.000	0.00	0.00	0.000	0.00	0.
6	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	25.42	0.435	2.56	0.79	74.810	1.14	0.
7	28.00	0.108	325.41	99.98	6550.788	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 9:20:42 AM
Printed by: wet


Ident: J2383-03DUP
Analysis from: 4/16/2018 1:03:20 PM
File: _2018-04-16_

Last save: 4/17/2018 9:20:42 AM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61915

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 34
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	0.00	0.000	0.00	0.00	0.000	0.00	0.
2	8.97	0.317	317.55	99.16	6349.682	98.85	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	0.00	0.000	0.00	0.00	0.000	0.00	0.
6	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	25.58	0.448	2.49	0.78	74.139	1.15	0.
7	28.00	0.109	320.04	99.94	6423.822	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 9:23:01 AM
Printed by: wet


Ident: J2383-03MS
Analysis from: 4/16/2018 1:34:14 PM
File: _2018-04-16_

Last save: 4/17/2018 9:23:02 AM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61916

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 35
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.88	0.779	1.32	0.40	63.513	0.96	0.
2	9.00	0.312	298.52	91.54	5882.059	88.70	0.
3	10.58	0.311	3.67	1.12	95.843	1.45	0.
4	13.60	0.234	4.47	1.37	69.981	1.06	0.
5	15.46	0.286	4.88	1.50	95.562	1.44	0.
6	21.88	0.527	3.79	1.16	158.012	2.38	0.
7	25.54	0.430	9.33	2.86	266.498	4.02	0.
7	28.00	0.411	325.98	99.96	6631.467	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 9:27:41 AM
Printed by: wet


Ident: J2383-01DLX20
Analysis from: 4/16/2018 2:05:09 PM
File: _2018-04-16_

Last save: 4/16/2018 2:32:58 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61917

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 37
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	0.00	0.000	0.00	0.00	0.000	0.00	0.
2	8.83	0.230	17.82	71.96	277.090	57.42	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	0.00	0.000	0.00	0.00	0.000	0.00	0.
6	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	25.48	0.441	6.86	27.71	202.376	41.94	0.
7	28.00	0.096	24.68	99.68	479.465	99.36	0.

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 9:28:19 AM
Printed by: wet


Ident: J2383-01DL2X100
Analysis from: 4/16/2018 2:36:03 PM
File: _2018-04-16_

Last save: 4/16/2018 3:03:52 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61918

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 36
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	0.00	0.000	0.00	0.00	0.000	0.00	0.
2	8.76	0.241	3.01	70.52	50.821	57.15	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	0.00	0.000	0.00	0.00	0.000	0.00	0.
6	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	25.38	0.452	1.26	29.45	38.111	42.85	0.
7	28.00	0.099	4.27	99.97	88.932	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 9:30:04 AM
Printed by: wet


Ident: J2383-02DLX10
Analysis from: 4/16/2018 3:06:58 PM
File: _2018-04-16_

Last save: 4/16/2018 3:34:46 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61919

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 39
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	0.00	0.000	0.00	0.00	0.000	0.00	0.
2	8.82	0.232	82.80	86.53	1280.776	77.40	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	15.48	0.328	0.44	0.46	9.785	0.59	0.
6	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	25.35	0.436	12.34	12.89	359.772	21.74	0.
7	28.00	0.142	95.58	99.88	1650.333	99.74	0.

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 9:30:26 AM
Printed by: wet


Ident: CCV
Analysis from: 4/16/2018 3:37:53 PM
File: _2018-04-16_

Last save: 4/16/2018 4:05:42 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61920

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.92	0.465	1.19	4.45	40.045	6.05	0.
2	8.86	0.242	2.84	10.59	46.239	6.99	0.
3	10.59	0.283	5.35	19.97	102.426	15.47	0.
4	13.49	0.290	3.50	13.05	67.700	10.23	0.
5	15.40	0.320	4.35	16.22	94.484	14.27	0.
6	21.74	0.484	3.42	12.74	130.349	19.69	0.
7	25.42	0.440	6.08	22.68	177.943	26.88	0.
7	28.00	0.360	26.72	99.69	659.186	99.59	0.

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 9:30:44 AM
Printed by: wet


Ident: CCB
Analysis from: 4/16/2018 4:08:47 PM
File: _2018-04-16_

Last save: 4/17/2018 9:30:42 AM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61921

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 9:31:34 AM
Printed by: wet


Ident: J2383-02DL2X200
Analysis from: 4/16/2018 4:39:42 PM
File: _2018-04-16_

Last save: 4/16/2018 5:07:30 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61922

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 38
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	0.00	0.000	0.00	0.00	0.000	0.00	0.
2	8.79	0.241	2.85	82.57	46.356	71.20	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	0.00	0.000	0.00	0.00	0.000	0.00	0.
6	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	25.45	0.454	0.52	15.14	15.716	24.14	0.
7	28.00	0.099	3.38	97.71	62.072	95.34	0.

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 9:37:50 AM
Printed by: wet


Ident: CCV
Analysis from: 4/16/2018 5:41:31 PM
File: _2018-04-16_

Last save: 4/16/2018 6:09:20 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61924

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 41
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.89	0.482	1.19	4.42	43.192	6.43	0.
2	8.77	0.242	2.86	10.60	46.760	6.96	0.
3	10.48	0.282	5.42	20.10	103.580	15.41	0.
4	13.38	0.290	3.53	13.10	68.366	10.17	0.
5	15.31	0.326	4.40	16.33	97.657	14.53	0.
6	21.69	0.486	3.43	12.70	130.632	19.44	0.
7	25.39	0.446	6.14	22.75	181.938	27.07	0.
7	28.00	0.365	26.98	99.99	672.125	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 9:38:01 AM
Printed by: wet


Ident: CCB
Analysis from: 4/16/2018 6:12:25 PM
File: _2018-04-16_

Last save: 4/16/2018 6:40:14 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61925

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 42
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	0.00	0.000	0.00	0.00	0.000	0.00	0.
2	0.00	0.000	0.00	0.00	0.000	0.00	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	0.00	0.000	0.00	0.00	0.000	0.00	0.
6	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	28.00	0.000	0.00	0.00	0.000	0.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 12:35:16 PM
Printed by: wet


Ident: CCV
Analysis from: 4/17/2018 10:34:05 AM
File: _2018-04-17_

Last save: 4/17/2018 11:01:54 AM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61937

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.84	0.507	1.15	4.33	43.799	6.63	0.
2	8.74	0.244	2.81	10.60	46.388	7.03	0.
3	10.49	0.282	5.41	20.38	103.916	15.74	0.
4	13.46	0.291	3.45	13.01	67.300	10.19	0.
5	15.40	0.325	4.26	16.06	94.432	14.30	0.
6	21.28	0.482	3.27	12.31	122.103	18.49	0.
7	24.72	0.443	6.09	22.96	179.177	27.14	0.
7	28.00	0.368	26.44	99.65	657.115	99.53	0.

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 12:35:33 PM
Printed by: wet


Ident: CCB
Analysis from: 4/17/2018 11:04:59 AM
File: _2018-04-17_

Last save: 4/17/2018 12:35:32 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61938

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 12:35:43 PM
Printed by: wet


Ident: LB94662BLW2
Analysis from: 4/17/2018 11:35:54 AM
File: _2018-04-17_

Last save: 4/17/2018 12:03:42 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61939

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	0.00	0.000	0.00	0.00	0.000	0.00	0.
2	0.00	0.000	0.00	0.00	0.000	0.00	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	0.00	0.000	0.00	0.00	0.000	0.00	0.
6	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	28.00	0.000	0.00	0.00	0.000	0.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 12:35:53 PM
Printed by: wet


Ident: LB94662BSW2
Analysis from: 4/17/2018 12:06:48 PM
File: _2018-04-17_

Last save: 4/17/2018 12:34:38 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61940

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.84	0.484	1.15	4.33	41.639	6.37	0.
2	8.70	0.242	2.73	10.26	44.684	6.83	0.
3	10.42	0.279	5.33	20.06	101.110	15.46	0.
4	13.33	0.290	3.42	12.86	66.008	10.09	0.
5	15.22	0.318	4.25	16.01	91.923	14.06	0.
6	21.35	0.481	3.32	12.49	123.983	18.96	0.
7	24.79	0.435	6.28	23.64	181.597	27.77	0.
7	28.00	0.361	26.47	99.65	650.944	99.55	0.

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 1:21:42 PM
Printed by: wet


Ident: J2383-03DLX100
Analysis from: 4/17/2018 12:38:08 PM
File: _2018-04-17_

Last save: 4/17/2018 1:05:56 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61941

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	0.00	0.000	0.00	0.00	0.000	0.00	0.
2	8.73	0.242	2.61	96.58	43.635	92.21	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	0.00	0.000	0.00	0.00	0.000	0.00	0.
6	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	28.00	0.035	2.61	96.58	43.635	92.21	0.

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 1:42:21 PM
Printed by: wet


Ident: CCV
Analysis from: 4/17/2018 1:09:03 PM
File: _2018-04-17_

Last save: 4/17/2018 1:36:52 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61942

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.83	0.472	1.23	4.48	45.477	6.75	0.
2	8.69	0.243	2.89	10.50	47.123	6.99	0.
3	10.40	0.281	5.54	20.15	105.937	15.72	0.
4	13.30	0.290	3.60	13.10	69.414	10.30	0.
5	15.18	0.318	4.48	16.29	96.706	14.35	0.
6	21.41	0.477	3.41	12.41	126.416	18.76	0.
7	24.87	0.433	6.25	22.73	179.962	26.70	0.
7	28.00	0.359	27.39	99.66	671.037	99.56	0.

This report has been created by IC Net
METROHM LTD

Report date: 4/17/2018 2:07:59 PM
Printed by: wet


Ident: CCB
Analysis from: 4/17/2018 1:39:57 PM
File: _2018-04-17_

Last save: 4/17/2018 2:07:46 PM

Method: ANIONS 03-19-18.mtw
Run operator: wet
Analysis number: 61943

Last save: 3/28/2018 6:04:32 PM

SAMPLE:
: AK/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD