



Chemtech Consulting Group

Analytical Review Report

Date Printed : 4/17/18
Analyst : AK
Data File : lb94663.csv

Approved By : 
Approved Date : 
Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units
										Line 2
Bromide										
STD1	STD1			W	3/19/18					
Bromide		PASS		0.000	0.000					mg/L
STD2	STD2			W	3/19/18					
Bromide		PASS		0.893	0.893					mg/L
STD3	STD3			W	3/19/18					
Bromide		PASS		2.157	2.16					mg/L
STD4	STD4			W	3/19/18					
Bromide		PASS		3.899	3.90					mg/L
STD5	STD5			W	3/19/18					
Bromide		PASS		9.406	9.41					mg/L
STD6	STD6			W	3/19/18					
Bromide		PASS		19.717	19.7					mg/L
STD7	STD7			W	3/19/18					
Bromide		PASS		25.427	25.4					mg/L
ICV1	ICV1			W	3/19/18					
Bromide		PASS		9.663	9.66	97.0	90	110		mg/L
ICB1	ICB1			W	3/19/18					
Bromide		PASS		0.000	0.000			+/-0.0660		mg/L
CCV1	CCV1			W	4/16/18					
Bromide		PASS		9.671	9.67	97.0	90	110		mg/L
CCB1	CCB1			W	4/16/18					
Bromide		PASS		0.000	0.000			+/-0.0660		mg/L
LB94663BSS	LB94663BSS			S	4/16/18					
Bromide		PASS		9.886	197.00	99.0	90.00	110.00		mg/Kg
LB94663BLS	LB94663BLS			S	4/16/18					
Bromide		PASS		0.000	0.000			+/-0.1900		mg/Kg
J2133-01	LOD-MDL-SOIL-01-QT2-2018		1	S	4/16/18					
Bromide		PASS		0.692	13.8					mg/Kg
J2133-03	MDL-SOIL-03-QT2-2018		1	S	4/16/18					
Bromide		PASS		0.698	14.0					mg/Kg
J2133-02	LOQ-SOIL-02-QT2-2018			S	4/16/18					
Bromide		PASS		1.357	27.10	136.0	50.00	150.00		mg/Kg
CCV2	CCV2			W	4/16/18					
Bromide		PASS		9.638	9.64	96.0	90	110		mg/L
CCB2	CCB2			W	4/16/18					
Bromide		PASS		0.000	0.000			+/-0.0660		mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 4/17/18
 Analyst : AK
 Data File : lb94663.csv

Approved By : 
 Approved Date :
 Worksheet # : 

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method	LCL	UCL	Line 2
Chloride											
STD1	STD1				W	3/19/18					
Chloride		PASS	0.000		0.000						mg/L
STD2	STD2				W	3/19/18					
Chloride		PASS	0.255		0.255						mg/L
STD3	STD3				W	3/19/18					
Chloride		PASS	0.637		0.637						mg/L
STD4	STD4				W	3/19/18					
Chloride		PASS	1.197		1.20						mg/L
STD5	STD5				W	3/19/18					
Chloride		PASS	2.856		2.86						mg/L
STD6	STD6				W	3/19/18					
Chloride		PASS	5.825		5.82						mg/L
STD7	STD7				W	3/19/18					
Chloride		PASS	7.680		7.68						mg/L
ICV1	ICV1				W	3/19/18					
Chloride		PASS	2.908		2.91	97.0	90	110			mg/L
ICB1	ICB1				W	3/19/18					
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L
CCV1	CCV1				W	4/16/18					
Chloride		PASS	2.950		2.95	98.0	90	110			mg/L
CCB1	CCB1				W	4/16/18					
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L
LB94663BSS	LB94663BSS				S	4/16/18					
Chloride		PASS	2.929		58.50	98.0	90.00	110.00			mg/Kg
LB94663BLS	LB94663BLS				S	4/16/18					
Chloride		PASS	0.000		0.000			+/-0.0320			mg/Kg
J2133-01	LOD-MDL-SOIL-01-QT2-2018	1			S	4/16/18					
Chloride		PASS	0.184		3.68						mg/Kg
J2133-03	MDL-SOIL-03-QT2-2018	1			S	4/16/18					
Chloride		PASS	0.183		3.66						mg/Kg
J2133-02	LOQ-SOIL-02-QT2-2018				S	4/16/18					
Chloride		PASS	0.377		7.54	126.0	50.00	150.00			mg/Kg
CCV2	CCV2				W	4/16/18					
Chloride		PASS	2.940		2.94	98.0	90	110			mg/L
CCB2	CCB2				W	4/16/18					
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 4/17/18
 Analyst : AK
 Data File : lb94663.csv



Approved By : 
 Approved Date : 4/18/18
 Worksheet # :

Lab Sample ID	Client ID	Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method	Line 1
Parameter			Final Conc	%Rec		LCL	UCL	Line 2
Fluoride								
STD1	STD1			W	3/19/18			
Fluoride		PASS	0.000	0.000				mg/L
STD2	STD2			W	3/19/18			
Fluoride		PASS	0.150	0.150				mg/L
STD3	STD3			W	3/19/18			
Fluoride		PASS	0.391	0.391				mg/L
STD4	STD4			W	3/19/18			
Fluoride		PASS	0.781	0.781				mg/L
STD5	STD5			W	3/19/18			
Fluoride		PASS	1.991	1.99				mg/L
STD6	STD6			W	3/19/18			
Fluoride		PASS	3.918	3.92				mg/L
STD7	STD7			W	3/19/18			
Fluoride		PASS	5.069	5.07				mg/L
ICV1	ICV1			W	3/19/18			
Fluoride		PASS	1.980	1.98	99.0	90	110	mg/L
ICB1	ICB1			W	3/19/18			
Fluoride		PASS	0.000	0.000			+/-0.0430	mg/L
CCV1	CCV1			W	4/16/18			
Fluoride		PASS	1.889	1.89	95.0	90	110	mg/L
CCB1	CCB1			W	4/16/18			
Fluoride		PASS	0.000	0.000			+/-0.0430	mg/L
LB94663BSS	LB94663BSS			S	4/16/18			
Fluoride		PASS	1.915	38.20	96.0	90.00	110.00	mg/Kg
LB94663BLS	LB94663BLS			S	4/16/18			
Fluoride		PASS	0.000	0.000			+/-0.0170	mg/Kg
J2133-01	LOD-MDL-SOIL-01-QT2-2018		1	S	4/16/18			
Fluoride		PASS	0.128	2.56				mg/Kg
J2133-03	MDL-SOIL-03-QT2-2018		1	S	4/16/18			
Fluoride		PASS	0.129	2.58				mg/Kg
J2133-02	LOQ-SOIL-02-QT2-2018			S	4/16/18			
Fluoride		PASS	0.185	3.70	93.0	50.00	150.00	mg/Kg
CCV2	CCV2			W	4/16/18			
Fluoride		PASS	1.921	1.92	96.0	90	110	mg/L
CCB2	CCB2			W	4/16/18			
Fluoride		PASS	0.000	0.000			+/-0.0430	mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 4/17/18
Analyst : AK
Data File : lb94663.csv


Approved By : 
Approved Date : 
Worksheet # : 

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis		Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method		Line 2
Nitrate										
STD1	STD1				W	3/19/18				
Nitrate		PASS	0.000		0.000				mg/L	
STD2	STD2				W	3/19/18				
Nitrate		PASS	0.230		0.230				mg/L	
STD3	STD3				W	3/19/18				
Nitrate		PASS	0.545		0.545				mg/L	
STD4	STD4				W	3/19/18				
Nitrate		PASS	0.979		0.979				mg/L	
STD5	STD5				W	3/19/18				
Nitrate		PASS	2.326		2.33				mg/L	
STD6	STD6				W	3/19/18				
Nitrate		PASS	4.937		4.94				mg/L	
STD7	STD7				W	3/19/18				
Nitrate		PASS	6.358		6.36				mg/L	
ICV1	ICV1				W	3/19/18				
Nitrate		PASS	2.388		2.39	96.0	90	110	mg/L	
ICB1	ICB1				W	3/19/18				
Nitrate		PASS	0.000		0.000			+/-0.0270	mg/L	
CCV1	CCV1				W	4/16/18				
Nitrate		PASS	2.444		2.44	98.0	90	110	mg/L	
CCB1	CCB1				W	4/16/18				
Nitrate		PASS	0.000		0.000			+/-0.0270	mg/L	
LB94663BSS	LB94663BSS				S	4/16/18				
Nitrate		PASS	2.477		49.40	99.0	90.00	110.00	mg/Kg	
LB94663BLS	LB94663BLS				S	4/16/18				
Nitrate		PASS	0.000		0.000			+/-0.0110	mg/Kg	
J2133-01	LOD-MDL-SOIL-01-QT2-2018	1			S	4/16/18				
Nitrate		PASS	0.177		3.54				mg/Kg	
J2133-03	MDL-SOIL-03-QT2-2018	1			S	4/16/18				
Nitrate		PASS	0.177		3.54				mg/Kg	
J2133-02	LOQ-SOIL-02-QT2-2018				S	4/16/18				
Nitrate		PASS	0.334		6.68	134.0	50.00	150.00	mg/Kg	
CCV2	CCV2				W	4/16/18				
Nitrate		PASS	2.381		2.38	95.0	90	110	mg/L	
CCB2	CCB2				W	4/16/18				
Nitrate		PASS	0.000		0.000			+/-0.0270	mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 4/17/18
 Analyst : AK
 Data File : lb94663.csv


Approved By : 
 Approved Date :
 Worksheet # : 

Lab Sample ID	Client ID	Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method	Line 1
Parameter			Final Conc	%Rec		LCL	UCL	Line 2
Nitrite								
STD1	STD1			W	3/19/18			
Nitrite		PASS	0.000	0.000				mg/L
STD2	STD2			W	3/19/18			
Nitrite		PASS	0.237	0.237				mg/L
STD3	STD3			W	3/19/18			
Nitrite		PASS	0.641	0.641				mg/L
STD4	STD4			W	3/19/18			
Nitrite		PASS	1.178	1.18				mg/L
STD5	STD5			W	3/19/18			
Nitrite		PASS	2.919	2.92				mg/L
STD6	STD6			W	3/19/18			
Nitrite		PASS	5.783	5.78				mg/L
STD7	STD7			W	3/19/18			
Nitrite		PASS	7.691	7.69				mg/L
ICV1	ICV1			W	3/19/18			
Nitrite		PASS	2.929	2.93	98.0	90	110	mg/L
ICB1	ICB1			W	3/19/18			
Nitrite		PASS	0.000	0.000			+/-0.0220	mg/L
CCV1	CCV1			W	4/16/18			
Nitrite		PASS	2.791	2.79	93.0	90	110	mg/L
CCB1	CCB1			W	4/16/18			
Nitrite		PASS	0.000	0.000			+/-0.0220	mg/L
LB94663BSS	LB94663BSS			S	4/16/18			
Nitrite		PASS	2.895	57.80	96.0	90.00	110.00	mg/Kg
LB94663BLS	LB94663BLS			S	4/16/18			
Nitrite		PASS	0.000	0.000			+/-0.0100	mg/Kg
J2133-01	LOD-MDL-SOIL-01-QT2-2018			S	4/16/18			
Nitrite		PASS	0.187	3.74				mg/Kg
J2133-03	MDL-SOIL-03-QT2-2018			S	4/16/18			
Nitrite		PASS	0.183	3.66				mg/Kg
J2133-02	LOQ-SOIL-02-QT2-2018			S	4/16/18			
Nitrite		PASS	0.374	7.48	125.0	50.00	150.00	mg/Kg
CCV2	CCV2			W	4/16/18			
Nitrite		PASS	2.803	2.80	93.0	90	110	mg/L
CCB2	CCB2			W	4/16/18			
Nitrite		PASS	0.000	0.000			+/-0.0220	mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 4/17/18
 Analyst : AK
 Data File : lb94663.csv


Approved By : 
 Approved Date : 
 Worksheet # :

Lab Sample ID	Client ID	Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method	Line 1
Parameter			Final Conc	%Rec		LCL	UCL	Line 2
Ortho Phosphate								
STD1	STD1			W	3/19/18			
Ortho Phosphate		PASS	0.000	0.000				mg/L
STD2	STD2			W	3/19/18			
Ortho Phosphate		PASS	0.156	0.156				mg/L
STD3	STD3			W	3/19/18			
Ortho Phosphate		PASS	0.916	0.916				mg/L
STD4	STD4			W	3/19/18			
Ortho Phosphate		PASS	2.096	2.10				mg/L
STD5	STD5			W	3/19/18			
Ortho Phosphate		PASS	5.113	5.11				mg/L
STD6	STD6			W	3/19/18			
Ortho Phosphate		PASS	10.139	10.1				mg/L
STD7	STD7			W	3/19/18			
Ortho Phosphate		PASS	12.330	12.3				mg/L
ICV1	ICV1			W	3/19/18			
Ortho Phosphate		PASS	5.261	5.26	105.0	90	110	mg/L
ICB1	ICB1			W	3/19/18			
Ortho Phosphate		PASS	0.000	0.000			+/-0.0100	mg/L
CCV1	CCV1			W	4/16/18			
Ortho Phosphate		PASS	5.269	5.27	105.0	90	110	mg/L
CCB1	CCB1			W	4/16/18			
Ortho Phosphate		PASS	0.000	0.000			+/-0.0100	mg/L
LB94663BSS	LB94663BSS			S	4/16/18			
Ortho Phosphate		PASS	5.053	101.00	101.0	90.00	110.00	mg/Kg
LB94663BLS	LB94663BLS			S	4/16/18			
Ortho Phosphate		PASS	0.000	0.000			+/-0.0020	mg/Kg
J2133-01	LOD-MDL-SOIL-01-QT2-2018			S	4/16/18			
Ortho Phosphate		PASS	0.077	1.54				mg/Kg
J2133-03	MDL-SOIL-03-QT2-2018			S	4/16/18			
Ortho Phosphate		PASS	0.062	1.24				mg/Kg
J2133-02	LOQ-SOIL-02-QT2-2018			S	4/16/18			
Ortho Phosphate		PASS	0.403	8.06	81.0	50.00	150.00	mg/Kg
CCV2	CCV2			W	4/16/18			
Ortho Phosphate		PASS	4.975	4.97	99.0	90	110	mg/L
CCB2	CCB2			W	4/16/18			
Ortho Phosphate		PASS	0.000	0.000			+/-0.0100	mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 4/17/18
 Analyst : AK
 Data File : lb94663.csv

Approved By : 
 Approved Date :
 Worksheet # :

Lab Sample ID	Client ID	Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method	Line 1
Parameter			Final Conc	%Rec		LCL	UCL	Line 2
Sulfate								
STD1	STD1			W	3/19/18			
Sulfate		PASS	0.000	0.000				mg/L
STD2	STD2			W	3/19/18			
Sulfate		PASS	1.311	1.31				mg/L
STD3	STD3			W	3/19/18			
Sulfate		PASS	3.197	3.20				mg/L
STD4	STD4			W	3/19/18			
Sulfate		PASS	5.899	5.90				mg/L
STD5	STD5			W	3/19/18			
Sulfate		PASS	14.149	14.1				mg/L
STD6	STD6			W	3/19/18			
Sulfate		PASS	29.550	29.6				mg/L
STD7	STD7			W	3/19/18			
Sulfate		PASS	38.144	38.1				mg/L
ICV1	ICV1			W	3/19/18			
Sulfate		PASS	15.206	15.20	101.0	90	110	mg/L
ICB1	ICB1			W	3/19/18			
Sulfate		PASS	0.000	0.000			+/-0.1320	mg/L
CCV1	CCV1			W	4/16/18			
Sulfate		PASS	14.782	14.80	99.0	90	110	mg/L
CCB1	CCB1			W	4/16/18			
Sulfate		PASS	0.000	0.000			+/-0.1320	mg/L
LB94663BSS	LB94663BSS			S	4/16/18			
Sulfate		PASS	15.298	305.00	102.0	90.00	110.00	mg/Kg
LB94663BLS	LB94663BLS			S	4/16/18			
Sulfate		PASS	0.000	0.000			+/-0.0960	mg/Kg
J2133-01	LOD-MDL-SOIL-01-QT2-2018		1	S	4/16/18			
Sulfate		PASS	1.050	21.0				mg/Kg
J2133-03	MDL-SOIL-03-QT2-2018		1	S	4/16/18			
Sulfate		PASS	1.053	21.1				mg/Kg
J2133-02	LOQ-SOIL-02-QT2-2018			S	4/16/18			
Sulfate		PASS	2.061	41.20	137.0	50.00	150.00	mg/Kg
CCV2	CCV2			W	4/16/18			
Sulfate		PASS	14.554	14.60	97.0	90	110	mg/L
CCB2	CCB2			W	4/16/18			
Sulfate		PASS	0.000	0.000			+/-0.1320	mg/L