

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/17/2018 12:31

Test: Reactive
~~Total~~ CN Ng/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	92.260	0.0	0.096	
ICB1	-0.522	0.0	0.004	
CCV1	235.971	0.0	0.239	
CCB1	-0.577	0.0	0.004	
PB108308BL	-0.934	0.0	0.003	
J2410-04	-0.396	0.0	0.004	
J2410-04DUP	-0.246	0.0	0.004	
J2410-08	-0.419	0.0	0.004	
J2411-03	0.059	0.0	0.004	
CCV2	243.754	0.0	0.246	
CCB2	-0.474	0.0	0.004	

N	11
Mean	51.680
SD	97.0740
CV%	187.84

Aquakem v. 7.2AQ1

Results from time period:

Tue Apr 17 12:02:01 2018

Tue Apr 17 12:02:11 2018

AS

Sample Id	Sam/Ctr/c	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-0.6176	µg/l	4/17/2018 9:52:44	
5.0PPBCN	A	Reactive CN	P	4.0269	µg/l	4/17/2018 9:52:45	
10PPBCN	A	Reactive CN	P	9.4045	µg/l	4/17/2018 9:52:46	
50PPBCN	A	Reactive CN	P	50.6368	µg/l	4/17/2018 9:52:47	
100PPBCN	A	Reactive CN	P	100.1988	µg/l	4/17/2018 9:52:48	
250PPBCN	A	Reactive CN	P	252.8648	µg/l	4/17/2018 9:52:49	
500PPBCN	A	Reactive CN	P	498.4858	µg/l	4/17/2018 9:52:50	
ICV1	S	Reactive CN	P	92.2604	µg/l	4/17/2018 12:02:01	
ICB1	S	Reactive CN	P	-0.5215	µg/l	4/17/2018 12:02:02	
CCV1	S	Reactive CN	P	235.9709	µg/l	4/17/2018 12:02:03	
CCB1	S	Reactive CN	P	-0.5769	µg/l	4/17/2018 12:02:04	
PB108308BL	S	Reactive CN	P	-0.9342	µg/l	4/17/2018 12:02:05	
J2410-04	S	Reactive CN	P	-0.3964	µg/l	4/17/2018 12:02:06	
J2410-04DUP	S	Reactive CN	P	-0.2457	µg/l	4/17/2018 12:02:07	
J2410-08	S	Reactive CN	P	-0.4192	µg/l	4/17/2018 12:02:08	
J2411-03	S	Reactive CN	P	0.0592	µg/l	4/17/2018 12:02:09	
CCV2	S	Reactive CN	P	243.7538	µg/l	4/17/2018 12:02:10	
CCB2	S	Reactive CN	P	-0.4744	µg/l	4/17/2018 12:02:11	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/17/2018 9:55


Test Total CN

Accepted 4/17/2018 9:55

Factor 1006
Bias 0.004

Coeff. of det. 0.999939

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.6176	0.0000	
2	5.0PPBCN	0.008	4.0269	5.0000	
3	10PPBCN	0.013	9.4045	10.0000	
4	50PPBCN	0.054	50.6368	50.0000	
5	100PPBCN	0.104	100.1988	100.0000	
6	250PPBCN	0.255	252.8648	250.0000	
7	500PPBCN	0.499	498.4858	500.0000	