

=====

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/3/2018 12:08

Test: ^{Reactive} ~~Total~~ CN Ugld

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.869	0.0	0.098	
ICB1	-1.160	0.0	0.003	
CCV1	242.367	0.0	0.242	
CCB1	-0.814	0.0	0.004	
PB108858BL	-0.505	0.0	0.004	
J2669-02	-0.119	0.0	0.004	
J2669-02DUP	-1.349	0.0	0.003	
J2669-04	-0.365	0.0	0.004	
J2670-02	-0.787	0.0	0.004	
J2679-02	-0.656	0.0	0.004	
J2679-04	-0.739	0.0	0.004	
J2679-06	-1.026	0.0	0.004	
J2681-02	-1.192	0.0	0.003	
J2681-04	-1.044	0.0	0.004	
CCV2	240.991	0.0	0.240	
CCB2	-0.990	0.0	0.004	
J2686-02	-0.729	0.0	0.004	
J2686-04	-0.662	0.0	0.004	
J2695-02	-1.600	0.0	0.003	
J2703-02	-0.933	0.0	0.004	
J2710-03	-0.958	0.0	0.004	
J2711-03	-0.813	0.0	0.004	
J2712-03	-1.326	0.0	0.003	
PB108859BL	-1.066	0.0	0.003	
J2703-04	-0.011	0.0	0.005	
J2703-04DUP	0.805	0.0	0.005	
CCV3	251.291	0.0	0.250	
CCB3	-0.773	0.0	0.004	

N 28
Mean 28.989
SD 78.3264
CV% 270.19

Aquakem v. 7.2AQ1

Results from time period:

Thu May 03 11:52:17 2018

Thu May 03 12:05:10 2018

AS

Sample Id	Sam/Ctr/c	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.7143	µg/l	5/3/2018 8:32:09	
5.OPPBCN	A	Reactive CN	P	3.8393	µg/l	5/3/2018 8:32:10	
10PPBCN	A	Reactive CN	P	9.6234	µg/l	5/3/2018 8:32:11	
50PPBCN	A	Reactive CN	P	50.3197	µg/l	5/3/2018 8:32:12	
100PPBCN	A	Reactive CN	P	100.9958	µg/l	5/3/2018 8:32:13	
250PPBCN	A	Reactive CN	P	252.2963	µg/l	5/3/2018 8:32:14	
500PPBCN	A	Reactive CN	P	498.6399	µg/l	5/3/2018 8:32:15	
ICV1	S	Reactive CN	P	95.8689	µg/l	5/3/2018 11:52:17	
ICB1	S	Reactive CN	P	-1.1605	µg/l	5/3/2018 11:52:18	
CCV1	S	Reactive CN	P	242.3666	µg/l	5/3/2018 11:52:19	
CCB1	S	Reactive CN	P	-0.8139	µg/l	5/3/2018 11:52:20	
PB108858BL	S	Reactive CN	P	-0.5049	µg/l	5/3/2018 11:52:21	
J2669-02	S	Reactive CN	P	-0.119	µg/l	5/3/2018 11:52:22	
J2669-02DUP	S	Reactive CN	P	-1.3493	µg/l	5/3/2018 11:52:23	
J2669-04	S	Reactive CN	P	-0.3651	µg/l	5/3/2018 11:52:24	
J2670-02	S	Reactive CN	P	-0.7869	µg/l	5/3/2018 11:52:25	
J2679-02	S	Reactive CN	P	-0.6561	µg/l	5/3/2018 11:52:26	
J2679-04	S	Reactive CN	P	-0.7387	µg/l	5/3/2018 11:52:27	
J2679-06	S	Reactive CN	P	-1.026	µg/l	5/3/2018 11:59:50	
J2681-02	S	Reactive CN	P	-1.1923	µg/l	5/3/2018 11:59:51	
J2681-04	S	Reactive CN	P	-1.0439	µg/l	5/3/2018 11:59:52	
CCV2	S	Reactive CN	P	240.9906	µg/l	5/3/2018 11:59:53	
CCB2	S	Reactive CN	P	-0.99	µg/l	5/3/2018 11:59:54	
J2686-02	S	Reactive CN	P	-0.7292	µg/l	5/3/2018 11:59:55	
J2686-04	S	Reactive CN	P	-0.6616	µg/l	5/3/2018 11:59:56	
J2695-02	S	Reactive CN	P	-1.6003	µg/l	5/3/2018 11:59:57	
J2703-02	S	Reactive CN	P	-0.9333	µg/l	5/3/2018 11:59:58	
J2710-03	S	Reactive CN	P	-0.9578	µg/l	5/3/2018 11:59:59	
J2711-03	S	Reactive CN	P	-0.8135	µg/l	5/3/2018 12:00:00	
J2712-03	S	Reactive CN	P	-1.3256	µg/l	5/3/2018 12:05:05	
PB108859BL	S	Reactive CN	P	-1.0658	µg/l	5/3/2018 12:05:06	
J2703-04	S	Reactive CN	P	-0.0108	µg/l	5/3/2018 12:05:07	
J2703-04DUP	S	Reactive CN	P	0.8048	µg/l	5/3/2018 12:05:08	
CCV3	S	Reactive CN	P	251.2913	µg/l	5/3/2018 12:05:09	
CCB3	S	Reactive CN	P	-0.7734	µg/l	5/3/2018 12:05:10	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/3/2018 8:53


Test Total CN

Accepted 5/3/2018 8:53

Factor 1022
Bias 0.005

Coeff. of det. 0.999950

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.7143	0.0000	
2	5.0PPBCN	0.008	3.8393	5.0000	
3	10PPBCN	0.014	9.6234	10.0000	
4	50PPBCN	0.054	50.3197	50.0000	
5	100PPBCN	0.103	100.9958	100.0000	
6	250PPBCN	0.251	252.2963	250.0000	
7	500PPBCN	0.492	498.6399	500.0000	