

=====

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AS5/3/2018 15:09
-----Test: Total CN ug/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.039	0.0	0.099	
ICB1	-0.412	0.0	0.004	
CCV1	239.818	0.0	0.239	
CCB1	-0.247	0.0	0.004	
PB108883BL	-1.021	0.0	0.004	
PB108883BS	96.447	0.0	0.099	
J2669-01	0.090	0.0	0.005	
J2669-01DUP	0.081	0.0	0.005	
J2669-01MS	28.853	0.0	0.033	
J2669-01MSD	28.810	0.0	0.033	
J2669-03	-0.212	0.0	0.004	
J2670-01	-0.789	0.0	0.004	
J2679-01	0.687	0.0	0.005	
J2679-03	2.661	0.0	0.007	
CCV2	235.648	0.0	0.235	
CCB2	-0.376	0.0	0.004	
J2679-05	-0.243	0.0	0.004	
J2681-01	0.010	0.0	0.005	
J2681-03	-0.188	0.0	0.004	
J2696-01	-0.249	0.0	0.004	
J2709-01	-0.407	0.0	0.004	
CCV3	236.506	0.0	0.236	
CCB3	-0.938	0.0	0.004	

N	23
Mean	41.764
SD	82.3132
CV%	197.09

Aquakem v. 7.2AQ1

Results from time period:

Thu May 03 14:58:27 2018

Thu May 03 15:08:50 2018

AS

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.7143	µg/l	5/3/2018 8:32:09	
5.OPPBCN	A	Total CN	P	3.8393	µg/l	5/3/2018 8:32:10	
10PPBCN	A	Total CN	P	9.6234	µg/l	5/3/2018 8:32:11	
50PPBCN	A	Total CN	P	50.3197	µg/l	5/3/2018 8:32:12	
100PPBCN	A	Total CN	P	100.9958	µg/l	5/3/2018 8:32:13	
250PPBCN	A	Total CN	P	252.2963	µg/l	5/3/2018 8:32:14	
500PPBCN	A	Total CN	P	498.6399	µg/l	5/3/2018 8:32:15	
ICV1	S	Total CN	P	96.0393	µg/l	5/3/2018 14:58:27	
ICB1	S	Total CN	P	-0.4124	µg/l	5/3/2018 14:58:28	
CCV1	S	Total CN	P	239.8184	µg/l	5/3/2018 14:58:29	
CCB1	S	Total CN	P	-0.2467	µg/l	5/3/2018 14:58:30	
PB108883BL	S	Total CN	P	-1.021	µg/l	5/3/2018 14:58:31	
PB108883BS	S	Total CN	P	96.4474	µg/l	5/3/2018 14:58:32	
J2669-01	S	Total CN	P	0.0896	µg/l	5/3/2018 14:58:33	
J2669-01DUP	S	Total CN	P	0.0809	µg/l	5/3/2018 14:58:34	
J2669-01MS	S	Total CN	P	28.8526	µg/l	5/3/2018 14:58:35	
J2669-01MSD	S	Total CN	P	28.8101	µg/l	5/3/2018 14:58:36	
J2669-03	S	Total CN	P	-0.2118	µg/l	5/3/2018 14:58:37	
J2670-01	S	Total CN	P	-0.7886	µg/l	5/3/2018 15:06:02	
J2679-01	S	Total CN	P	0.6866	µg/l	5/3/2018 15:06:03	
J2679-03	S	Total CN	P	2.6611	µg/l	5/3/2018 15:06:04	
CCV2	S	Total CN	P	235.6478	µg/l	5/3/2018 15:06:05	
CCB2	S	Total CN	P	-0.3761	µg/l	5/3/2018 15:06:06	
J2679-05	S	Total CN	P	-0.2432	µg/l	5/3/2018 15:06:07	
J2681-01	S	Total CN	P	0.0096	µg/l	5/3/2018 15:06:08	
J2681-03	S	Total CN	P	-0.1879	µg/l	5/3/2018 15:06:09	
J2696-01	S	Total CN	P	-0.249	µg/l	5/3/2018 15:06:10	
J2709-01	S	Total CN	P	-0.4075	µg/l	5/3/2018 15:06:11	
CCV3	S	Total CN	P	236.5065	µg/l	5/3/2018 15:06:12	
CCB3	S	Total CN	P	-0.9382	µg/l	5/3/2018 15:08:50	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/3/2018 14:17


Test Total CN

Accepted 5/3/2018 8:53

Factor 1022
Bias 0.005

Coeff. of det. 0.999950

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.7143	0.0000	
2	5.0PPBCN	0.008	3.8393	5.0000	
3	10PPBCN	0.014	9.6234	10.0000	
4	50PPBCN	0.054	50.3197	50.0000	
5	100PPBCN	0.103	100.9958	100.0000	
6	250PPBCN	0.251	252.2963	250.0000	
7	500PPBCN	0.492	498.6399	500.0000	