

Daily Analysis Runlog For Sequence/QC Batch ID # LB95340

Review By

Review On

SubDirectory

LB95340

Test :

Cyanide

STD. NAME		STD REF.#				
ICAL Standard		WP64442,WP64443,WP64444,WP64445,WP64446,WP64447,WP64448				
ICV Standard		W1542				
CCV Standard		WP64449				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		WP62802				
Chk Standard		WP64452,WP59958,WP60878				

Sr#	SampleID	ClientID	QcType	Date	Comment	Status
1	0.0PPBCN	0.0PPBCN	CAL1	05/14/18 11:08		OK
2	5.0PPBCN	5.0PPBCN	CAL2	05/14/18 11:08		OK
3	10PPBCN	10PPBCN	CAL3	05/14/18 11:08		OK
4	50PPBCN	50PPBCN	CAL4	05/14/18 11:08		OK
5	100PPBCN	100PPBCN	CAL5	05/14/18 11:08		OK
6	250PPBCN	250PPBCN	CAL6	05/14/18 11:08		OK
7	500PPBCN	500PPBCN	CAL7	05/14/18 11:08		OK
8	ICV1	ICV1	ICV	05/14/18 14:58		OK
9	ICB1	ICB1	ICB	05/14/18 14:58		OK
10	CCV1	CCV1	CCV	05/14/18 14:58		OK
11	CCB1	CCB1	CCB	05/14/18 14:58		OK
12	PB109239BL	PB109239BL	MB	05/14/18 14:58		OK
13	PB109239BS	PB109239BS	LCS	05/14/18 14:58		OK
14	LOWPB109239	LOWPB109239	SAM	05/14/18 14:58		OK
15	HIGHPB109239	HIGHPB109239	SAM	05/14/18 14:58		OK
16	J2880-02	PLANTER-AREA-2-F	SAM	05/14/18 14:58		OK
17	J2880-02DUP	PLANTER-AREA-2-F	DUP	05/14/18 14:58		OK
18	J2880-02MS	PLANTER-AREA-2-F	MS	05/14/18 14:58		OK
19	J2880-02MSD	PLANTER-AREA-2-F	MSD	05/14/18 15:05		OK
20	J2880-05	PLANTER-AREA-3-F	SAM	05/14/18 15:05		OK
21	J2888-01	RRT-N1	SAM	05/14/18 15:05		OK

Daily Analysis Runlog For Sequence/QC Batch ID # LB95340

Review By			Review On			
SubDirectory		LB95340		Test : Cyanide		
STD. NAME		STD REF.#				
ICAL Standard		WP64442,WP64443,WP64444,WP64445,WP64446,WP64447,WP64448				
ICV Standard		W1542				
CCV Standard		WP64449				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		WP62802				
Chk Standard		WP64452,WP59958,WP60878				
22	CCV2	CCV2	CCV	05/14/18 15:05		OK
23	CCB2	CCB2	CCB	05/14/18 15:05		OK
24	J2888-05	RRT-N2	SAM	05/14/18 15:05		OK
25	J2888-09	RRT-S1	SAM	05/14/18 15:05		OK
26	J2888-13	RRT-S2	SAM	05/14/18 15:05		OK
27	J2888-17	26KV-N1	SAM	05/14/18 15:05		OK
28	J2888-21	26KV-N2	SAM	05/14/18 15:05		OK
29	J2888-25	26KV-S1	SAM	05/14/18 15:06		OK
30	J2888-29	26KV-S2	SAM	05/14/18 15:13		OK
31	PB109240BL	PB109240BL	MB	05/14/18 15:13		OK
32	J2877-01	MS-WATER	SAM	05/14/18 15:13		OK
33	CCV3	CCV3	CCV	05/14/18 15:13		OK
34	CCB3	CCB3	CCB	05/14/18 15:13		OK
35	J2877-01DUP	MS-WATERDUP	DUP	05/14/18 15:13		OK
36	J2877-01MS	MS-WATERMS	MS	05/14/18 15:13		OK
37	J2877-01MSD	MS-WATERMSD	MSD	05/14/18 15:13		OK
38	J2883-03	DRUM-1-5	SAM	05/14/18 15:13		OK
39	PB109240BS	PB109240BS	LCS	05/14/18 15:35		OK
40	CCV4	CCV4	CCV	05/14/18 15:35		OK
41	CCB4	CCB4	CCB	05/14/18 15:35		OK



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB95340	Date Received:	
Client Sample ID:	HIGHPB109239	SDG No.:	LB95340
Lab Sample ID:	HIGHPB109239	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.49		1	0.003		0.005	mg/L		05/14/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	5/11/2018 12:00:00 AM
Project:	LB95340	Date Received:	5/11/2018 12:00:00 AM
Client Sample ID:	MS-WATER	SDG No.:	LB95340
Lab Sample ID:	J2877-01	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.0033	J	1	0.003		0.005	mg/L	05/14/2018	05/14/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	5/10/2018 12:00:00 AM
Project:	LB95340	Date Received:	5/11/2018 12:00:00 AM
Client Sample ID:	PLANTER-AREA-2-PRECHAR	SDG No.:	LB95340
Lab Sample ID:	J2880-02	Matrix:	Solid
Level (low/med):	low	% Solid:	84.8

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.2	J	1	0.037		0.28	mg/Kg	05/14/2018	05/14/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	5/10/2018 12:00:00 AM
Project:	LB95340	Date Received:	5/11/2018 12:00:00 AM
Client Sample ID:	PLANTER-AREA-3-PRECHAR	SDG No.:	LB95340
Lab Sample ID:	J2880-05	Matrix:	Solid
Level (low/med):	low	% Solid:	89.7

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.21	J	1	0.035		0.27	mg/Kg	05/14/2018	05/14/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	5/11/2018 12:00:00 AM
Project:	LB95340	Date Received:	5/11/2018 12:00:00 AM
Client Sample ID:	DRUM-1-5	SDG No.:	LB95340
Lab Sample ID:	J2883-03	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.0041	J	1	0.003		0.005	mg/L	05/14/2018	05/14/2018	9012B

Report of Analysis

Client:		Date Collected:	5/11/2018 12:00:00 AM
Project:	LB95340	Date Received:	5/11/2018 12:00:00 AM
Client Sample ID:	RRT-N1	SDG No.:	LB95340
Lab Sample ID:	J2888-01	Matrix:	Solid
Level (low/med):	low	% Solid:	100

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.086	J	1	0.031		0.24	mg/Kg	05/14/2018	05/14/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	5/11/2018 12:00:00 AM
Project:	LB95340	Date Received:	5/11/2018 12:00:00 AM
Client Sample ID:	RRT-N2	SDG No.:	LB95340
Lab Sample ID:	J2888-05	Matrix:	Solid
Level (low/med):	low	% Solid:	100

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.2	J	1	0.031		0.24	mg/Kg	05/14/2018	05/14/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	5/11/2018 12:00:00 AM
Project:	LB95340	Date Received:	5/11/2018 12:00:00 AM
Client Sample ID:	RRT-S1	SDG No.:	LB95340
Lab Sample ID:	J2888-09	Matrix:	Solid
Level (low/med):	low	% Solid:	100

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.15	J	1	0.032		0.24	mg/Kg	05/14/2018	05/14/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	5/11/2018 12:00:00 AM
Project:	LB95340	Date Received:	5/11/2018 12:00:00 AM
Client Sample ID:	RRT-S2	SDG No.:	LB95340
Lab Sample ID:	J2888-13	Matrix:	Solid
Level (low/med):	low	% Solid:	100

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.081	J	1	0.032		0.24	mg/Kg	05/14/2018	05/14/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	5/11/2018 12:00:00 AM
Project:	LB95340	Date Received:	5/11/2018 12:00:00 AM
Client Sample ID:	26KV-N1	SDG No.:	LB95340
Lab Sample ID:	J2888-17	Matrix:	Solid
Level (low/med):	low	% Solid:	100

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.2	J	1	0.032		0.25	mg/Kg	05/14/2018	05/14/2018	9012B

Report of Analysis

Client:		Date Collected:	5/11/2018 12:00:00 AM
Project:	LB95340	Date Received:	5/11/2018 12:00:00 AM
Client Sample ID:	26KV-N2	SDG No.:	LB95340
Lab Sample ID:	J2888-21	Matrix:	Solid
Level (low/med):	low	% Solid:	100

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.46		1	0.033		0.25	mg/Kg	05/14/2018	05/14/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	5/11/2018 12:00:00 AM
Project:	LB95340	Date Received:	5/11/2018 12:00:00 AM
Client Sample ID:	26KV-S1	SDG No.:	LB95340
Lab Sample ID:	J2888-25	Matrix:	Solid
Level (low/med):	low	% Solid:	100

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.14	J	1	0.031		0.24	mg/Kg	05/14/2018	05/14/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	5/11/2018 12:00:00 AM
Project:	LB95340	Date Received:	5/11/2018 12:00:00 AM
Client Sample ID:	26KV-S2	SDG No.:	LB95340
Lab Sample ID:	J2888-29	Matrix:	Solid
Level (low/med):	low	% Solid:	100

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.17	J	1	0.031		0.23	mg/Kg	05/14/2018	05/14/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB95340	Date Received:	
Client Sample ID:	LOWPB109239	SDG No.:	LB95340
Lab Sample ID:	LOWPB109239	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.011		1	0.003		0.005	mg/L		05/14/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB95340	Date Received:	
Client Sample ID:	LOWPB109239	SDG No.:	LB95340
Lab Sample ID:	LOWPB109239	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
-----	-----------	-------	------	----	-----	-----	------------	-------	-----------	-----------	----------

Color Before:

Clarity Before:

Texture:

Color After:

Clarity After:

Artifacts:

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB95340

SDG No.: LB95340

Contract: _____ Lab Code: CHEM

Case No.: LB95340

SAS No.: LB95340

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV1	Cyanide	0.095	0.099	96		85 - 115		05/14/2018	14:58	lb95340



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB95340

SDG No.: LB95340

Contract: _____ Lab Code: CHEM

Case No.: LB95340

SAS No.: LB95340

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	Cyanide	0.24	0.25	96		90 - 110		05/14/2018	14:58	lb95340
CCV2	Cyanide	0.24	0.25	96		90 - 110		05/14/2018	15:05	lb95340
CCV3	Cyanide	0.24	0.25	96		90 - 110		05/14/2018	15:13	lb95340
CCV4	Cyanide	0.26	0.25	104		90 - 110		05/14/2018	15:35	lb95340



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB95340

SDG No.: LB95340

Contract:

Lab Code: CHEM

Case No.: LB95340

SAS No.: LB95340

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	Cyanide	0.00061	+/-0.005	U		0.005		05/14/2018	14:58	lb95340



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB95340

SDG No.: LB95340

Contract:

Lab Code: CHEM

Case No.: LB95340

SAS No.: LB95340

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	Cyanide	0.001	+/-0.005	U		0.005		05/14/2018	14:58	1b95340
CCB2	Cyanide	0.00067	+/-0.005	U		0.005		05/14/2018	15:05	1b95340
CCB3	Cyanide	0.00077	+/-0.005	U		0.005		05/14/2018	15:13	1b95340
CCB4	Cyanide	0.0013	+/-0.005	U		0.005		05/14/2018	15:35	1b95340



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB95340

SDG No.: LB95340

Contract: Lab Code: CHEM

Case No.: LB95340

SAS No.: LB95340

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



GENCHEM
- 3b -
PREPARATION BLANK SUMMARY

Client: LB95340

SDG No.: LB95340

Instrument: Konelab 20

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	LOD mg/Kg	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB109239BL	Cyanide	SOLID		Batch Number:		PB109239		Prep Date:	05/14/2018	
		0.037	<0.25	J		0.25		05/14/2018	14:58	lb95340
Sample ID	Analyte	Result (mg/L)	Acceptance Limit	Conc Qual	LOD mg/L	CRQL mg/L	M	Analysis Date	Analysis Time	Run
PB109240BL	Cyanide	WATER		Batch Number:		PB109240		Prep Date:	05/14/2018	
		0.0011	<0.005	U		0.005		05/14/2018	15:13	lb95340



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB95340 level: low sdg no.: LB95340
contract: lab code: CHEM case no.: LB95340 sas no.: LB95340
matrix: WATER sample id: J2877-01 client id: MS-WATERMS
Percent Solids for Sample: 0 Spiked ID: J2877-01MS Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/L	75 - 125	0.039		0.0033	J	0.04	89		



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB95340 level: low sdg no.: LB95340
contract: lab code: CHEM case no.: LB95340 sas no.: LB95340
matrix: WATER sample id: J2877-01 client id: MS-WATERMSD
Percent Solids for Sample: 0 Spiked ID: J2877-01MSD Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/L	75 - 125	0.04		0.0033	J	0.04	92		



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB95340 level: low sdg no.: LB95340
contract: lab code: CHEM case no.: LB95340 sas no.: LB95340
matrix: Solid sample id: J2880-02 client id: PLANTER-AREA-2-PRECHARMS
Percent Solids for Sample: 84.8 Spiked ID: J2880-02MS Percent Solids for Spike Sample: 84.8

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	1.1		0.2	J	2.3	39	*	



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB95340 level: low sdg no.: LB95340
contract: lab code: CHEM case no.: LB95340 sas no.: LB95340
matrix: Solid sample id: J2880-02 client id: PLANTER-AREA-2-PRECHARMSD
Percent Solids for Sample: 84.8 Spiked ID: J2880-02MSD Percent Solids for Spike Sample: 84.8

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	1.1		0.2	J	2.3	39	*	

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB95340 **Level:** LOW **SDG No.:** LB95340
Contract: **Lab Code:** CHEM **Case No.:** LB95340 **SAS No.:** LB95340
Matrix: WATER **Sample ID:** J2877-01 **Client ID:** MS-WATERDUP
Percent Solids for Sample: 0 **Duplicate ID** J2877-01DUP **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/L	20	0.0033	J	0.0028	U	200.0	*	

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB95340 **Level:** LOW **SDG No.:** LB95340
Contract: **Lab Code:** CHEM **Case No.:** LB95340 **SAS No.:** LB95340
Matrix: WATER **Sample ID:** J2877-01MS **Client ID:** MS-WATERMSD
Percent Solids for Sample: 0 **Duplicate ID** J2877-01MSD **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/L	20	0.039		0.04		3		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB95340 **Level:** LOW **SDG No.:** LB95340
Contract: **Lab Code:** CHEM **Case No.:** LB95340 **SAS No.:** LB95340
Matrix: Solid **Sample ID:** J2880-02 **Client ID:** PLANTER-AREA-2-PRECHARDUP
Percent Solids for Sample: 84.8 **Duplicate ID** J2880-02DUP **Percent Solids for Spike Sample:** 84.8

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	0.2	J	0.15	J	29	*	

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB95340 **Level:** LOW **SDG No.:** LB95340
Contract: **Lab Code:** CHEM **Case No.:** LB95340 **SAS No.:** LB95340
Matrix: Solid **Sample ID:** J2880-02MS **Client ID:** PLANTER-AREA-2-PRECHARMSD
Percent Solids for Sample: 84.8 **Duplicate ID** J2880-02MSD **Percent Solids for Spike Sample:** 84.8

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	1.1		1.1		0		



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB95340

SDG No.: LB95340

Contract:

Lab Code: CHEM

Case No.: LB95340

SAS No.: LB95340

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB109239BS Cyanide	mg/Kg	5	4.8		96	85 - 115	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB95340

SDG No.: LB95340

Contract:

Lab Code: CHEM

Case No.: LB95340

SAS No.: LB95340

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB109240BS Cyanide	mg/L	0.1	0.1		100	85 - 115	