

21395361

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/15/2018 11:11

Test: *Reaction*
Total CN *ug/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.851	0.0	0.098	
ICB1	-0.091	0.0	0.004	
CCV1	243.507	0.0	0.244	
CCB1	-0.196	0.0	0.004	
PB109264BL	-0.166	0.0	0.004	
J2899-03	-0.312	0.0	0.004	
J2899-03DUP	0.051	0.0	0.004	
J2899-07	0.059	0.0	0.004	
J2899-10	0.304	0.0	0.004	
J2901-09	0.343	0.0	0.004	
J2901-10	-0.309	0.0	0.004	
J2901-11	-0.265	0.0	0.004	
J2901-12	0.088	0.0	0.004	
J2902-05	-0.438	0.0	0.004	
CCV2	240.639	0.0	0.241	
CCB2	-0.539	0.0	0.004	
J2902-06	-0.293	0.0	0.004	
PB109263BL	-0.601	0.0	0.004	
J2877-02	-0.372	0.0	0.004	
J2877-02DUP	-0.164	0.0	0.004	
J2883-04	-0.255	0.0	0.004	
PB109285BL	-0.622	0.0	0.004	
J1759-02	-0.893	0.0	0.003	
J1759-02DUP	-0.687	0.0	0.003	
J2267-02	-0.881	0.0	0.003	
CCV3	250.680	0.0	0.251	
CCB3	-0.241	0.0	0.004	
N	27			
Mean	30.526			
SD	79.4392			
CV%	260.24			

Aquakem v. 7.2AQ1

Results from time period:

Tue May 15 10:38:19 2018

Tue May 15 11:08:09 2018

AS-

Sample Id	Sam/Ctr/c/	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-0.9093	µg/l	5/15/2018 8:11:33	
5.0PPBCN	A	Reactive CN	P	3.8993	µg/l	5/15/2018 8:11:34	
10PPBCN	A	Reactive CN	P	9.1663	µg/l	5/15/2018 8:11:35	
50PPBCN	A	Reactive CN	P	50.0986	µg/l	5/15/2018 8:11:36	
100PPBCN	A	Reactive CN	P	101.5247	µg/l	5/15/2018 8:11:37	
250PPBCN	A	Reactive CN	P	253.015	µg/l	5/15/2018 8:11:38	
500PPBCN	A	Reactive CN	P	498.2053	µg/l	5/15/2018 8:11:39	
ICV1	S	Reactive CN	P	95.8514	µg/l	5/15/2018 10:38:19	
ICB1	S	Reactive CN	P	-0.0911	µg/l	5/15/2018 10:38:20	
CCV1	S	Reactive CN	P	243.5068	µg/l	5/15/2018 10:38:21	
CCB1	S	Reactive CN	P	-0.196	µg/l	5/15/2018 10:38:22	
PB109264BL	S	Reactive CN	P	-0.1664	µg/l	5/15/2018 10:38:23	
J2899-03	S	Reactive CN	P	-0.3123	µg/l	5/15/2018 10:38:24	
J2899-03DUP	S	Reactive CN	P	0.0512	µg/l	5/15/2018 10:38:25	
J2899-07	S	Reactive CN	P	0.0589	µg/l	5/15/2018 10:38:26	
J2899-10	S	Reactive CN	P	0.3042	µg/l	5/15/2018 10:38:27	
J2901-09	S	Reactive CN	P	0.3426	µg/l	5/15/2018 10:38:28	
J2901-10	S	Reactive CN	P	-0.3092	µg/l	5/15/2018 10:38:29	
J2901-11	S	Reactive CN	P	-0.2654	µg/l	5/15/2018 10:45:54	
J2901-12	S	Reactive CN	P	0.0881	µg/l	5/15/2018 10:45:55	
J2902-05	S	Reactive CN	P	-0.4378	µg/l	5/15/2018 10:45:56	
CCV2	S	Reactive CN	P	240.6386	µg/l	5/15/2018 10:45:57	
CCB2	S	Reactive CN	P	-0.539	µg/l	5/15/2018 10:45:58	
J2902-06	S	Reactive CN	P	-0.2929	µg/l	5/15/2018 10:45:59	
PB109263BL	S	Reactive CN	P	-0.6014	µg/l	5/15/2018 10:46:00	
J2877-02	S	Reactive CN	P	-0.3721	µg/l	5/15/2018 10:46:01	
J2877-02DUP	S	Reactive CN	P	-0.1639	µg/l	5/15/2018 10:46:02	
J2883-04	S	Reactive CN	P	-0.2553	µg/l	5/15/2018 10:46:03	
PB109285BL	S	Reactive CN	P	-0.6216	µg/l	5/15/2018 11:08:04	
J1759-02	S	Reactive CN	P	-0.8932	µg/l	5/15/2018 11:08:05	
J1759-02DUP	S	Reactive CN	P	-0.6874	µg/l	5/15/2018 11:08:06	
J2267-02	S	Reactive CN	P	-0.8811	µg/l	5/15/2018 11:08:07	
CCV3	S	Reactive CN	P	250.6798	µg/l	5/15/2018 11:08:08	
CCB3	S	Reactive CN	P	-0.2414	µg/l	5/15/2018 11:08:09	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS.

5/15/2018 8:12


Test Total CN

Accepted 5/15/2018 8:11

Factor 1016
 Bias 0.004

Coeff. of det. 0.999915

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.9093	0.0000	
2	5.0PPBCN	0.008	3.8993	5.0000	
3	10PPBCN	0.013	9.1663	10.0000	
4	50PPBCN	0.053	50.0986	50.0000	
5	100PPBCN	0.104	101.5247	100.0000	
6	250PPBCN	0.253	253.0150	250.0000	
7	500PPBCN	0.494	498.2053	500.0000	