

21395375

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/15/2018 16:08

Test: Total CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.816	0.0	0.097	
ICB1	0.067	0.0	0.004	
CCV1	238.303	0.0	0.239	
CCB1	-0.171	0.0	0.004	
PB109286BL	-0.645	0.0	0.003	
PB109286BS	95.831	0.0	0.098	
J1759-01	4.354	0.0	0.008	
J1759-01DUP	4.412	0.0	0.008	
J1759-01MS	38.221	0.0	0.042	
J1759-01MSD	38.408	0.0	0.042	
J2267-01	3.108	0.0	0.007	
J2910-01	0.498	0.0	0.005	
J2910-03	0.290	0.0	0.004	
CCV2	254.502	0.0	0.255	
CCB2	-0.248	0.0	0.004	

N 15
Mean 51.450
SD 85.7788
CV% 166.72

Aquakem v. 7.2AQ1

Results from time period:

Tue May 15 15:51:34 2018

Tue May 15 15:55:56 2018

AS

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.9093	µg/l	5/15/2018 8:11:33	
5.0PPBCN	A	Total CN	P	3.8993	µg/l	5/15/2018 8:11:34	
10PPBCN	A	Total CN	P	9.1663	µg/l	5/15/2018 8:11:35	
50PPBCN	A	Total CN	P	50.0986	µg/l	5/15/2018 8:11:36	
100PPBCN	A	Total CN	P	101.5247	µg/l	5/15/2018 8:11:37	
250PPBCN	A	Total CN	P	253.015	µg/l	5/15/2018 8:11:38	
500PPBCN	A	Total CN	P	498.2053	µg/l	5/15/2018 8:11:39	
ICV1	S	Total CN	P	94.8156	µg/l	5/15/2018 15:51:34	
ICB1	S	Total CN	P	0.0674	µg/l	5/15/2018 15:51:35	
CCV1	S	Total CN	P	238.303	µg/l	5/15/2018 15:51:36	
CCB1	S	Total CN	P	-0.1708	µg/l	5/15/2018 15:51:37	
PB109286BL	S	Total CN	P	-0.6454	µg/l	5/15/2018 15:51:38	
PB109286BS	S	Total CN	P	95.8305	µg/l	5/15/2018 15:51:39	
J1759-01	S	Total CN	P	4.3538	µg/l	5/15/2018 15:51:40	
J1759-01DUP	S	Total CN	P	4.4122	µg/l	5/15/2018 15:51:41	
J1759-01MS	S	Total CN	P	38.2205	µg/l	5/15/2018 15:51:42	
J1759-01MSD	S	Total CN	P	38.4083	µg/l	5/15/2018 15:51:43	
J2267-01	S	Total CN	P	3.1076	µg/l	5/15/2018 15:51:44	
J2910-01	S	Total CN	P	0.4981	µg/l	5/15/2018 15:55:53	
J2910-03	S	Total CN	P	0.2896	µg/l	5/15/2018 15:55:54	
CCV2	S	Total CN	P	254.5021	µg/l	5/15/2018 15:55:55	
CCB2	S	Total CN	P	-0.2484	µg/l	5/15/2018 15:55:56	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/15/2018 16:10


Test Total CN

Accepted 5/15/2018 8:11

Factor 1016
Bias 0.004

Coeff. of det. 0.999915

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.9093	0.0000	
2	5.0PPBCN	0.008	3.8993	5.0000	
3	10PPBCN	0.013	9.1663	10.0000	
4	50PPBCN	0.053	50.0986	50.0000	
5	100PPBCN	0.104	101.5247	100.0000	
6	250PPBCN	0.253	253.0150	250.0000	
7	500PPBCN	0.494	498.2053	500.0000	