

LB95434

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/17/2018 14:51

Reactive
Test: Total CN Mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.272	0.0	0.097	
ICB1	0.122	0.0	0.004	
CCV1	239.121	0.0	0.239	
CCB1	-0.066	0.0	0.004	
PB109391BL	-0.193	0.0	0.004	
J2966-04	-0.330	0.0	0.003	
J2966-04DUP	-0.252	0.0	0.004	
J2966-08	-0.227	0.0	0.004	
J2967-06	-0.258	0.0	0.004	
PB109392BL	-0.547	0.0	0.003	
J2966-10	-1.001	0.0	0.003	
J2966-10DUP	-0.537	0.0	0.003	
CCV2	256.245	0.0	0.256	
CCB2	-0.674	0.0	0.003	

N	14
Mean	41.905
SD	90.8653
CV%	216.83

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/17/2018 9:25


Test Total CN

Accepted 5/17/2018 9:25

Factor 1017
Bias 0.004

Coeff. of det. 0.999906

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.7108	0.0000	
2	5.0PPBCN	0.008	4.5501	5.0000	
3	10PPBCN	0.013	9.7387	10.0000	
4	50PPBCN	0.052	48.8552	50.0000	
5	100PPBCN	0.103	100.7825	100.0000	
6	250PPBCN	0.253	253.6330	250.0000	
7	500PPBCN	0.494	498.1512	500.0000	

Aquakem v. 7.2AQ1

AS.

Results from time period:

Thu May 17 14:38:18 2018

Thu May 17 14:42:11 2018

Sample Id	Sam/Ctr/c	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-0.7108	µg/l	5/17/2018 9:18:31	
5.0PPBCN	A	Reactive CN	P	4.5501	µg/l	5/17/2018 9:18:32	
10PPBCN	A	Reactive CN	P	9.7387	µg/l	5/17/2018 9:18:33	
50PPBCN	A	Reactive CN	P	48.8552	µg/l	5/17/2018 9:18:34	
100PPBCN	A	Reactive CN	P	100.7825	µg/l	5/17/2018 9:18:35	
250PPBCN	A	Reactive CN	P	253.633	µg/l	5/17/2018 9:18:36	
500PPBCN	A	Reactive CN	P	498.1512	µg/l	5/17/2018 9:18:37	
ICV1	S	Reactive CN	P	95.272	µg/l	5/17/2018 14:38:18	
ICB1	S	Reactive CN	P	0.1215	µg/l	5/17/2018 14:38:19	
CCV1	S	Reactive CN	P	239.1206	µg/l	5/17/2018 14:38:20	
CCB1	S	Reactive CN	P	-0.0658	µg/l	5/17/2018 14:38:21	
PB109391BL	S	Reactive CN	P	-0.1932	µg/l	5/17/2018 14:38:22	
J2966-04	S	Reactive CN	P	-0.3296	µg/l	5/17/2018 14:38:23	
J2966-04DUP	S	Reactive CN	P	-0.2516	µg/l	5/17/2018 14:38:24	
J2966-08	S	Reactive CN	P	-0.2267	µg/l	5/17/2018 14:38:25	
J2967-06	S	Reactive CN	P	-0.2584	µg/l	5/17/2018 14:38:26	
PB109392BL	S	Reactive CN	P	-0.547	µg/l	5/17/2018 14:38:27	
J2966-10	S	Reactive CN	P	-1.0007	µg/l	5/17/2018 14:38:28	
J2966-10DUP	S	Reactive CN	P	-0.537	µg/l	5/17/2018 14:42:09	
CCV2	S	Reactive CN	P	256.2452	µg/l	5/17/2018 14:42:10	
CCB2	S	Reactive CN	P	-0.674	µg/l	5/17/2018 14:42:11	