

Daily Analysis Runlog For Sequence/QC Batch ID # LB95499

Review By		Review On				
SubDirectory		LB95499		Test : TOC		
STD. NAME		STD REF.#				
ICAL Standard		wp64486,wp64487,wp64488,wp64489,wp64490,wp64491,wp64492,wp64493				
ICV Standard		WP64495				
CCV Standard		wp64492				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		WP64495				
Chk Standard		wp62726				
Sr#	SampleID	ClientID	QcType	Date	Comment	Status
1	0.0PPM	0.0PPM	CAL1	05/15/18 11:50		OK
2	0.5PPM	0.5PPM	CAL2	05/15/18 12:03		OK
3	1.0PPM	1.0PPM	CAL3	05/15/18 12:16		OK
4	2.0PPM	2.0PPM	CAL4	05/15/18 12:29		OK
5	5.0PPM	5.0PPM	CAL5	05/15/18 12:43		OK
6	10.0PPM	10.0PPM	CAL6	05/15/18 12:57		OK
7	20.0PPM	20.0PPM	CAL7	05/15/18 13:12		OK
8	ICV1	ICV1	ICV	05/15/18 13:28		OK
9	ICB1	ICB1	ICB	05/15/18 13:41		OK
10	CCV1	CCV1	CCV	05/21/18 11:51		OK
11	CCB1	CCB1	CCB	05/21/18 12:05		OK
12	LB95499BLW	LB95499BLW	MB	05/21/18 12:28		OK
13	LB95499BSW	LB95499BSW	LCS	05/21/18 12:58		OK
14	J2923-05	LOQ-NPW-Precision	LOQ	05/21/18 13:25		OK
15	J2923-06	LOQ-NPW-Precision	LOQ	05/21/18 13:52		OK
16	J2923-07	LOQ-NPW-Precision	LOQ	05/21/18 14:44		OK
17	J2923-08	LOQ-NPW-Precision	LOQ	05/21/18 15:10		OK
18	CCV2	CCV2	CCV	05/21/18 15:25		OK
19	CCB2	CCB2	CCB	05/21/18 15:38		OK



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB95499

SDG No.: LB95499

Contract: _____

Lab Code: CHEM

Case No.: LB95499

SAS No.: LB95499

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV1	TOC	10.2	10	102		90 - 110		05/15/2018	13:28	lb95499



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB95499

SDG No.: LB95499

Contract: _____ Lab Code: CHEM

Case No.: LB95499

SAS No.: LB95499

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	TOC	10.1	10	101		90 - 110		05/21/2018	11:51	lb95499
CCV2	TOC	9.5	10	95		90 - 110		05/21/2018	15:25	lb95499

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	TOC	0.21	+/-0.5	J			0.5	05/15/2018	13:41	lb95499



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB95499

SDG No.: LB95499

Contract:

Lab Code: CHEM

Case No.: LB95499

SAS No.: LB95499

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	TOC	0.38	+/-0.5	J		0.5		05/21/2018	12:05	1b95499
CCB2	TOC	0.41	+/-0.5	J		0.5		05/21/2018	15:38	1b95499



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB95499

SDG No.: LB95499

Contract: Lab Code: CHEM

Case No.: LB95499

SAS No.: LB95499

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



GENCHEM
- 3b -
PREPARATION BLANK SUMMARY

Client: LB95499

SDG No.: LB95499

Instrument: Appolo-9000

Sample ID	Analyte	Result (mg/L)	Acceptance Limit	Conc Qual	LOD mg/L	CRQL mg/L	M	Analysis Date	Analysis Time	Run
LB95499BLW	TOC	WATER 0.083	<0.5	Batch Number: U		0.5		Prep Date: 05/21/2018	12:28	lb95499



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB95499

SDG No.: LB95499

Contract:

Lab Code: CHEM

Case No.: LB95499

SAS No.: LB95499

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
J2923-05 TOC	mg/L	1	0.84		84	50 - 150	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB95499

SDG No.: LB95499

Contract:

Lab Code: CHEM

Case No.: LB95499

SAS No.: LB95499

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
J2923-06 TOC	mg/L	1	0.72		72	50 - 150	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB95499

SDG No.: LB95499

Contract:

Lab Code: CHEM

Case No.: LB95499

SAS No.: LB95499

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
J2923-07 TOC	mg/L	1	0.71		71	50 - 150	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB95499

SDG No.: LB95499

Contract:

Lab Code: CHEM

Case No.: LB95499

SAS No.: LB95499

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
J2923-08 TOC	mg/L	1	0.63		63	50 - 150	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB95499

SDG No.: LB95499

Contract:

Lab Code: CHEM

Case No.: LB95499

SAS No.: LB95499

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
LB95499BSW TOC	mg/L	10	10.4		104	90 - 110	