

Daily Analysis Runlog For Sequence/QC Batch ID # LB95529

Review By		Review On				
SubDirectory		LB95529		Test : TOC		
STD. NAME		STD REF.#				
ICAL Standard		WP64254,WP64233,WP64234,WP64236,WP64238				
ICV Standard		WP64240				
CCV Standard		WP64233				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		WP64240				
Chk Standard		wp61490				
Sr#	SampleID	ClientID	QcType	Date	Comment	Status
1	BLANK	BLANK	CAL1	05/04/18 09:51		OK
2	250mg/l	250mg/l	CAL2	05/04/18 09:57		OK
3	500mg/l	500mg/l	CAL3	05/04/18 10:02		OK
4	1000mg/l	1000mg/l	CAL4	05/04/18 10:13		OK
5	2000mg/l	2000mg/l	CAL5	05/04/18 10:19		OK
6	ICV1	ICV1	ICV	05/04/18 10:27		OK
7	ICB1	ICB1	ICB	05/04/18 10:32		OK
8	CCV1	CCV1	CCV	05/23/18 10:53		OK
9	CCB1	CCB1	CCB	05/23/18 10:56		OK
10	LB95529BLS	LB95529BLS	MB	05/23/18 11:06		OK
11	LB95529BSS	LB95529BSS	LCS	05/23/18 11:20		OK
12	J2923-01	LOQ-SCM-Precision-	LOQ	05/23/18 11:30		OK
13	J2923-02	LOQ-SCM-Precision-	LOQ	05/23/18 12:07		OK
14	J2923-03	LOQ-SCM-Precision-	LOQ	05/23/18 12:25		OK
15	J2923-04	LOQ-SCM-Precision-	LOQ	05/23/18 12:32		OK
16	CCV2	CCV2	CCV	05/23/18 12:37		OK
17	CCB2	CCB2	CCB	05/23/18 12:40		OK



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB95529

SDG No.: LB95529

Contract: _____ Lab Code: CHEM

Case No.: LB95529

SAS No.: LB95529

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	TOC	973	1000	97.3		90 - 110		05/23/2018	10:53	LB95529
CCV2	TOC	992	1000	99.2		90 - 110		05/23/2018	12:37	LB95529



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB95529 **SDG No.:** LB95529
Contract: **Lab Code:** CHEM **Case No.:** LB95529 **SAS No.:** LB95529

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	TOC	13.3	+/-250	J		250		05/04/2018	10:32	LB95529



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB95529

SDG No.: LB95529

Contract:

Lab Code: CHEM

Case No.: LB95529

SAS No.: LB95529

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	TOC	15.1	+/-250	J		250		05/23/2018	10:56	LB95529
CCB2	TOC	7	+/-250	U		250		05/23/2018	12:40	LB95529



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB95529


SDG No.: LB95529

Contract: Lab Code: CHEM

Case No.: LB95529

SAS No.: LB95529

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



SDG No.: LB95529

Instrument: Appolo-9000

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	LOD mg/Kg	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
LB95529BLS		SOLID		Batch Number:				Prep Date:		
	TOC	7.7	<250	U		250		05/23/2018	11:06	LB95529



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB95529

SDG No.: LB95529

Contract:

Lab Code: CHEM

Case No.: LB95529

SAS No.: LB95529

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
J2923-01 TOC	mg/Kg	500	267		53.4	50 - 150	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB95529

SDG No.: LB95529

Contract:

Lab Code: CHEM

Case No.: LB95529

SAS No.: LB95529

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
J2923-02 TOC	mg/Kg	500	267		53.4	50 - 150	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB95529

SDG No.: LB95529

Contract:

Lab Code: CHEM

Case No.: LB95529

SAS No.: LB95529

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
J2923-03 TOC	mg/Kg	500	249	J	49.8	50 - 150	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB95529

SDG No.: LB95529

Contract:

Lab Code: CHEM

Case No.: LB95529

SAS No.: LB95529

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
J2923-04 TOC	mg/Kg	500	247	J	49.4	50 - 150	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB95529

SDG No.: LB95529

Contract:

Lab Code: CHEM

Case No.: LB95529

SAS No.: LB95529

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
LB95529BSS TOC	mg/Kg	1000	1070		107	90 - 110	