

2095590

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/24/2018 12:02

Test: *Reactive*
~~Total~~ CN *Myd*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.665	0.0	0.100	
ICB1	0.684	0.0	0.004	
CCV1	241.575	0.0	0.247	
CCB1	0.477	0.0	0.004	
PB109628BL	0.293	0.0	0.003	
J2670-06	0.563	0.0	0.004	
J2670-06DUP	0.509	0.0	0.004	
J2672-08	0.290	0.0	0.003	
J2678-04	1.185	0.0	0.004	
J3084-04	0.535	0.0	0.004	
J3084-08	1.660	0.0	0.005	
J3084-12	0.618	0.0	0.004	
J3086-02	0.299	0.0	0.003	
J3094-03	0.276	0.0	0.003	
CCV2	239.701	0.0	0.245	
CCB2	0.686	0.0	0.004	
J3095-02	0.009	0.0	0.003	
J3099-03	0.422	0.0	0.004	
J3099-06	0.487	0.0	0.004	
J3104-02	0.408	0.0	0.004	
J3104-04	0.228	0.0	0.003	
PB109629BL	1.051	0.0	0.004	
J3094-05	0.062	0.0	0.003	
J3094-05DUP	0.459	0.0	0.004	
CCV3	252.920	0.0	0.259	
CCB3	0.134	0.0	0.003	

N	26
Mean	32.354
SD	80.4310
CV%	248.60

Aquakem v. 7.2AQ1

Results from time period:

Thu May 24 11:50:19 2018

Thu May 24 12:02:14 2018

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	0.1339	µg/l	5/24/2018 8:05:07	
5.OPPBCN	A	Reactive CN	P	4.8734	µg/l	5/24/2018 8:05:08	
10PPBCN	A	Reactive CN	P	10.5266	µg/l	5/24/2018 8:05:09	
50PPBCN	A	Reactive CN	P	49.5826	µg/l	5/24/2018 8:05:10	
100PPBCN	A	Reactive CN	P	101.3357	µg/l	5/24/2018 8:05:11	
250PPBCN	A	Reactive CN	P	247.5649	µg/l	5/24/2018 8:05:12	
500PPBCN	A	Reactive CN	P	500.9828	µg/l	5/24/2018 8:05:13	
ICV1	S	Reactive CN	P	95.6651	µg/l	5/24/2018 11:50:19	
ICB1	S	Reactive CN	P	0.6841	µg/l	5/24/2018 11:50:20	
CCV1	S	Reactive CN	P	241.5747	µg/l	5/24/2018 11:50:21	
CCB1	S	Reactive CN	P	0.477	µg/l	5/24/2018 11:50:22	
PB109628BL	S	Reactive CN	P	0.2928	µg/l	5/24/2018 11:50:23	
J2670-06	S	Reactive CN	P	0.5634	µg/l	5/24/2018 11:50:24	
J2670-06DUP	S	Reactive CN	P	0.5086	µg/l	5/24/2018 11:50:25	
J2672-08	S	Reactive CN	P	0.2899	µg/l	5/24/2018 11:50:26	
J2678-04	S	Reactive CN	P	1.1851	µg/l	5/24/2018 11:50:27	
J3084-04	S	Reactive CN	P	0.5349	µg/l	5/24/2018 11:50:28	
J3084-08	S	Reactive CN	P	1.6603	µg/l	5/24/2018 11:50:29	
J3084-12	S	Reactive CN	P	0.6175	µg/l	5/24/2018 11:57:52	
J3086-02	S	Reactive CN	P	0.2992	µg/l	5/24/2018 11:57:53	
J3094-03	S	Reactive CN	P	0.2756	µg/l	5/24/2018 11:57:54	
CCV2	S	Reactive CN	P	239.7014	µg/l	5/24/2018 11:57:55	
CCB2	S	Reactive CN	P	0.6862	µg/l	5/24/2018 11:57:56	
J3095-02	S	Reactive CN	P	0.0088	µg/l	5/24/2018 11:57:57	
J3099-03	S	Reactive CN	P	0.4217	µg/l	5/24/2018 11:57:58	
J3099-06	S	Reactive CN	P	0.4872	µg/l	5/24/2018 11:57:59	
J3104-02	S	Reactive CN	P	0.4075	µg/l	5/24/2018 11:58:00	
J3104-04	S	Reactive CN	P	0.2284	µg/l	5/24/2018 11:58:01	
PB109629BL	S	Reactive CN	P	1.0515	µg/l	5/24/2018 11:58:02	
J3094-05	S	Reactive CN	P	0.062	µg/l	5/24/2018 12:02:11	
J3094-05DUP	S	Reactive CN	P	0.4593	µg/l	5/24/2018 12:02:12	
CCV3	S	Reactive CN	P	252.9204	µg/l	5/24/2018 12:02:13	
CCB3	S	Reactive CN	P	0.1338	µg/l	5/24/2018 12:02:14	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/24/2018 8:09


Test Total CN

Accepted 5/24/2018 8:09

Factor 989.2
Bias 0.003

Coeff. of det. 0.999955

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.1339	0.0000	
2	5.0PPBCN	0.008	4.8734	5.0000	
3	10PPBCN	0.014	10.5266	10.0000	
4	50PPBCN	0.053	49.5826	50.0000	
5	100PPBCN	0.106	101.3357	100.0000	
6	250PPBCN	0.253	247.5649	250.0000	
7	500PPBCN	0.510	500.9828	500.0000	