

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AS

5/24/2018 15:28

Test: Total CN ug/l

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.315	0.0	0.099	
ICB1	0.929	0.0	0.004	
CCV1	251.641	0.0	0.257	
CCB1	0.753	0.0	0.004	
PB109648BL	0.772	0.0	0.004	
PB109648BS	94.817	0.0	0.099	
MIDPB109648	237.098	0.0	0.243	
J3093-12	17.013	0.0	0.020	
J3093-12DUP	17.161	0.0	0.020	
J3093-12MS	52.476	0.0	0.056	
J3093-12MSD	52.005	0.0	0.056	
CCV2	253.645	0.0	0.260	
CCB2	0.521	0.0	0.004	

N	13
Mean	82.627
SD	99.6833
CV%	120.64

Aquakem v. 7.2AQ1

AS

Results from time period:

Thu May 24 14:55:47 2018

Thu May 24 15:22:45 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	0.1339	µg/l	5/24/2018 8:05:07	
5.OPPBCN	A	Total CN	P	4.8734	µg/l	5/24/2018 8:05:08	
10PPBCN	A	Total CN	P	10.5266	µg/l	5/24/2018 8:05:09	
50PPBCN	A	Total CN	P	49.5826	µg/l	5/24/2018 8:05:10	
100PPBCN	A	Total CN	P	101.3357	µg/l	5/24/2018 8:05:11	
250PPBCN	A	Total CN	P	247.5649	µg/l	5/24/2018 8:05:12	
500PPBCN	A	Total CN	P	500.9828	µg/l	5/24/2018 8:05:13	
ICV1	S	Total CN	P	95.3149	µg/l	5/24/2018 15:19:21	
ICB1	S	Total CN	P	0.9286	µg/l	5/24/2018 15:19:22	
CCV1	S	Total CN	P	251.6407	µg/l	5/24/2018 15:19:23	
CCB1	S	Total CN	P	0.753	µg/l	5/24/2018 15:19:24	
PB109648BL	S	Total CN	P	0.7715	µg/l	5/24/2018 15:19:25	
PB109648BS	S	Total CN	P	94.8171	µg/l	5/24/2018 15:19:26	
MIDPB109648	S	Total CN	P	237.0979	µg/l	5/24/2018 15:19:27	
J3093-12	S	Total CN	P	17.0127	µg/l	5/24/2018 15:19:28	
J3093-12DUP	S	Total CN	P	17.1613	µg/l	5/24/2018 15:19:29	
J3093-12MS	S	Total CN	P	52.476	µg/l	5/24/2018 15:19:30	
J3093-12MSD	S	Total CN	P	52.0052	µg/l	5/24/2018 15:19:31	
CCV2	S	Total CN	P	253.6452	µg/l	5/24/2018 15:22:44	
CCB2	S	Total CN	P	0.5207	µg/l	5/24/2018 15:22:45	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/24/2018 14:45

Test Total CN


Accepted 5/24/2018 8:09

Factor 989.2

Bias 0.003

Coeff. of det. 0.999955

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.1339	0.0000	
2	5.0PPBCN	0.008	4.8734	5.0000	
3	10PPBCN	0.014	10.5266	10.0000	
4	50PPBCN	0.053	49.5826	50.0000	
5	100PPBCN	0.106	101.3357	100.0000	
6	250PPBCN	0.253	247.5649	250.0000	
7	500PPBCN	0.510	500.9828	500.0000	