

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/24/2018 15:57

Test: Total CN Mg/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.630	0.0	0.100	
ICB1	0.696	0.0	0.004	
CCV1	240.467	0.0	0.246	
CCB1	0.791	0.0	0.004	
PB109643BL	0.241	0.0	0.003	
PB109643BS	97.318	0.0	0.101	
J3094-04	5.008	0.0	0.008	
J3094-04DUP	4.910	0.0	0.008	
J3094-04MS	42.187	0.0	0.046	
J3094-04MSD	43.076	0.0	0.047	
CCV2	240.774	0.0	0.247	
CCB2	0.091	0.0	0.003	

N	12
Mean	64.266
SD	89.7151
CV%	139.60

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/24/2018 14:46

Test Total CN


Accepted 5/24/2018 8:09

Factor 989.2

Bias 0.003

Coeff. of det. 0.999955

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.1339	0.0000	
2	5.0PPBCN	0.008	4.8734	5.0000	
3	10PPBCN	0.014	10.5266	10.0000	
4	50PPBCN	0.053	49.5826	50.0000	
5	100PPBCN	0.106	101.3357	100.0000	
6	250PPBCN	0.253	247.5649	250.0000	
7	500PPBCN	0.510	500.9828	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Thu May 24 15:52:15 2018

Thu May 24 15:55:03 2018

AS

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.1339	µg/l	5/24/2018 8:05:07	
5.0PPBCN	A	Total CN	P	4.8734	µg/l	5/24/2018 8:05:08	
10PPBCN	A	Total CN	P	10.5266	µg/l	5/24/2018 8:05:09	
50PPBCN	A	Total CN	P	49.5826	µg/l	5/24/2018 8:05:10	
100PPBCN	A	Total CN	P	101.3357	µg/l	5/24/2018 8:05:11	
250PPBCN	A	Total CN	P	247.5649	µg/l	5/24/2018 8:05:12	
500PPBCN	A	Total CN	P	500.9828	µg/l	5/24/2018 8:05:13	
ICV1	S	Total CN	P	95.6304	µg/l	5/24/2018 15:52:15	
ICB1	S	Total CN	P	0.6962	µg/l	5/24/2018 15:52:16	
CCV1	S	Total CN	P	240.4666	µg/l	5/24/2018 15:52:17	
CCB1	S	Total CN	P	0.7908	µg/l	5/24/2018 15:52:18	
PB109643BL	S	Total CN	P	0.2413	µg/l	5/24/2018 15:52:19	
PB109643BS	S	Total CN	P	97.3179	µg/l	5/24/2018 15:52:20	
J3094-04	S	Total CN	P	5.0084	µg/l	5/24/2018 15:52:21	
J3094-04DUP	S	Total CN	P	4.9096	µg/l	5/24/2018 15:52:22	
J3094-04MS	S	Total CN	P	42.1866	µg/l	5/24/2018 15:52:23	
J3094-04MSD	S	Total CN	P	43.076	µg/l	5/24/2018 15:52:24	
CCV2	S	Total CN	P	240.7741	µg/l	5/24/2018 15:52:25	
CCB2	S	Total CN	P	0.0911	µg/l	5/24/2018 15:55:03	