

LB95677

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AL

5/25/2018 16:28

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.075	0.0	0.095	
ICB1	1.067	0.0	0.004	
CCV1	244.687	0.0	0.235	
CCB1	1.087	0.0	0.004	
PB109700BL	1.063	0.0	0.004	
PB109700BS	98.281	0.0	0.096	
MIDPB109700	246.134	0.0	0.237	
J3093-12	17.341	0.0	0.019	
J3093-12DUP	17.024	0.0	0.019	
CCV2	262.826	0.0	0.253	
CCB2	0.807	0.0	0.004	

N	11
Mean	89.763
SD	109.8077
CV%	122.33

LB 95677

AK

Aquakem v. 7.2AQ1

Results from time period:

Fri May 25 16:13:15 2018

Fri May 25 16:16:39 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	0.5505	µg/l	5/25/2018 11:25:04	
5.OPPBCN	A	Total CN	P	4.2413	µg/l	5/25/2018 11:25:05	
10PPBCN	A	Total CN	P	10.0267	µg/l	5/25/2018 11:25:06	
50PPBCN	A	Total CN	P	50.5658	µg/l	5/25/2018 11:25:07	
100PPBCN	A	Total CN	P	100.0107	µg/l	5/25/2018 11:25:08	
250PPBCN	A	Total CN	P	249.3132	µg/l	5/25/2018 11:25:09	
500PPBCN	A	Total CN	P	500.2917	µg/l	5/25/2018 11:25:10	
ICV1	S	Total CN	P	97.0752	µg/l	5/25/2018 16:13:15	
ICB1	S	Total CN	P	1.0672	µg/l	5/25/2018 16:13:16	
CCV1	S	Total CN	P	244.6873	µg/l	5/25/2018 16:13:17	
CCB1	S	Total CN	P	1.087	µg/l	5/25/2018 16:13:18	
PB109700BL	S	Total CN	P	1.0635	µg/l	5/25/2018 16:13:19	
PB109700BS	S	Total CN	P	98.2806	µg/l	5/25/2018 16:13:20	
MIDPB109700	S	Total CN	P	246.1343	µg/l	5/25/2018 16:13:21	
J3093-12	S	Total CN	P	17.3413	µg/l	5/25/2018 16:13:24	
J3093-12DUP	S	Total CN	P	17.0241	µg/l	5/25/2018 16:13:25	
CCV2	S	Total CN	P	262.8257	µg/l	5/25/2018 16:16:38	
CCB2	S	Total CN	P	0.8067	µg/l	5/25/2018 16:16:39	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/25/2018 15:36


Test Total CN

Accepted 5/25/2018 11:25

Factor 1053
Bias 0.003

Coeff. of det. 0.999991

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.5505	0.0000	
2	5.0PPBCN	0.007	4.2413	5.0000	
3	10PPBCN	0.012	10.0267	10.0000	
4	50PPBCN	0.051	50.5658	50.0000	
5	100PPBCN	0.098	100.0107	100.0000	
6	250PPBCN	0.240	249.3132	250.0000	
7	500PPBCN	0.478	500.2917	500.0000	