

UB95680

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

5/29/2018 15:02

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.551	0.0	0.091	
ICB1	2.220	0.0	0.004	
CCV1	253.259	0.0	0.239	
CCB1	2.432	0.0	0.004	
PB109749BL	2.285	0.0	0.004	
PB109749BS	96.240	0.0	0.092	
MIDPB109749	255.818	0.0	0.242	
J3173-01	6.348	0.0	0.008	
J3173-01DUP	6.036	0.0	0.007	
J3173-01MS	45.865	0.0	0.045	
J3173-01MSD	46.369	0.0	0.045	
CCV3	270.827	0.0	0.256	
CCB3	1.999	0.0	0.004	

N 13
Mean 83.481
SD 106.0957
CV% 127.09

=====

Calibration results

Aquakem 7.2AQ1

Page:

0895678 AR
5/29/18

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AR

5/29/2018 11:32


Test Total CN

Accepted 5/29/2018 11:32

Factor 1065
Bias 0.002

Coeff. of det. 0.999955

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	1.3492	0.0000	
2	5.0PPBCN	0.007	5.1542	5.0000	
3	10PPBCN	0.012	11.3410	10.0000	
4	50PPBCN	0.047	48.7076	50.0000	
5	100PPBCN	0.094	98.5577	100.0000	
6	250PPBCN	0.235	249.0020	250.0000	
7	500PPBCN	0.472	500.8884	500.0000	

UB 95680

Aquakem v. 7.2AQ1

Results from time period:

AG

Tue May 29 14:59:02 2018

Tue May 29 15:02:26 2018

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	1.3492	µg/l	5/29/2018 11:20:30	
5.OPPBCN	A	Total CN	P	5.1542	µg/l	5/29/2018 11:20:31	
10PPBCN	A	Total CN	P	11.341	µg/l	5/29/2018 11:20:32	
50PPBCN	A	Total CN	P	48.7076	µg/l	5/29/2018 11:20:33	
100PPBCN	A	Total CN	P	98.5577	µg/l	5/29/2018 11:20:34	
250PPBCN	A	Total CN	P	249.002	µg/l	5/29/2018 11:20:35	
500PPBCN	A	Total CN	P	500.8884	µg/l	5/29/2018 11:20:36	
ICV1	S	Total CN	P	95.5511	µg/l	5/29/2018 14:59:02	
ICB1	S	Total CN	P	2.2198	µg/l	5/29/2018 14:59:03	
CCV1	S	Total CN	P	253.2592	µg/l	5/29/2018 14:59:04	
CCB1	S	Total CN	P	2.4324	µg/l	5/29/2018 14:59:05	
PB109749BL	S	Total CN	P	2.2854	µg/l	5/29/2018 14:59:06	
PB109749BS	S	Total CN	P	96.2404	µg/l	5/29/2018 14:59:07	
MIDPB109749	S	Total CN	P	255.8177	µg/l	5/29/2018 14:59:08	
J3173-01	S	Total CN	P	6.3484	µg/l	5/29/2018 14:59:09	
J3173-01DUP	S	Total CN	P	6.0363	µg/l	5/29/2018 14:59:10	
J3173-01MS	S	Total CN	P	45.8649	µg/l	5/29/2018 14:59:11	
J3173-01MSD	S	Total CN	P	46.3693	µg/l	5/29/2018 14:59:12	
CCV3	S	Total CN	P	270.8275	µg/l	5/29/2018 15:02:25	
CCB3	S	Total CN	P	1.9987	µg/l	5/29/2018 15:02:26	