

4395699

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AR

5/30/2018 11:56

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.024	0.0	0.109	
ICB1	0.004	0.0	0.003	
CCV1	1.046	0.0	0.111	
CCB1	-0.001	0.0	0.003	
PB109711BL	0.004	0.0	0.003	
PB109711BS	1.035	0.0	0.110	
J3135-01	0.004	0.0	0.003	
J3135-01DUP	0.003	0.0	0.003	
J3135-01MS	1.014	0.0	0.108	
J3135-01MSD	1.019	0.0	0.109	
J3135-02	0.029	0.0	0.006	
J3135-03	0.007	0.0	0.004	
J3135-04	0.014	0.0	0.004	
PB109766BL	0.003	0.0	0.003	
CCV2	1.050	0.0	0.112	
CCB2	0.007	0.0	0.004	
PB109766BS	1.024	0.0	0.109	
J3169-01	0.006	0.0	0.004	
J3169-04	0.007	0.0	0.004	
J3169-07	0.003	0.0	0.003	
J3169-07DUP	0.009	0.0	0.004	
J3169-08MS	0.988	0.0	0.105	
J3169-09MSD	1.000	0.0	0.107	
J3169-10	-0.000	0.0	0.003	
J3169-14	-0.019	0.0	0.001	
CCV3	1.041	0.0	0.111	
CCB3	-0.000	0.0	0.003	

N	27
Mean	0.382
SD	0.5017
CV%	131.26

Aquakem v. 7.2AQ1

Results from time period:

Wed May 30 10:18:12 2018

Wed May 30 11:49:09 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0545	mg/l	5/30/2018 10:18:12	
0.1PPM	A	Phenolics-	P	0.1045	mg/l	5/30/2018 10:18:13	
0.25PPM	A	Phenolics-	P	0.2515	mg/l	5/30/2018 10:18:14	
0.50PPM	A	Phenolics-	P	0.5049	mg/l	5/30/2018 10:18:15	
1.0PPM	A	Phenolics-	P	0.9726	mg/l	5/30/2018 10:18:16	
2.0PPM	A	Phenolics-	P	2.0119	mg/l	5/30/2018 10:18:17	
ICV1	S	Phenolics-	P	1.0237	mg/l	5/30/2018 10:57:25	
ICB1	S	Phenolics-	P	0.0038	mg/l	5/30/2018 10:57:26	
CCV1	S	Phenolics-	P	1.0464	mg/l	5/30/2018 10:57:27	
CCB1	S	Phenolics-	P	-0.0005	mg/l	5/30/2018 10:57:28	
PB109711BL	S	Phenolics-	P	0.0039	mg/l	5/30/2018 10:57:29	
PB109711BS	S	Phenolics-	P	1.0345	mg/l	5/30/2018 10:57:30	
J3135-01	S	Phenolics-	P	0.0042	mg/l	5/30/2018 10:57:31	
J3135-01DUP	S	Phenolics-	P	0.0031	mg/l	5/30/2018 10:57:32	
J3135-01MS	S	Phenolics-	P	1.0136	mg/l	5/30/2018 10:57:33	
J3135-01MSD	S	Phenolics-	P	1.0185	mg/l	5/30/2018 10:57:34	
J3135-02	S	Phenolics-	P	0.0293	mg/l	5/30/2018 10:57:35	
J3135-03	S	Phenolics-	P	0.0073	mg/l	5/30/2018 10:57:36	
J3135-04	S	Phenolics-	P	0.0142	mg/l	5/30/2018 11:07:27	
PB109766BL	S	Phenolics-	P	0.0031	mg/l	5/30/2018 11:07:28	
CCV2	S	Phenolics-	P	1.0495	mg/l	5/30/2018 11:07:29	
CCB2	S	Phenolics-	P	0.0069	mg/l	5/30/2018 11:07:30	
PB109766BS	S	Phenolics-	P	1.0243	mg/l	5/30/2018 11:07:31	
J3169-01	S	Phenolics-	P	0.0055	mg/l	5/30/2018 11:07:32	
J3169-04	S	Phenolics-	P	0.0069	mg/l	5/30/2018 11:07:33	
J3169-07	S	Phenolics-	P	0.0035	mg/l	5/30/2018 11:07:34	
J3169-07DUP	S	Phenolics-	P	0.0094	mg/l	5/30/2018 11:07:35	
J3169-08MS	S	Phenolics-	P	0.9878	mg/l	5/30/2018 11:07:36	
J3169-09MSD	S	Phenolics-	P	0.9998	mg/l	5/30/2018 11:07:37	
J3169-10	S	Phenolics-	P	-0.0002	mg/l	5/30/2018 11:07:38	
J3169-14	S	Phenolics-	P	-0.0188	mg/l	5/30/2018 11:12:21	
CCV3	S	Phenolics-	P	1.0413	mg/l	5/30/2018 11:12:22	
CCB3	S	Phenolics-	P	-0.0005	mg/l	5/30/2018 11:12:23	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

5/30/2018 10:18

Test Phenolics-


Accepted 5/30/2018 10:18

Factor 9.651

Bias 0.003

Coeff. of det. 0.999657

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0545	0.0500	
2	Phenol-2	0.014	0.1045	0.1000	
3	Phenol-2	0.029	0.2515	0.2500	
4	Phenol-2	0.055	0.5049	0.5000	
5	Phenol-2	0.104	0.9726	1.0000	
6	Phenol-2	0.211	2.0119	2.0000	