

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

6/1/2018 8:44

Test: *Reactive*
~~Total~~ CN *ug/l*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.132	0.0	0.098	
ICB1	-0.237	0.0	0.003	
CCV1	242.504	0.0	0.244	
CCB1	-0.356	0.0	0.003	
PB109814BL	-0.520	0.0	0.003	
J2669-12	-0.506	0.0	0.003	
J2669-12DUP	-0.338	0.0	0.003	
J2678-06	-0.358	0.0	0.003	
J3202-04	-0.294	0.0	0.003	
J3203-03	-0.476	0.0	0.003	
J3203-06	-0.120	0.0	0.004	
J3217-04	-0.363	0.0	0.003	
J3217-07	-0.212	0.0	0.003	
J3217-11	-0.357	0.0	0.003	
CCV2	243.851	0.0	0.245	
CCB2	-0.084	0.0	0.004	
J3217-15	-0.097	0.0	0.004	
J3218-04	-0.361	0.0	0.003	
J3218-08	-0.247	0.0	0.003	
J3219-04	-0.019	0.0	0.004	
PB109786BL	-0.244	0.0	0.003	
J3191-03	-0.453	0.0	0.003	
J3197-04	-0.808	0.0	0.003	
J3197-08	-0.115	0.0	0.004	
J3188-02	-0.445	0.0	0.003	
J3188-02DUP	-0.309	0.0	0.003	
CCV3	259.584	0.0	0.261	
CCB3	0.214	0.0	0.004	

N	28
Mean	29.784
SD	79.3186
CV%	266.31

Aquakem v. 7.2AQ1

AS

Results from time period:

Fri Jun 01 07:38:08 2018

Fri Jun 01 08:43:25 2018

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and tin	Stat
0.OPPBCN	A	Reactive CN	P	-0.6849	µg/l	6/1/2018 7:38:08	
5.OPPBCN	A	Reactive CN	P	4.2679	µg/l	6/1/2018 7:38:09	
10PPBCN	A	Reactive CN	P	9.2375	µg/l	6/1/2018 7:38:10	
50PPBCN	A	Reactive CN	P	49.6646	µg/l	6/1/2018 7:38:11	
100PPBCN	A	Reactive CN	P	100.5569	µg/l	6/1/2018 7:38:12	
250PPBCN	A	Reactive CN	P	254.0263	µg/l	6/1/2018 7:38:13	
500PPBCN	A	Reactive CN	P	497.9315	µg/l	6/1/2018 7:38:14	
ICV1	S	Reactive CN	P	95.1318	µg/l	6/1/2018 8:10:16	
ICB1	S	Reactive CN	P	-0.2369	µg/l	6/1/2018 8:10:17	
CCV1	S	Reactive CN	P	242.5038	µg/l	6/1/2018 8:10:18	
CCB1	S	Reactive CN	P	-0.3557	µg/l	6/1/2018 8:10:19	
PB109814BL	S	Reactive CN	P	-0.5204	µg/l	6/1/2018 8:10:20	
J2669-12	S	Reactive CN	P	-0.506	µg/l	6/1/2018 8:10:21	
J2669-12DUP	S	Reactive CN	P	-0.3382	µg/l	6/1/2018 8:10:22	
J2678-06	S	Reactive CN	P	-0.358	µg/l	6/1/2018 8:10:23	
J3202-04	S	Reactive CN	P	-0.2945	µg/l	6/1/2018 8:10:24	
J3203-03	S	Reactive CN	P	-0.4762	µg/l	6/1/2018 8:10:25	
J3203-06	S	Reactive CN	P	-0.1196	µg/l	6/1/2018 8:10:26	
J3217-04	S	Reactive CN	P	-0.3633	µg/l	6/1/2018 8:17:49	
J3217-07	S	Reactive CN	P	-0.2119	µg/l	6/1/2018 8:17:50	
J3217-11	S	Reactive CN	P	-0.3566	µg/l	6/1/2018 8:17:51	
CCV2	S	Reactive CN	P	243.851	µg/l	6/1/2018 8:17:52	
CCB2	S	Reactive CN	P	-0.084	µg/l	6/1/2018 8:17:53	
J3217-15	S	Reactive CN	P	-0.0965	µg/l	6/1/2018 8:17:54	
J3218-04	S	Reactive CN	P	-0.3614	µg/l	6/1/2018 8:17:55	
J3218-08	S	Reactive CN	P	-0.2466	µg/l	6/1/2018 8:17:56	
J3219-04	S	Reactive CN	P	-0.0195	µg/l	6/1/2018 8:17:57	
PB109786BL	S	Reactive CN	P	-0.244	µg/l	6/1/2018 8:17:58	
J3191-03	S	Reactive CN	P	-0.453	µg/l	6/1/2018 8:23:05	
J3197-04	S	Reactive CN	P	-0.8081	µg/l	6/1/2018 8:23:06	
J3197-08	S	Reactive CN	P	-0.1147	µg/l	6/1/2018 8:23:07	
J3188-02	S	Reactive CN	P	-0.4446	µg/l	6/1/2018 8:43:22	
J3188-02DUP	S	Reactive CN	P	-0.3092	µg/l	6/1/2018 8:43:23	
CCV3	S	Reactive CN	P	259.5837	µg/l	6/1/2018 8:43:24	
CCB3	S	Reactive CN	P	0.214	µg/l	6/1/2018 8:43:25	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

6/1/2018 7:40


Test Total CN

Accepted 6/1/2018 7:40

Factor 1010
Bias 0.004

Coeff. of det. 0.999891

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.6849	0.0000	
2	5.0PPBCN	0.008	4.2679	5.0000	
3	10PPBCN	0.013	9.2375	10.0000	
4	50PPBCN	0.053	49.6646	50.0000	
5	100PPBCN	0.103	100.5569	100.0000	
6	250PPBCN	0.255	254.0263	250.0000	
7	500PPBCN	0.496	497.9315	500.0000	