

2B95758

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : As

6/1/2018 10:52

Test: Total CN µg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.525	0.0	0.098	
ICB1	0.359	0.0	0.004	
CCV1	243.102	0.0	0.244	
CCB1	0.244	0.0	0.004	
PB109791BL	-0.264	0.0	0.003	
PB109791BS	97.269	0.0	0.100	
MIDPB109791	246.928	0.0	0.248	
J3192-12	22.476	0.0	0.026	
J3192-12DUP	22.579	0.0	0.026	
J3192-12MS	55.127	0.0	0.058	
J3192-12MSD	54.702	0.0	0.058	
PB109833BL	-0.021	0.0	0.004	
PB109833BS	97.537	0.0	0.100	
J3212-12	23.869	0.0	0.027	
CCV2	246.833	0.0	0.248	
CCB2	0.406	0.0	0.004	
J3212-12DUP	23.636	0.0	0.027	
J3212-12MS	56.257	0.0	0.059	
J3212-12MSD	56.119	0.0	0.059	
CCV3	247.280	0.0	0.248	
CCB3	0.363	0.0	0.004	
N	21			
Mean	75.730			
SD	90.6366			
CV%	119.68			

Aquakem v. 7.2AQ1

AS

Results from time period:

Fri Jun 01 10:39:55 2018

Fri Jun 01 10:47:11 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.6849	µg/l	6/1/2018 7:38:08	
5.0PPBCN	A	Total CN	P	4.2679	µg/l	6/1/2018 7:38:09	
10PPBCN	A	Total CN	P	9.2375	µg/l	6/1/2018 7:38:10	
50PPBCN	A	Total CN	P	49.6646	µg/l	6/1/2018 7:38:11	
100PPBCN	A	Total CN	P	100.5569	µg/l	6/1/2018 7:38:12	
250PPBCN	A	Total CN	P	254.0263	µg/l	6/1/2018 7:38:13	
500PPBCN	A	Total CN	P	497.9315	µg/l	6/1/2018 7:38:14	
ICV1	S	Total CN	P	95.5252	µg/l	6/1/2018 10:39:55	
ICB1	S	Total CN	P	0.3586	µg/l	6/1/2018 10:39:56	
CCV1	S	Total CN	P	243.1019	µg/l	6/1/2018 10:39:57	
CCB1	S	Total CN	P	0.2445	µg/l	6/1/2018 10:39:58	
PB109791BL	S	Total CN	P	-0.2644	µg/l	6/1/2018 10:39:59	
PB109791BS	S	Total CN	P	97.2686	µg/l	6/1/2018 10:40:00	
MIDPB109791	S	Total CN	P	246.9276	µg/l	6/1/2018 10:40:01	
J3192-12	S	Total CN	P	22.4755	µg/l	6/1/2018 10:40:02	
J3192-12DUP	S	Total CN	P	22.5793	µg/l	6/1/2018 10:40:03	
J3192-12MS	S	Total CN	P	55.1267	µg/l	6/1/2018 10:40:04	
J3192-12MSD	S	Total CN	P	54.7018	µg/l	6/1/2018 10:40:05	
PB109833BL	S	Total CN	P	-0.0213	µg/l	6/1/2018 10:47:02	
PB109833BS	S	Total CN	P	97.5372	µg/l	6/1/2018 10:47:03	
J3212-12	S	Total CN	P	23.8686	µg/l	6/1/2018 10:47:04	
CCV2	S	Total CN	P	246.8332	µg/l	6/1/2018 10:47:05	
CCB2	S	Total CN	P	0.4062	µg/l	6/1/2018 10:47:06	
J3212-12DUP	S	Total CN	P	23.6363	µg/l	6/1/2018 10:47:07	
J3212-12MS	S	Total CN	P	56.2574	µg/l	6/1/2018 10:47:08	
J3212-12MSD	S	Total CN	P	56.119	µg/l	6/1/2018 10:47:09	
CCV3	S	Total CN	P	247.2797	µg/l	6/1/2018 10:47:10	
CCB3	S	Total CN	P	0.3627	µg/l	6/1/2018 10:47:11	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

6/1/2018 7:40


Test Total CN

Accepted 6/1/2018 7:40

Factor 1010
Bias 0.004

Coeff. of det. 0.999891

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.6849	0.0000	
2	5.0PPBCN	0.008	4.2679	5.0000	
3	10PPBCN	0.013	9.2375	10.0000	
4	50PPBCN	0.053	49.6646	50.0000	
5	100PPBCN	0.103	100.5569	100.0000	
6	250PPBCN	0.255	254.0263	250.0000	
7	500PPBCN	0.496	497.9315	500.0000	