

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

6/5/2018 11:58

Test: *Reactive*
~~Total~~ CN *ug/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.638	0.0	0.096	
ICB1	-0.224	0.0	0.004	
CCV1	244.449	0.0	0.234	
CCB1	-0.503	0.0	0.004	
PB109978BL	-0.692	0.0	0.004	
J2309-06	-0.810	0.0	0.004	
J2309-06DUP	-1.292	0.0	0.003	
J2309-10	1.704	0.0	0.006	
J3195-11	-1.142	0.0	0.003	
J3300-04	-0.418	0.0	0.004	
J3304-03	-0.255	0.0	0.004	
J3304-06	-1.144	0.0	0.003	
J3305-06	-0.808	0.0	0.004	
J3307-01	-0.549	0.0	0.004	
CCV2	259.560	0.0	0.248	
CCB2	-0.682	0.0	0.004	

N	16
Mean	37.177
SD	87.3904
CV%	235.07

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

6/5/2018 8:12


Test Total CN

Accepted 6/5/2018 8:12

Factor 1063
Bias 0.004

Coeff. of det. 0.999905

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-1.2826	0.0000	
2	5.0PPBCN	0.008	3.7822	5.0000	
3	10PPBCN	0.014	9.7855	10.0000	
4	50PPBCN	0.052	50.5947	50.0000	
5	100PPBCN	0.099	100.5307	100.0000	
6	250PPBCN	0.243	253.4773	250.0000	
7	500PPBCN	0.473	498.1122	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Tue Jun 05 11:50:56 2018

Tue Jun 05 11:55:47 2018

AS

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-1.2826	µg/l	6/5/2018 8:05:31	
5.OPPBCN	A	Reactive CN	P	3.7822	µg/l	6/5/2018 8:05:32	
10PPBCN	A	Reactive CN	P	9.7855	µg/l	6/5/2018 8:05:33	
50PPBCN	A	Reactive CN	P	50.5947	µg/l	6/5/2018 8:05:34	
100PPBCN	A	Reactive CN	P	100.5307	µg/l	6/5/2018 8:05:35	
250PPBCN	A	Reactive CN	P	253.4773	µg/l	6/5/2018 8:05:36	
500PPBCN	A	Reactive CN	P	498.1122	µg/l	6/5/2018 8:05:37	
ICV1	S	Reactive CN	P	97.6379	µg/l	6/5/2018 11:50:56	
ICB1	S	Reactive CN	P	-0.2241	µg/l	6/5/2018 11:50:57	
CCV1	S	Reactive CN	P	244.4495	µg/l	6/5/2018 11:50:58	
CCB1	S	Reactive CN	P	-0.5035	µg/l	6/5/2018 11:50:59	
PB109978BL	S	Reactive CN	P	-0.6919	µg/l	6/5/2018 11:51:00	
J2309-06	S	Reactive CN	P	-0.8097	µg/l	6/5/2018 11:51:01	
J2309-06DUP	S	Reactive CN	P	-1.2915	µg/l	6/5/2018 11:51:02	
J2309-10	S	Reactive CN	P	1.7044	µg/l	6/5/2018 11:51:03	
J3195-11	S	Reactive CN	P	-1.1421	µg/l	6/5/2018 11:51:04	
J3300-04	S	Reactive CN	P	-0.4181	µg/l	6/5/2018 11:51:05	
J3304-03	S	Reactive CN	P	-0.2553	µg/l	6/5/2018 11:51:06	
J3304-06	S	Reactive CN	P	-1.1443	µg/l	6/5/2018 11:55:43	
J3305-06	S	Reactive CN	P	-0.808	µg/l	6/5/2018 11:55:44	
J3307-01	S	Reactive CN	P	-0.5487	µg/l	6/5/2018 11:55:45	
CCV2	S	Reactive CN	P	259.5597	µg/l	6/5/2018 11:55:46	
CCB2	S	Reactive CN	P	-0.6818	µg/l	6/5/2018 11:55:47	