

21395865

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

6/7/2018 11:21

Test: *Reactive*
~~Total~~ CN, *mg/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	98.331	0.0	0.106	
ICB1	-0.687	0.0	0.004	
CCV1	246.996	0.0	0.259	
CCB1	-0.917	0.0	0.003	
PB110032BL	-0.360	0.0	0.004	
J3252-02	-0.956	0.0	0.003	
J3252-02DUP	-0.444	0.0	0.004	
J3344-04	-1.027	0.0	0.003	
J3344-08	-0.656	0.0	0.004	
J3344-12	-0.772	0.0	0.004	
J3352-04	-0.767	0.0	0.004	
CCV2	259.338	0.0	0.272	
CCB2	-0.950	0.0	0.003	

N	13
Mean	45.933
SD	95.9646
CV%	208.92

Aquakem v. 7.2AQ1

AS

Results from time period:

Thu Jun 07 09:13:03 2018

Thu Jun 07 11:21:32 2018

Sample Id	Sam/Ctr/c/	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-1.2083	µg/l	6/7/2018 10:07:46	
5.0PPBCN	A	Reactive CN	P	4.229	µg/l	6/7/2018 10:07:47	
10PPBCN	A	Reactive CN	P	9.6063	µg/l	6/7/2018 10:07:48	
50PPBCN	A	Reactive CN	P	50.2503	µg/l	6/7/2018 10:07:49	
100PPBCN	A	Reactive CN	P	100.9506	µg/l	6/7/2018 10:07:50	
250PPBCN	A	Reactive CN	P	252.7433	µg/l	6/7/2018 10:07:51	
500PPBCN	A	Reactive CN	P	498.4288	µg/l	6/7/2018 10:07:52	
ICV1	S	Reactive CN	P	98.331	µg/l	6/7/2018 11:18:08	
ICB1	S	Reactive CN	P	-0.6868	µg/l	6/7/2018 11:18:09	
CCV1	S	Reactive CN	P	246.9961	µg/l	6/7/2018 11:18:10	
CCB1	S	Reactive CN	P	-0.9175	µg/l	6/7/2018 11:18:11	
PB110032BL	S	Reactive CN	P	-0.3604	µg/l	6/7/2018 11:18:12	
J3252-02	S	Reactive CN	P	-0.9556	µg/l	6/7/2018 11:18:13	
J3252-02DUP	S	Reactive CN	P	-0.4444	µg/l	6/7/2018 11:18:14	
J3344-04	S	Reactive CN	P	-1.027	µg/l	6/7/2018 11:18:15	
J3344-08	S	Reactive CN	P	-0.6562	µg/l	6/7/2018 11:18:16	
J3344-12	S	Reactive CN	P	-0.7724	µg/l	6/7/2018 11:18:17	
J3352-04	S	Reactive CN	P	-0.767	µg/l	6/7/2018 11:18:18	
CCV2	S	Reactive CN	P	259.3378	µg/l	6/7/2018 11:21:31	
CCB2	S	Reactive CN	P	-0.9503	µg/l	6/7/2018 11:21:32	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AB

6/7/2018 10:08


Test Total CN

Accepted 6/7/2018 10:08

Factor 968.7
Bias 0.004

Coeff. of det. 0.999936

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-1.2083	0.0000	
2	5.0PPBCN	0.009	4.2290	5.0000	
3	10PPBCN	0.014	9.6063	10.0000	
4	50PPBCN	0.056	50.2503	50.0000	
5	100PPBCN	0.109	100.9506	100.0000	
6	250PPBCN	0.265	252.7433	250.0000	
7	500PPBCN	0.519	498.4288	500.0000	