

2B95902

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

6/8/2018 12:53

Test: *Reactive* ~~Total~~ CN *µg/L*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.023	0.0	0.102	
ICB1	-0.859	0.0	0.004	
CCV1	241.625	0.0	0.252	
CCB1	-0.880	0.0	0.004	
PB110075BL	-1.150	0.0	0.003	
J3240-02	-0.821	0.0	0.004	
J3240-02DUP	-0.974	0.0	0.004	
J3247-02	-0.678	0.0	0.004	
J3360-03	-0.605	0.0	0.004	
J3377-02	-0.827	0.0	0.004	
J3377-05	-0.763	0.0	0.004	
J3377-09	-0.813	0.0	0.004	
CCV2	258.420	0.0	0.269	
CCB2	-1.086	0.0	0.003	

N	14
Mean	41.829
SD	91.8633
CV%	219.61

Aquakem v. 7.2AQ1

Results from time period:

Fri Jun 08 12:49:28 2018

Fri Jun 08 12:53:21 2018



Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.9092	µg/l	6/8/2018 9:59:56	
5.OPPBCN	A	Reactive CN	P	4.0392	µg/l	6/8/2018 9:59:57	
10PPBCN	A	Reactive CN	P	10.5541	µg/l	6/8/2018 9:59:58	
50PPBCN	A	Reactive CN	P	49.6974	µg/l	6/8/2018 9:59:59	
100PPBCN	A	Reactive CN	P	101.1604	µg/l	6/8/2018 10:00:00	
250PPBCN	A	Reactive CN	P	251.3228	µg/l	6/8/2018 10:00:01	
500PPBCN	A	Reactive CN	P	499.1353	µg/l	6/8/2018 10:00:02	
ICV1	S	Reactive CN	P	95.0229	µg/l	6/8/2018 12:49:28	
ICB1	S	Reactive CN	P	-0.8588	µg/l	6/8/2018 12:49:29	
CCV1	S	Reactive CN	P	241.625	µg/l	6/8/2018 12:49:30	
CCB1	S	Reactive CN	P	-0.88	µg/l	6/8/2018 12:49:31	
PB110075BL	S	Reactive CN	P	-1.1496	µg/l	6/8/2018 12:49:32	
J3240-02	S	Reactive CN	P	-0.8206	µg/l	6/8/2018 12:49:33	
J3240-02DUP	S	Reactive CN	P	-0.9741	µg/l	6/8/2018 12:49:34	
J3247-02	S	Reactive CN	P	-0.6778	µg/l	6/8/2018 12:49:35	
J3360-03	S	Reactive CN	P	-0.6046	µg/l	6/8/2018 12:49:36	
J3377-02	S	Reactive CN	P	-0.8269	µg/l	6/8/2018 12:49:37	
J3377-05	S	Reactive CN	P	-0.7634	µg/l	6/8/2018 12:49:38	
J3377-09	S	Reactive CN	P	-0.8129	µg/l	6/8/2018 12:53:19	
CCV2	S	Reactive CN	P	258.4199	µg/l	6/8/2018 12:53:20	
CCB2	S	Reactive CN	P	-1.0864	µg/l	6/8/2018 12:53:21	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

6/8/2018 10:15


Test Total CN

Accepted 6/8/2018 10:15

Factor 977.7
 Bias 0.005

Coeff. of det. 0.999971

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.9092	0.0000	
2	5.0PPBCN	0.009	4.0392	5.0000	
3	10PPBCN	0.015	10.5541	10.0000	
4	50PPBCN	0.055	49.6974	50.0000	
5	100PPBCN	0.108	101.1604	100.0000	
6	250PPBCN	0.262	251.3228	250.0000	
7	500PPBCN	0.515	499.1353	500.0000	