


7471B

LB 96008
INSTRUMENT ID: CV4

Linear

 μ Abs.:

43835



A= 0.0000e+000

B= 2.2849e-004

C= -1.9090e-003

Rho= 0.9997776

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	39	0.000	39				
0.2	0.200	0.207	0.007	916	0.0 %	916				
2.5	2.500	2.575	0.075	11277	0.0 %	11277				
5.0	5.000	4.836	-0.164	21175	0.0 %	21175				
7.5	7.500	7.561	0.061	33098	0.0 %	33098				
10.0	10.000	10.014	0.014	43835	0.0 %	43835				

LB96008 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Jun 2018 14:35:02

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	39	-	0.00007471B	PPB		14 Jun 2018 14:37:58	S	Std
0.2 - 1	50.2	916	-	0.20007471B	PPB		14 Jun 2018 14:40:16	S	Std
2.5 - 1	52.5	11277	-	2.50007471B	PPB		14 Jun 2018 14:42:34	S	Std
5.0 - 1	55	21175	-	5.00007471B	PPB		14 Jun 2018 14:44:51	S	Std
7.5 - 1	57.5	33098	-	7.50007471B	PPB		14 Jun 2018 14:47:08	S	Std
10.0 - 1	510	43835	-	10.00007471B	PPB		14 Jun 2018 14:49:26	S	Std
ICV02 - 1	ICV02	16937	3.8680	-7471B	PPB		14 Jun 2018 14:52:06	U	SMPL
ICB02 - 1	ICB02	-133	-0.0323	-7471B	PPB		14 Jun 2018 14:54:22	U	SMPL
CCV29 - 1	CCV29	22696	5.1839	-7471B	PPB		14 Jun 2018 14:56:39	U	SMPL
CCB29 - 1	CCB29	-147	-0.0355	-7471B	PPB		14 Jun 2018 14:58:55	U	SMPL
CRA - 1	CRA	893	0.2021	-7471B	PPB		14 Jun 2018 15:01:11	U	SMPL
HighStd - 1	HighStd	43687	9.9801	-7471B	PPB		14 Jun 2018 15:03:27	U	SMPL
ChkStd - 1	ChkStd	31070	7.0972	-7471B	PPB		14 Jun 2018 15:05:43	U	SMPL
PB110228BL - 1	PBS	-198	-0.0471	-7471B	PPB		14 Jun 2018 15:08:01	U	SMPL
PB110228BS - 1	LCSS	16287	3.7195	-7471B	PPB		14 Jun 2018 15:10:18	U	SMPL
J3239-09 - 1	AU-06-061318-A	2518	0.5734	-7471B	PPB		14 Jun 2018 15:12:35	U	SMPL
J3239-11 - 1	AU-06-061318-B	1607	0.3653	-7471B	PPB		14 Jun 2018 15:14:53	U	SMPL
J3239-13 - 1	AU-06-061318-C	2453	0.5586	-7471B	PPB		14 Jun 2018 15:17:10	U	SMPL
J3241-07 - 1	HR-05-061318-A	4292	0.9788	-7471B	PPB		14 Jun 2018 15:19:28	U	SMPL
J3241-09 - 1	HR-05-061318-B	9399	2.1457	-7471B	PPB		14 Jun 2018 15:21:44	U	SMPL
CCV30 - 1	CCV30	21766	4.9714	-7471B	PPB		14 Jun 2018 15:24:00	U	SMPL
CCB30 - 1	CCB30	-107	-0.0264	-7471B	PPB		14 Jun 2018 15:26:16	U	SMPL
J3241-11 - 1	HR-05-061318-C	5517	1.2587	-7471B	PPB		14 Jun 2018 15:28:32	U	SMPL
J3241-11DUP - 1	HR-05-061318-CDUP	6035	1.3770	-7471B	PPB		14 Jun 2018 15:30:48	U	SMPL
J3241-11MS - 1	HR-05-061318-CMS	22164	5.0623	-7471B	PPB		14 Jun 2018 15:33:04	U	SMPL
J3241-11MSD - 1	HR-05-061318-CMSD	23246	5.3095	-7471B	PPB		14 Jun 2018 15:35:20	U	SMPL
J3459-01 - 1	MH-RO-341	1019	0.2309	-7471B	PPB		14 Jun 2018 15:37:37	U	SMPL
J3465-01 - 1	TP-A1-WC-1	519	0.1167	-7471B	PPB		14 Jun 2018 15:39:55	U	SMPL
J3465-03 - 1	TP-A2-WC-1	766	0.1731	-7471B	PPB		14 Jun 2018 15:42:13	U	SMPL
J3465-05 - 1	TP-B4-WC-1	531	0.1194	-7471B	PPB		14 Jun 2018 15:44:30	U	SMPL
J3471-02 - 1	#1	4821	1.0996	-7471B	PPB		14 Jun 2018 15:46:48	U	SMPL
J3471-04 - 1	#2	5380	1.2274	-7471B	PPB		14 Jun 2018 15:49:04	U	SMPL
CCV31 - 1	CCV31	21879	4.9972	-7471B	PPB		14 Jun 2018 15:51:20	U	SMPL
CCB31 - 1	CCB31	-105	-0.0259	-7471B	PPB		14 Jun 2018 15:53:36	U	SMPL
J3480-01 - 1	OILY-STONES	60	0.0118	-7471B	PPB		14 Jun 2018 15:55:52	U	SMPL
J3481-01 - 1	FK-GF-02-021-A	3254	0.7416	-7471B	PPB		14 Jun 2018 15:58:08	U	SMPL
J3481-03 - 1	FK-GF-02-021-B	2247	0.5115	-7471B	PPB		14 Jun 2018 16:00:25	U	SMPL
J3482-01 - 1	FK-PS-01-029	561	0.1263	-7471B	PPB		14 Jun 2018 16:02:41	U	SMPL
J3487-01 - 1	OILY-SOIL	5219	1.1906	-7471B	PPB		14 Jun 2018 16:04:57	U	SMPL
J3488-01 - 1	FS-RO-3091	8532	1.9476	-7471B	PPB		14 Jun 2018 16:07:14	U	SMPL
J3490-01 - 1	CONCRETE-BASE	57	0.0111	-7471B	PPB		14 Jun 2018 16:09:32	U	SMPL
J3493-01 - 1	WC-1	1321	0.2999	-7471B	PPB		14 Jun 2018 16:11:50	U	SMPL
PB110229BL - 1	PBS	33	0.0056	-7471B	PPB		14 Jun 2018 16:14:08	U	SMPL
PB110229BS - 1	LCSS	17336	3.9592	-7471B	PPB		14 Jun 2018 16:16:24	U	SMPL
CCV32 - 1	CCV32	21936	5.0102	-7471B	PPB		14 Jun 2018 16:18:40	U	SMPL
CCB32 - 1	CCB32	-128	-0.0312	-7471B	PPB		14 Jun 2018 16:20:57	U	SMPL
J3498-11 - 1	TP-E-S	2232	0.5081	-7471B	PPB		14 Jun 2018 16:23:13	U	SMPL
J3498-15 - 1	TP-D-S	2336	0.5318	-7471B	PPB		14 Jun 2018 16:25:29	U	SMPL
J3501-03 - 1	TP-10-S	6961	1.5886	-7471B	PPB		14 Jun 2018 16:27:46	U	SMPL
J3501-07 - 1	TP-10-D	578	0.1302	-7471B	PPB		14 Jun 2018 16:30:02	U	SMPL
J3501-11 - 1	TP-9	1014	0.2298	-7471B	PPB		14 Jun 2018 16:32:19	U	SMPL
J3501-11DUP - 1	TP-9DUP	1178	0.2673	-7471B	PPB		14 Jun 2018 16:34:36	U	SMPL
J3501-11MS - 1	TP-9MS	17919	4.0924	-7471B	PPB		14 Jun 2018 16:36:53	U	SMPL
J3501-11MSD - 1	TP-9MSD	17666	4.0346	-7471B	PPB		14 Jun 2018 16:39:10	U	SMPL
CCV33 - 1	CCV33	21924	5.0075	-7471B	PPB		14 Jun 2018 16:41:29	U	SMPL
CCB33 - 1	CCB33	-126	-0.0307	-7471B	PPB		14 Jun 2018 16:43:46	U	SMPL
J3241-11LX5 - 1		1029	0.2332	-7471B	PPB		14 Jun 2018 16:46:02	U	SMPL
J3241-11A - 1		23209	5.3011	-7471B	PPB		14 Jun 2018 16:48:18	U	SMPL
J3501-11LX5 - 1		75	0.0152	-7471B	PPB		14 Jun 2018 16:50:35	U	SMPL
J3501-11A - 1		18862	4.3078	-7471B	PPB		14 Jun 2018 16:52:51	U	SMPL
CCV34 - 1	CCV34	21943	5.0118	-7471B	PPB		14 Jun 2018 16:55:07	U	SMPL
CCB34 - 1	CCB34	-119	-0.0291	-7471B	PPB		14 Jun 2018 16:57:24	U	SMPL