

2B96083

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

6/18/2018 14:17

Test: ^{Reactive} Total CN ^{Mg/d}

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.348	0.0	0.091	
ICB1	0.287	0.0	0.004	
CCV1	247.397	0.0	0.226	
CCB1	0.053	0.0	0.004	
PB110325BL	-0.020	0.0	0.003	
J3240-04	0.426	0.0	0.004	
J3240-04DUP	-0.174	0.0	0.003	
J3244-02	0.350	0.0	0.004	
J3538-06	0.155	0.0	0.004	
J3538-12	0.011	0.0	0.004	
J3551-05	-0.286	0.0	0.003	
J3566-04	-0.128	0.0	0.003	
J3566-08	-0.275	0.0	0.003	
J3566-12	0.395	0.0	0.004	
CCV2	254.618	0.0	0.233	
CCB2	0.069	0.0	0.004	
J3566-16	0.086	0.0	0.004	
J3566-20	-0.433	0.0	0.003	
J3567-04	-0.931	0.0	0.003	
J3567-08	0.113	0.0	0.004	
J3567-12	-0.066	0.0	0.003	
PB110326BL	-0.094	0.0	0.003	
J3532-02	0.312	0.0	0.004	
J3532-02DUP	-0.119	0.0	0.003	
CCV3	264.006	0.0	0.241	
CCB3	0.149	0.0	0.004	

N 26
Mean 33.202
SD 84.0362
CV% 253.11

Aquakem v. 7.2AQ1

AS

Results from time period:

Mon Jun 18 13:31:39 2018

Mon Jun 18 14:12:04 2018

Sample Id	Sam/Ctr/c	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.7059	µg/l	6/18/2018 13:31:39	
5.OPPBCN	A	Reactive CN	P	4.879	µg/l	6/18/2018 13:31:40	
10PPBCN	A	Reactive CN	P	9.5709	µg/l	6/18/2018 13:31:41	
50PPBCN	A	Reactive CN	P	50.4565	µg/l	6/18/2018 13:31:42	
100PPBCN	A	Reactive CN	P	99.6579	µg/l	6/18/2018 13:31:43	
250PPBCN	A	Reactive CN	P	252.2184	µg/l	6/18/2018 13:31:44	
500PPBCN	A	Reactive CN	P	498.9234	µg/l	6/18/2018 13:31:45	
ICV1	S	Reactive CN	P	97.348	µg/l	6/18/2018 14:00:09	
ICB1	S	Reactive CN	P	0.287	µg/l	6/18/2018 14:00:10	
CCV1	S	Reactive CN	P	247.3973	µg/l	6/18/2018 14:00:11	
CCB1	S	Reactive CN	P	0.0528	µg/l	6/18/2018 14:00:12	
PB110325BL	S	Reactive CN	P	-0.0195	µg/l	6/18/2018 14:00:13	
J3240-04	S	Reactive CN	P	0.4256	µg/l	6/18/2018 14:00:14	
J3240-04DUP	S	Reactive CN	P	-0.1737	µg/l	6/18/2018 14:00:15	
J3244-02	S	Reactive CN	P	0.3496	µg/l	6/18/2018 14:00:16	
J3538-06	S	Reactive CN	P	0.1545	µg/l	6/18/2018 14:00:17	
J3538-12	S	Reactive CN	P	0.0115	µg/l	6/18/2018 14:00:18	
J3551-05	S	Reactive CN	P	-0.2857	µg/l	6/18/2018 14:00:19	
J3566-04	S	Reactive CN	P	-0.1283	µg/l	6/18/2018 14:07:42	
J3566-08	S	Reactive CN	P	-0.2753	µg/l	6/18/2018 14:07:43	
J3566-12	S	Reactive CN	P	0.3953	µg/l	6/18/2018 14:07:44	
CCV2	S	Reactive CN	P	254.6176	µg/l	6/18/2018 14:07:45	
CCB2	S	Reactive CN	P	0.0691	µg/l	6/18/2018 14:07:46	
J3566-16	S	Reactive CN	P	0.0859	µg/l	6/18/2018 14:07:47	
J3566-20	S	Reactive CN	P	-0.4329	µg/l	6/18/2018 14:07:48	
J3567-04	S	Reactive CN	P	-0.9314	µg/l	6/18/2018 14:07:49	
J3567-08	S	Reactive CN	P	0.1125	µg/l	6/18/2018 14:07:50	
J3567-12	S	Reactive CN	P	-0.0658	µg/l	6/18/2018 14:07:51	
PB110326BL	S	Reactive CN	P	-0.0937	µg/l	6/18/2018 14:07:52	
J3532-02	S	Reactive CN	P	0.3122	µg/l	6/18/2018 14:12:01	
J3532-02DUP	S	Reactive CN	P	-0.1189	µg/l	6/18/2018 14:12:02	
CCV3	S	Reactive CN	P	264.0064	µg/l	6/18/2018 14:12:03	
CCB3	S	Reactive CN	P	0.149	µg/l	6/18/2018 14:12:04	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : As

6/18/2018 13:32


Test Total CN

Accepted 6/18/2018 13:32

Factor 1110
Bias 0.004

Coeff. of det. 0.999965

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.7059	0.0000	
2	5.0PPBCN	0.008	4.8790	5.0000	
3	10PPBCN	0.012	9.5709	10.0000	
4	50PPBCN	0.049	50.4565	50.0000	
5	100PPBCN	0.093	99.6579	100.0000	
6	250PPBCN	0.231	252.2184	250.0000	
7	500PPBCN	0.453	498.9234	500.0000	